


Name: Shanghai Institute of Quality Inspection and Technical Research
Address: No.900, Jiangyue Road, Minhang District, Shanghai, China (Metrology Testing)
Registration No. CNAS L0128
Accreditation Criteria: ISO/IEC 17025:2017 and relevant requirements of CNAS
Effective Date: 2024-09-20 Expiry Date: 2030-10-11

SCHEDULE 5 ACCREDITED CALIBRATION AND MEASUREMENT CAPABILITY SCOPE

Note: The instruments with * represents onsite calibration can be performed.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
一、Thermometry							
1	*Digital Temperature Indicators and Controllers	Temperature	V.R.of Digital Temperature Indicators and Controllers JJG617	thermal resistor(-200~600)℃	U=0.20℃		
				Type T Thermocouple (-200~0)℃	U=0.7℃		
				Type T Thermocouple (0~400)℃	U=0.6℃		
				Type K Thermocouple (-200~-100)℃	U=0.7℃		
				Type K Thermocouple (-100~1372)℃	U=0.6℃		
				Type J Thermocouple (-210~-100)℃	U=0.6℃		
				Type J Thermocouple (-100~1200)℃	U=0.5℃		



No. CNAS L0128

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				Type B Thermocouple (600~800)°C	U=1.4°C		
				Type B Thermocouple (800~1820)°C	U=1.2°C		
				Type R Thermocouple (-20~0)°C	U=1.6°C		
				Type R Thermocouple (0~100)°C	U=1.5°C		
				Type R Thermocouple (100~1767)°C	U=1.3°C		
				Type S Thermocouple (-20~0)°C	U=1.6°C		
				Type S Thermocouple (0~200)°C	U=1.5°C		
				Type S Thermocouple (200~1400)°C	U=1.3°C		
				Type S Thermocouple (1400~1767)°C	U=1.4°C		
				Type E Thermocouple (-200~600)°C	U=0.6°C		
				Type E Thermocouple (600~1000)°C	U=0.5°C		
				Type N Thermocouple (-200~-100)°C	U=0.9°C		
				Type N Thermocouple (-100~900)°C	U=0.8°C		
				Type N Thermocouple (900~1300)°C	U=0.6°C		
2	*Vacuum drying oven	temperature	Calibration Specification for Vacuum Ovens JJF(min)1093	(0~300)°C	U=0.4°C		



No. CNAS L0128

第 2 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Pressure		(-90~200) kPa	U=0.8kPa		
3	*High temperature test chambers	Temperature	Calibration Specification for the Low/high temperature test chambers SQI/JL-JF-46	(300~500) °C	U=1.2°C		
4	*Environmental test equipment	Temperature	Calibration specification for Environmental testing Equipment for Temperature and Humidity Parameters. JJF1101	(-80~300) °C	U=0.22°C		
		Humidity		(10~98) %RH	U=1.2%RH		
5	*Aging life test chamber/Low pressure test chamber	Differential Pressure	Calibration Specification for the Equipment of the Environmental Testing for Differential Pressure and Wind Speed and illuminance SQI/JL-JF-47	(1~190) Pa	U=0.5Pa		
		Wind Speed		(0.2~20)m/s	U=0.10m/s		
6	*Recorders for Industrial Process Measurement	Temperature	Recorders for Industrial Process Measurement JJG74	thermal resistor:(-200~600)°C	U=0.2°C		
				Type T Thermocouple (-200~0)°C	U=0.7°C		
				Type T Thermocouple(0~400)°C	U=0.6°C		
				Type K Thermocouple (-200~-100)°C	U=0.7°C		
				Type K Thermocouple (-100~1372) °C	U=0.6°C		
				Type J Thermocouple (-210~-100)°C	U=0.6°C		
				Type J Thermocouple(-100~1200)°C	U=0.5°C		




No. CNAS L0128

第 3 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				Type B Thermocouple (600~800) °C	$U=1.4^{\circ}\text{C}$		
				Type B Thermocouple (800~1820) °C	$U=1.2^{\circ}\text{C}$		
				Type R Thermocouple (-20~0) °C	$U=1.6^{\circ}\text{C}$		
				Type R Thermocouple (0~100) °C	$U=1.5^{\circ}\text{C}$		
				Type R Thermocouple (100~1767) °C	$U=1.3^{\circ}\text{C}$		
				Type S Thermocouple (-20~0) °C	$U=1.6^{\circ}\text{C}$		
				Type S Thermocouple (0~200) °C	$U=1.5^{\circ}\text{C}$		
				Type S Thermocouple (200~1400) °C	$U=1.3^{\circ}\text{C}$		
				Type S Thermocouple (1400~1767) °C	$U=1.4^{\circ}\text{C}$		
				Type E Thermocouple (-200~600) °C	$U=0.6^{\circ}\text{C}$		
				Type E Thermocouple (600~1000) °C	$U=0.5^{\circ}\text{C}$		
				Type N Thermocouple (-200~-100) °C	$U=0.9^{\circ}\text{C}$		
				Type N Thermocouple (-100~900) °C	$U=0.8^{\circ}\text{C}$		
				Type N Thermocouple (900~1300) °C	$U=0.6^{\circ}\text{C}$		
				input : (0.001~15) V, (0.001~22)mA	$U_{\text{rel}}=0.030\%$		



No. CNAS L0128

第 4 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
7	*Carbon dioxide incubator	temperature	Carbon dioxide incubator calibration specification SQI/JL-JF-49	(10~50) °C	U=0.3°C		
		Concentration		(1~15) %	U _{rel} =3%		
8	Liquid-in-Glass Thermometers	Temperature	Liquid-in-Glass Thermometers for Working JJG130	(-80~300) °C	U=0.06°C		
9	*Temperature Itinerant Detecting Instrument	Temperature	Calibration Specification for Temperature Itinerant Detecting Instrument JJF 1171	(-60~300) °C	U=0.12°C		
10	Mechanical Thermo-hygrometers	Temperature	Mechanical Thermo-hygrometers JJG205	(5~50) °C	U=0.3°C		
		Humidity		(30~95) %RH	U=1.4%RH		
11	Radiation Thermometers	Temperature	V.R. of the Working Radiation Thermometers JJG 856	(-50~50) °C	U=1.0°C		
				(50~500) °C	U=1.0°C		
				(500~1100) °C	U=2°C		
12	*Thermostatic Baths for Temperature Calibration	Temperature	Measurement and Test Norm of Metrological Characteristics of Thermostatic Baths for Temperature Calibration JJF1030	Temperature fluctuation (-80~300) °C	U=0.003°C		
				Horizontal temperature difference (-80~300) °C	U=0.002°C		
				Maximum temperature difference (-80~0) °C	U=0.006°C		
				Maximum temperature difference (0~100) °C	U=0.003°C		
				Maximum temperature difference (100~300) °C	U=0.006°C		




No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
13	*Digital Temperature-hygrometers	Humidity	Calibration Specification for Digital Temperature-hygrometers JJF 1076	(10~90) %RH	U=1.0%RH		
		Temperature		(5~50) °C	U=0.10°C		
14	*Platinum Copper Resistance	Temperature	V.R. of Industrial Platinum Copper Resistance Thermometer JJG229-2010 JJG 229	Copper Resistance(-50~0) °C	U=0.027 °C		
				Copper Resistance(0 ~ 150) °C	U=0.032°C		
				Platinum(-80~0) °C	U=0.027 °C		
				Platinum(0 ~ 100) °C	U=0.034°C		
				Platinum(100 ~ 300) °C	U=0.042°C		
15	*Thermometers of Clinic Autoclave	Temperature	Calibration Specification for Thermometers of Clinic Autoclave JJF1308	(20~150)°C	U=0.28 °C		
16	Base Metal Thermocouple	Temperature	Calibration Specification for Base Metal Thermocouples JJF 1637	(-40~300) °C	U=0.5°C		
				(300~400) °C	U=1.1°C		
				(400~1100) °C	U=1.2°C		
17	Thermal Imagers	Temperature	C.S. for Thermal Imagers JJF1187	(-15~200) °C	U=1.1 °C		
				(200~450) °C	U=1.2°C		
18	Bimetallic Thermometer	Temperature	Calibration Specification for bimetallic thermometers JJF 1908	(-80~500)°C	U=0.4°C	(- 80 ~ 500) °C bimetallic thermometer	



No. CNAS L0128

第 6 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
19	Filled System thermometer	Temperature	Calibration Specification for Filled system thermometers JJF 1909	(-80~600)℃	U=0.4℃	A round scale steam Filled system with a measuring range of (-30-200) ℃, a round scale gas Filled system with a measuring range of (-80-600) ℃, and a liquid Filled system thermometer with a measuring range of (-40-250) ℃	
20	*Enthalpy Potential Testing Apparatus	Power	Calibration Specification for Enthalpy Potential Testing Apparatus JJF (Jixie)1005	0.01W~80kW	U _{rel} =0.05%		
		Temperature		Environmental: (-40~100) ℃	U=0.22℃		
				Pt100: (-50~100) ℃	U=0.034℃		
				Pt100: (100~150) ℃	U=0.05℃		
				Thermocouple: (-50~300) ℃	U=0.4℃		
		Humidity		(20~95) %RH	U=1.2%RH		



No. CNAS L0128

第 7 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Pressure	ilac-MRA	(-0.1~60)MPa	$U=0.2\%FS$		
		Tachometer		(20~33000) r/min	$U_{rel}=0.02\%$		
		voltage		0.01V~1000V	$U_{rel}=0.05\%$		
		current		0.001A~80A	$U_{rel}=0.05\%$		
		frequency		40Hz~30kHz	$U_{rel}=0.01\%$		
		Power factor		0~1	$U=0.0006$		
		electric energy		0.01V~1000V; 0.001A~80A	$U_{rel}=0.05\%$		
		flow		0.2m ³ /h~40m ³ /h	$U_{rel}=0.2\%$		
		wind speed		0.2m/s~20m/s	$U_{rel}=6\%$		
		Nozzle size		(5~250)mm	$U=0.03mm$		
21	*Temperature Transmitter	Temperature	Calibration Specification of the Temperature Transmitter JJF1183	(-200~800)°C Resistance Thermometer input (without sensor)	$U=(0.3\sim0.6)^{\circ}C$	合格评定国家认可委员会 认可证书专用章	
				(-200°C~1800)°C Thermocouple input (without sensor)	$U=(0.5\sim1.6)^{\circ}C$		
				(-80°C~660)°C Resistance Thermometer input (with sensor)	$U=0.3^{\circ}C$		
				(-40~1100)°C Thermocouple input (with sensor)	$U=1.0^{\circ}C$		



No. CNAS L0128

第 8 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
22	*Box type resistance furnace	Temperature	Calibration specification for box type resistance furnace JJF 1376	(300～1100) °C	<i>U</i> =2.2℃		
23	*Baby Incubator	Temperature	Calibration Specification for Baby Incubator JJF 1260	(20～50) °C	<i>U</i> =0.12℃		
		Humidity		(10～90) %RH	<i>U</i> =2.0%RH		
		Oxygen concentration		30%～40%	<i>U</i> =2.0%		
24	*Disintegration Analyzers	Temperature	Calibration Specification for Disintegration Analyzers JJF 1449	(0～100) °C	<i>U</i> =0.34℃		
		Length		(0～300)mm	<i>U</i> =0.12mm		
		time		1s～30min	<i>U</i> _{rel} =0.09%		
		Disintegration time		(1～520)s	<i>U</i> _{rel} =10.6%		
25	Flow Intergration Meters	Flow	Verification Regulation of Flow Intergration Meters JJG1003	(1～100)% range	<i>U</i> _{rel} =0.04%		
26	Surface Thermometer	Temperature	Calibration Specification for the Surface Thermometers JJF1409	(20～300) °C	<i>U</i> =1.0℃		
				(300～400)℃	<i>U</i> =1.2℃		
27	Gas Float Meter	Flow-rate	Verification Regulation of float meter JJG257	(0.2～30)m³/h	<i>U</i> _{rel} =0.4%		
28	*heat distortion and vicat softening temperature apparatus	temperature	Calibration specification for heat distortion and vicat softening temperature apparatus JJF (zhe) 1051	室温～300℃	<i>U</i> =0.20℃		
		heating rate		(50～120)℃/h	<i>U</i> =0.10℃/h		
		Length		(1.005～10.000) mm	<i>U</i> =2.0 μ m		
		mass		(0.001～500) g	<i>U</i> =0.006g		
				(500～2000) g	<i>U</i> =0.012g		



No. CNAS L0128

第 9 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
				(2000～5000) g	<i>U</i> =0.017g		
29	*Extrusion Plastometer	temperature	Verification Regulation of Extrusion Plastometer JJG 878	(100～400)℃	<i>U</i> =0.2℃		
		Time		(1～600)s	<i>U</i> =0.3s		
		Mass		0.01mg～35kg	<i>U</i> =(1.5～2.2)g		
		melt flow rate		(1.70～1.96)g/10min	<i>U</i> =0.17g/10min		
30	Liquid level measuring device	current	Verification Regulation of theLiquid level measuring device JJG 971	(4～20)mA	<i>U</i> =0.11%FS		
		Level		Horizontal simulation method: (0～2000)mm	<i>U</i> =0.32mm		
				Horizontal simulation method: (2000～ 10000)mm	<i>U</i> _{rel} =0.016%		
				Vertical measurement method: (0～2000)mm	<i>U</i> =2mm		
		Pressure		(0～100)kPa	<i>U</i> =0.09kPa		
		voltage		(0.1～50)V	<i>U</i> =0.11%FS		
31	*Natural ventilation thermal aging test oven	Temperature	Verification procedure for test equipment of rubber plastic wire and cable - Part6: Natural ventilation thermal aging test oven JB/T 4278.6	(-40～600) ℃	<i>U</i> =0.5℃		
		ventilation rate		(8～20)/h	<i>U</i> _{rel} =2%		
32	Precision dew point hygrometer	Dew point temperature	Verification regulation of precision dew point meter JJG499	(-70～20)℃CDP	<i>U</i> =0.20 ℃CDP		
33	Temperature Block Calibrator	Temperature	Calibration Guideline of the Temperature Block	(-80～150) ℃	<i>U</i> =0.12℃		



No. CNAS L0128

第 10 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
			Calibrators JJF1257	(150~420) °C	$U=0.44^{\circ}\text{C}$		
				(420~660) °C	$U=0.60^{\circ}\text{C}$		
				(660~700) °C	$U=0.66^{\circ}\text{C}$		
				(700~1100) °C	$U=1.3^{\circ}\text{C}$		
34	*Digital Thermometer	Temperature	The Working Digital Thermometer JJF(lu)90 JJF(lu)90	(-80~300) °C	$U=0.06^{\circ}\text{C}$		
				(300~600) °C	$U=0.10^{\circ}\text{C}$		
				(600~1100) °C	$U=1.0^{\circ}\text{C}$		
35	*Fabric shrinkage rate testing machine	Temperature	Calibration Specification for Fabric Shrinkage Testers JJF (Fabric) 052	(25~98) °C	$U=0.6^{\circ}\text{C}$		
		Rotating speed		(50~1200) r/min	$U_{\text{rel}}=0.28\%$		
		time		(0~30) min	$U=0.3\text{s}$		
		Level		(100~300) mm	$U=0.7\text{mm}$		
36	*salt fog test chamber	Temperature	Calibration specification for salt fog test chamber JJF (Zhe) 1125	(5~70) °C	$U=0.4^{\circ}\text{C}$		
		Salt spray deposition rate		(1.0~2.0)mL/ (80cm ² •h)	$U=0.3\text{mL}/(80\text{cm}^2\cdot\text{h})$		
37	*Drug Dissolution Meter	Temperature	Calibration Specification for Drug Dissolution Meter SQI/JL-JF-62	(0~50) °C	$U=0.20^{\circ}\text{C}$		
		speed		(20~500) r/min	$U_{\text{rel}}=0.6\%$		
		Coaxiality		(0~5) mm	$U=0.09\text{mm}$		
		Shaft oscillation		(0~5) mm	$U=0.11\text{mm}$		
		Cup swing		(0~5) mm	$U=0.08\text{mm}$		



No. CNAS L0128

第 11 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		Depth		(0~30) mm	<i>U</i> =0.21mm		
		Verticality of Dissolution Cup		(-2~92)°	<i>U</i> =0.12°		
		Verticality of Rotary Shaft		(-2~92)°	<i>U</i> =0.12°		
		Sampling Pump Speed		(0~10) mL/min	<i>U</i> =0.1mL/min		
38	Thermal Mass Gas Flowmeters	Flow	Thermal Mass Gas Flowmeters JJG1132	(5~50000) mL/min	<i>U</i> _{rel} =0.44%		
39	Temperature Switchs	Temperature	Calibration Specification for Temperature Parameters of Temperature Switchs JJF 1632	(-30~300)°C	Action Temperature: <i>U</i> =0.3 °C； On-off Temperature Difference: <i>U</i> =0.20°C		
40	*Washing fastness testing machine	Temperature	Calibration Specification for Washing Fastness Testing Machine JJF (Textile) 026	(0~150) °C (deviation)	<i>U</i> =0.7°C		
				(0~150) °C (Indication error)	<i>U</i> =0.7°C		
				(0~150) °C (Volatility)	<i>U</i> =0.6°C		
		Rotate speed		40r/min	<i>U</i> =0.20r/min		
		Time		30min、45min	<i>U</i> =2.0s		
		capacity		550mL	<i>U</i> =4mL		
		length		125mm (Cup depth)	<i>U</i> =1.0mm		
				45mm (The distance from the shaft core to the bottom of the test cup)	<i>U</i> =1.0mm		



No. CNAS L0128

第 12 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
41	*Temperature and humidity standard chambers	Temperature	Calibration specification for Temperature and humidity standard chambers JJF 1564	Temperature Uniformity: (5~50) °C, Temperature Fluctuation: (5~50) °C, Temperature rate of change: (0~5) °C/min	Temperature Uniformity: $U=0.05$ °C; Temperature Fluctuation: $U=0.02$ °C; Temperature rate of change: $U=0.03$ °C/min		
		Humidity		Humidity Uniformity: (10~90) %RH, Humidity Fluctuation: (10~90) %RH, Humidity rate of change: (0~10) %RH/min	Humidity Uniformity: $U=0.5\%$ RH; Humidity Fluctuation: $U=0.2\%$ RH; Humidity rate of change: $U=0.3\%$ RH/min.		
42	*Infrared Thermometers for Measurement of Human Temperature	Temperature	Calibration specification of Infrared Thermometers for Measurement of Human Temperature JJF 1107	(22~40) °C	$U=0.10$ °C		
43	*Illumination test chamber	Illuminance	Calibration Specification for the Equipment of the Environmental Testing for Differential Pressure and Wind Speed and illuminance SQI/JL-JF-47	(1~3000) lx	$U=49$ lx		
44	Copper/Copper-Nickel Thermocouple	Temperature	V.R. of Working Copper / Copper-Nickel Thermocouple JJG368	(-80~300) °C	$U=0.3$ °C		



No. CNAS L0128

第 13 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
45	Infrared devices for instant screening of human skin temperature	Temperature	Calibration Specification of Infrared Thermometers for Measurement of Human Temperature JJF 1107	(22～40) °C	<i>U</i> =0.10 °C		
46	Sheathed Thermocouple	Temperature	Calibration Specification for Sheathed Thermocouples JJF 1262	(-40～300) °C	<i>U</i> =0.3 °C		
				(300～400) °C	<i>U</i> =1.1 °C		
				(400～1100) °C	<i>U</i> =1.2 °C		
47	Temperature Data Acquisition Instruments	temperature	Calibration Specification of Temperature Data Acquisition Instruments JJF 1366	(-80～500) °C	<i>U</i> =0.05 °C		
48	Temperature Calibration Devices for Polymerase Chain Reaction Analyzers	Temperature	Calibration Specification of Temperature Calibration Devices for Polymerase Chain Reaction Analyzers JJF 1821	(0～120) °C	<i>U</i> =0.04 °C		
49	*Dissolution Testers	Temperature	Calibration Specification for Dissolution Testers JJF(zhe)1096	(0～50) °C	<i>U</i> =0.10 °C		
		Revolution Speed		(50～200) rpm	<i>U</i> _{rel} =0.6%		
		Length		coaxiality:(0～5) mm	<i>U</i> =0.06 mm		
				swing scope:(0～5) mm	<i>U</i> =0.08 mm		
50	Thermistor thermometers	Temperature	Calibration Specification of Thermistor Thermometers JJF 1379	(-50～200) °C	resolution0.001 °C: <i>U</i> =0.030 °C , resolution0.01 °C: <i>U</i> =0.06 °C , resolution0.1 °C: <i>U</i> =0.10 °C		



在线扫码获取验证

No. CNAS L0128

第 14 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
51	Resistance-capacitance dew point hygrometer	Dew point temperature	Calibration specification for dew point hygrometer with resistance-capacitance method JJF 1272	$(-70\sim 20)^{\circ}\text{C DP}$	$U=0.4^{\circ}\text{C DP}$		
52	*Temperature Indicators	Temperature	Calibration Specification for Temperature Indicators JJF1664	Digit thermal resistor($-200\sim 600$) $^{\circ}\text{C}$	$U=(0.10\sim 0.3)^{\circ}\text{C}$		
				Digit Thermocouple ($-200\sim 1800$) $^{\circ}\text{C}$	$U=(0.5\sim 1.6)^{\circ}\text{C}$		
		Temperature		Analog thermal resistor($-200\sim 600$) $^{\circ}\text{C}$	$U=(0.3\sim 0.5)^{\circ}\text{C}$		
				Analog Thermocouple ($-200\sim 1800$) $^{\circ}\text{C}$	$U=(0.6\sim 1.6)^{\circ}\text{C}$		
53	*Divided Flow Humidity Generator	Temperature	Calibration Specification for Divided Flow Humidity Generator JJF(military)42	$(5\sim 50)^{\circ}\text{C}$	$U=0.20^{\circ}\text{C}$		
		Dew point temperature		$(-70\sim 20)^{\circ}\text{C DP}$	$U=0.34^{\circ}\text{C DP}$		
		Relative humidity		$(10\sim 90)\% \text{RH}$	$U=1.6\% \text{RH}$		
54	*Liquid constant temperature testing equipment	Temperature	Measurement Specification for Temperature Performance of Liquid Constant Temperature Testing Equipment JJF 2019	$(-80\sim 300)^{\circ}\text{C}$	$U=(0.05\sim 0.3)^{\circ}\text{C}$		
55	*Short Base Metal Thermocouples	Temperature	Calibration Specification for Short Base Metal Thermocouples JJF 1991	$(-40\sim 300)^{\circ}\text{C}$	$U=0.5^{\circ}\text{C}$	合格评定国家认可委员会 认可证书专用章	
				$(300\sim 400)^{\circ}\text{C}$	$U=1.1^{\circ}\text{C}$		
				$(400\sim 1000)^{\circ}\text{C}$	$U=1.2^{\circ}\text{C}$		
56	*Stream Sterilizer	Temperature	Calibration Specification for Temperature and Pressure parameters of Stream	$(20\sim 150)^{\circ}\text{C}$	$U=0.15^{\circ}\text{C}$		
		Pressure		$(0\sim 400)\text{ kPa}$	$U=1.7\text{ kPa}$		



No. CNAS L0128

第 15 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Time	Sterilizer JJF (Shanghai) 60	(10~3600) s	U=0.4 s		
57	Distributed fiber optic thermometer	Temperature	Calibration specification for distributed fiber optic thermometers JJF 1630	(-20~100) °C	U=0.20℃		
58	Passive Medical Cold Boxes	Temperature	Calibration Specification for Temperature Parameter of Passive Medical Cold Boxes JJF 1676	(-20~20) °C	U=0.20℃		
59	Calibration Specification for Surface Temperature Sources	Temperature	Calibration Specification for Surface Temperature Sources JJF (Lu) 137	Stability: (30~400) °C	U=0.12℃		
				Uniformity: (30~400) °C	U=0.3℃		
				Indication error: (30~400) °C	U=0.26℃		
60	Thermometers of WBGT-index Meters	Temperature	Calibration Specification for Thermometers of WBGT-index Meters JJF 1407	(5~120) °C	U=0.3℃		
61	*Large Steam Sterilizers	temperature	Calibration Specification for Temperature, Pressure and Time Parameters of Large Steam Sterilizers JJF 2088	(100~150) °C	U=0.14℃		
		time		(1~3600) s	U=1s		
		pressure		(70~400) kPa.abs	U=1.2kPa		
62	*Forced air heat aging test oven	Temperature	Verification procedure for test equipment of rubber plastic wire and cable -Part 13:Forced air heat aging test oven JB/T 4278.13	(-80~300) °C	U=0.5℃		
		Air change rate		(0~200) time/h	U=1 次/h		
		Time		1ms~3600s	U=0.6s		
二、Mechanics							
1	*Moisture Meters	Quality	Verification Regulation of	(0~1)kg	U=(0.2~10)mg		

No. CNAS L0128

第 16 页 共 143 页



The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Moisture	Thermogravimetric Moisture Meters JJG 658	(0~100)%	$U=0.09\%$		
2	*Electronic Balance	Quality	Verification Regulation of Electronic Balance JJG 1036, Calibration Specification of Electronic Balances JJF1847	(1~100) mg	$U=(0.004\sim0.005)$ mg		
				100mg~1g	$U=(0.005\sim0.008)$ mg		
				(1~100) g	$U=(0.008\sim0.06)$ mg		
				100g~1kg	$U=(0.06\sim2)$ mg		
				(1~100) kg	$U=2\text{mg}\sim0.5\text{g}$		
				(100~200) kg	$U=(0.5\sim5)$ g		
3	*Flow Meters	Flow	Specification for on-line flow Calibration by master meter method SQI/JL-JF-54	(0.2~40) m ³ /h, DN (8-50) mm	$U_{\text{rel}}=0.20\%$	Accredited only for Water Medium	
4	Tachometer	rotate speed	Verification Regulation of Tachometer JJG 105	(20~60000) r/min	$U_{\text{rel}}=0.02\%$		
5	*Weights	Quality	Verification Regulation of Weights JJG 99	F ₁ Class, 1mg~500mg	$U=(0.004\sim0.010)\text{mg}$	Field calibration is not carried out for grade F2 and above	
				F ₁ Class, 1g~500g	$U=(0.02\sim0.4)\text{mg}$		
				F ₂ Class, 1mg~500mg	$U=(0.02\sim0.03)\text{mg}$		
				F ₁ Class, 1kg~5kg	$U=(0.6\sim4)\text{mg}$		
				F ₂ Class, 1g~500g	$U=(0.03\sim0.6)$ mg		
				F ₂ Class, 1kg~5kg	$U=(2\sim8)\text{mg}$		
				M ₁ Class, 1mg~500mg	$U=(0.02\sim0.08)\text{mg}$		
				M ₁ Class, 1g~500g	$U=(0.3\sim3)\text{mg}$		



No. CNAS L0128

第 17 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				M ₁ Class: 1kg~25kg	$U=5\text{mg}\sim 0.2\text{g}$		
6	*Electromagnetic flowmeter	Flow	Calibration Specification for Electromagnetic Flowmeters JJF (Su) 228	(8~40)m ³ /h, DN50	$U_{\text{rel}}=3.6\%$		
				(30~400)m ³ /h, DN(100~150)	$U_{\text{rel}}=2.8\%$		
				(80~400)m ³ /h, DN200	$U_{\text{rel}}=2.5\%$		
				(240~1200)m ³ /h, DN(250~300)	$U_{\text{rel}}=2.4\%$		
7	*Mass flowmeter	flow	Online calibration specification for Mass flowmeters by Mass method SQI/JL-JF-43	(0.01~40) m ³ /h, DN(4-50)	$U_{\text{rel}}=0.4\%$	Accredited only for Water Medium	
8	*Mechanical Balance	Quality	Verification Regulation of Mechanical Balance JJG 98	(0~20)g $e=0.01\text{mg}$	$U=0.03\text{mg}$		
				(0~200)g $e=0.1\text{mg}$	$U=0.08\text{mg}$		
				(1~5)kg $e=1\text{mg}\sim 5\text{mg}$	$U=(0.8\sim 1.6)\text{mg}$		
				(20~200)kg $e=(10\sim 100)\text{mg}$	$U=(0.15\sim 2.4)\text{g}$		
9	Force Transducers	force	Verification Regulation of Force Transducers JJG 391	0.1N~1MN	$U_{\text{rel}}=0.06\%$		
10	*Falling Weight Impact Testing Machines	speed	Calibration Specification for Falling Weight Impact Testing Machines JJF 1445	(0.1~30)m/s	$U_{\text{rel}}=0.8\%$		
		height		(0.01~5) m	$U=2\text{mm}$		
		Quality		(0.01~35) kg	$U=1\text{g}$		
		Radius		(0.5~500) mm	$U=0.06\text{mm}$		
		energy loss		(0.01~100) %	$U_{\text{rel}}=1.5\%$		



No. CNAS L0128

第 18 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
11	Standard Glass Hydrometers	Density	Verification Regulation of Standard Glass Hydrometers JJG 86	(650~1500) kg/m ³	$U=0.20 \text{ kg/m}^3$		
12	Vibration Displacement Transducer	displacement	Verification Regulation of Vibration Displacement Transducer JJG 644	0.002mm~1000mm (20Hz~2000Hz)	$U_{\text{rel}}=1.4\%$		
		Sensitivity		0.002mm~1000mm (0.2Hz~10kHz)	$U_{\text{rel}}=1.4\%$		
13	Working Glass Hydrometers	Density	Verification Regulation of Working Glass Hydrometers JJG 42	(650~2000) kg/m ³	$U=0.4 \text{ kg/m}^3$		
		Soil Degree		(-5~50) S°	$U=0.20 \text{ S}^\circ$		
		alcohol concentration		(0~100) %	$U=0.4\%$		
14	*Falling Body Type Shock Testing Machine	acceleration	Verification Regulation of Shock and Bump Testing Machines JJG 1174	(100~5000) m/s ²	$U_{\text{rel}}=4\%$		
		Pulse width		(0.05~240) ms	$U_{\text{rel}}=0.8\%$		
15	Working Glass Container	Capacity	Verification Regulation of Working Glass Container JJG196	(0.1~100)mL	$U=(0.002\sim0.019)\text{mL}$		
				(100~500)mL	$U=(0.019\sim0.08)\text{mL}$		
				(500~2000) mL	$U=(0.08\sim0.26)\text{mL}$		
16	*Portable Vibration Calibrator	acceleration	Verification Regulation of Portable Vibration Calibrator JJG 1062	(0.5~500) m/s ²	$U_{\text{rel}}=3\%$		
		displacement		(1~25)mm	$U_{\text{rel}}=3\%$		
		Frequency		(10~2000)Hz	$U=0.6\text{Hz}$		
17	Locomotive Pipett	Capacity	Verification Regulation of Locomotive Pipette JJG646	(0.1~50) μL	$U=(0.02\sim0.2) \mu\text{L}$		
				(50~500) μL	$U=(0.2\sim0.5) \mu\text{L}$		
				(500~10000) μL	$U=(0.5\sim2.2) \mu\text{L}$		



No. CNAS L0128

第 19 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
18	*Laboratory Oscillation-type Liquid Density Meters	Density	Laboratory Oscillation-type Liquid Density Meters JJG 1058	(0.65~1.50)g/cm ³ (1.50~2.00)g/cm ³	$U=0.0002 \text{ g/cm}^3$ $U=0.0003 \text{ g/cm}^3$	/	
19	*Equipment of Power Measuring	torque	Verification Regulation of Equipment of Power Measuring JJG 653	(0.01~108.2) Nm (108.2~3000) Nm (3000~10000) Nm	$U_{\text{rel}}=0.28\%$ $U_{\text{rel}}=0.27\%$ $U_{\text{rel}}=0.30\%$		
		Rotate Speed		(20~20000) r/min	$U_{\text{rel}}=0.17\%$		
20	*Mass Comparators	Mass	Calibration Specification for Mass Comparators JJF 1326	repeatability: (0~1000)g	$U=(2 \times 10^{-3} \sim 2)\text{mg}$	/	
		Mass		Eccentric load error: (0~1000)g	$U=(3 \times 10^{-3} \sim 3)\text{mg}$		
		Mass		Partial indication error: (0~1000)g	$U=(5 \times 10^{-3} \sim 4)\text{mg}$		
21	Shore D Durometer	Hardness	Verification Regulation of Shore D Durometer JJG 1039	0~100HD	$U=0.14\text{HD}$		
		Force		(0.1~50) N	$U_{\text{rel}}=0.2\%$		
		Length		(0~10)mm	$U=6 \mu\text{m}$		
		Angle		(0~90)°	$U=0.03^\circ$		
22	Electromagnetic Velocity Transducer	speed	Verification Regulation of Electromagnetic Velocity Transducer JJG 134	(0.01~20)m/s (20Hz~2000Hz)	$U_{\text{rel}}=1.1\%$		
		Sensitivity		1cm/s~10cm/s (20Hz~2000Hz)	$U_{\text{rel}}=1.1\%$		
23	*Metallic Rockwell Hardness Testers	Hardness	Standard Test Methods for Rockwell Hardness and Rockwell Superficial	(80~88)HRA (85~95)HRB	$U=0.6\text{HRA}$ $U=0.8\text{HRBw}$		



No. CNAS L0128

第 20 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Hardness of Metallic Materials ASTM E18, Verification Regulation of Metallic Rockwell Hardness Testers (Scales A, B, C, D, E, F, G, H, K, N, T) JJG 112		(60~70)HRC	$U=0.6\text{HRC}$		
				(35~55)HRC	$U=0.6\text{HRC}$		
				(20~30)HRC	$U=0.7\text{HRC}$		
				(89~91)HR15N	$U=0.7\text{HR15N}$		
				(74~80)HR30N	$U=0.7\text{HR30N}$		
				(42~54)HR30N	$U=0.9\text{HR30N}$		
				(32~61)HR45N	$U=0.9\text{HR45N}$		
				(88~93)HR15T	$U=1.2\text{HR15TW}$		
				(70~82)HR30T	$U=1.2\text{HR30TW}$		
				(55~72)HR45T	$U=1.7\text{HR45TW}$		
		Force		(10kgf~150kgf) (100~1500) N	$U_{\text{rel}}=0.3\%$		
		Depth		(0~1) mm	$U=0.6\text{ }\mu\text{m}$		
		Angle		0~180°	$U=0.03^\circ$		
		Size		(0~1) mm	$U=6\text{ }\mu\text{m}$		
		Time		(1~60) s	$U_{\text{rel}}=1.0\%$		
24	Standard Dynamometers	force	Verification Regulation of Standard Dynamometers JJG 144	0.1N~10N	$U_{\text{rel}}=0.06\%$		
				10N~10kN	$U_{\text{rel}}=0.03\%$		
				10kN~1MN	$U_{\text{rel}}=0.06\%$		
25	*Calibration	torque	Verification Regulation of	(0.02~10) Nm	$U_{\text{rel}}=0.8\%$		

No. CNAS L0128

第 21 页 共 143 页



The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
	Instrument for Torque Wrenches		Calibration Instrument for Torque Wrenches JJG 797	(10~3000) Nm	$U_{rel}=0.9\%$		
26	*Torsion Testing Machines	torque angle	Verification Regulation of Torsion Testing Machines JJG 269	(0.02~3000) Nm	$U_{rel}=0.5\%$		
				(3000~10000) Nm	$U_{rel}=0.7\%$		
				$0.1^{\circ} \sim 1080^{\circ}$	$U_{rel}=0.5\%$		
27	*Analogue Indicating Weighing Instruments	Weight	Verification Regulation of Analogue Indicating Weighing Instruments JJG 13	(0.2~8) kg	$U=(3.3\sim 3.7) g$		
28	*Non-self-indicating Weighing Instruments	Weight	Verification Regulation of Non-self-indicating Weighing Instruments JJG 14	(0.04~3)kg	$U=(0.5 \sim 0.6)g$		
				(0.1~5)kg	$U=(0.5 \sim 0.6)g$		
				(0.1~10)kg	$U=(0.5 \sim 0.8)g$		
				(1~100)kg	$U=(6\sim 11)g$		
				(4~300)kg	$U=(21\sim 38)g$		
				(4~500)kg	$U=(21\sim 59)g$		
				(10~1000)kg	$U=(0.05\sim 0.12)kg$		
				(20~2000)kg	$U=(0.05\sim 0.24)kg$		
				(20~3000)kg	$U=(0.05\sim 0.42)kg$		
				(40~5000)kg	$U=(0.1\sim 0.6)kg$		
				(5~10000)kg	$U=(0.2\sim 1.4)kg$		
				(200~30000) kg	$U=(0.4\sim 4.5)kg$		



No. CNAS L0128

第 22 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				(400~60000) kg	$U=U=(2.9\sim 12)\text{kg}$		
29	*Digital Indicating Weighing Instruments	Weight	Verification Regulation of Digital Indicating Weighing Instruments JJG 539	(2~300)g	$U=(10\sim 30)\text{mg}$		
				(2~500)g	$U=(10\sim 40)\text{mg}$		
				(2~1000)g	$U=(10\sim 60)\text{mg}$		
				(0.02~3)kg	$U=(0.3\sim 0.4)\text{g}$		
				(0.04~6)kg	$U=(0.3\sim 0.9)\text{g}$		
				(0.1~15)kg	$U=(1.2\sim 1.5)\text{g}$		
				(0.2~30)kg	$U=(3\sim 4)\text{g}$		
				(0.4~60)kg	$U=(6\sim 7)\text{g}$		
				(1~150)kg	$U=(6\sim 10)\text{g}$		
				(2~300)kg	$U=(20\sim 40)\text{g}$		
				(2~500)kg	$U=(20\sim 50)\text{g}$		
				(4~1000)kg	$U=(0.07\sim 0.13)\text{kg}$		
				(10~2000)kg	$U=(0.1\sim 0.4)\text{kg}$		
				(20~3000)kg	$U=(0.1\sim 0.5)\text{kg}$		
				(20~5000)kg	$U=(0.1\sim 0.6)\text{kg}$		
				(200~10000) kg	$U=(2\sim 3)\text{kg}$		
				(200~20000) kg	$U=(2\sim 4)\text{kg}$		
				(200~30000) kg	$U=(2\sim 5)\text{kg}$		



No. CNAS L0128

第 23 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				(400~60000) kg	U=(3~11)kg		
				(400~80000) kg	U=(3~12)kg		
				(400~100000) kg	U=(3~14)kg		
30	*Automatic Gravimetric Filling Instruments	Weight	Verification Regulation of Automatic Gravimetric Filling Instruments JJG 564	X (0.1) Class, (0.01~200) kg	U _{rel} =(0.25~0.02)%		
				X (0.2) Class, (0.01~200) kg	U _{rel} =(0.25~0.02)%		
				X (0.5) Class, (0.01~200) kg	U _{rel} =(0.26~0.03)%		
				X (1) Class, (0.01~200) kg	U _{rel} =(0.27~0.06)%		
				X (2) Class, (0.01~200) kg	U _{rel} =(0.30~0.12)%		
31	*Quantitative Filling Machine for Liquid Material	Weight	Verification Regulation of Quantitative Filling Machine for Liquid Material JJG 687	(50~32000) g	U _{rel} =(1.2~0.01)%	use the weighing method to verify the volume of the filling machine	
		Volume		(32~150) kg	U _{rel} =(0.20~0.08)%		
				(50~32000) mL	U _{rel} = (1.2~0.04) %		
				(32~150) L	U _{rel} = (0.28~0.10) %		
32	*Precise Pressure Gauges	pressure	Verification Ragulation of Elastic Element Precise Pressure Gauges and Vacuum Gauges JJG49	(-0.1~0.01) MPa	U=(0.08~0.13)%FS		
				(0.01~250) MPa	U= (0.08~0.13) %FS		
33	*Elastic Element Pressure Gauges, Pressure-Vacuum Gauges and	pressure	Verification Ragulation of Elastic Element Pressure Gauges,Pressure-Wacuum Gauges and Vacuum Gauges for General Use JJG52	(-0.1~0.01) MPa	U=0.34%FS		

No. CNAS L0128


第 24 页 共 143 页



The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	Vacuum Gauges for General Use			(0.01~250) MPa	$U=0.34\%FS$		
34	*Digital Pressure Gauges	pressure	Verification Regulation of Digital Pressure Gauges JJG875	(-0.1~0.01) MPa	$U=(0.01\sim0.07)\%FS$		
				(0.01~250) MPa	$U=(0.01\sim0.07)\%FS$		
35	*Pressure Regulators with Bourdon Tube Pressure Gauge	pressure	Calibration Specification for Pressure Regulators with Bourdon Tube Pressure Gauge JJF1328	0.005MPa~25MPa	$U_{rel}=1\%$		
36	*Pressure Controllers	pressure	Verification Regulation of Pressure Controllers JJG544	(-0.1~0.01) MPa	$U=0.12\%FS$		
				(0.01~60) MPa	$U=0.12\%FS$		
37	*Pressure Transmitter	pressure	Verification Regulation of the Pressure Transmitter JJG882	(-0.1~0.01) MPa	$U=0.02\%FS$		
				(0.01~250) MPa	$U=0.02\%FS$		
38	Industrial Thermal Conductivity Vacuum Gauges	Vacuum	Calibration Specification of Industrial Thermal Conductivity Vacuum Gauges JJF1050	$(2.0\times10^{-1}\sim1.0\times10^5)$ Pa	$U_{rel}=4\%$		
39	Channel Reference Leaks	Flow	calibration specification of channel reference leaks SQI/JL-JF-40	(0.1~60) mL/min	$U_{rel}=3\%$		
40	Working Dynamometers	force	Verification Regulation of Working Dynamometers JJG455	1.96mN~2000kN	$U_{rel}=0.2\%$		
41	*Dedicated work force measuring machine	force	Calibration specification for dedicated working dynamometer JJF 1134	0.1N~2000kN	$U_{rel}=0.4\%$		
42	*electronic universal testing machine	force	Verification Regulation of Electronic Universal Testing Machine JJG475	9.8 μ N~500kN	$U_{rel}=0.18\%$		



№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		concentricity Displacement speed		500kN～2000kN	<i>U</i> _{rel} =0.50%		
				2000kN～10000kN	<i>U</i> _{rel} =0.50%		
				(0.1～40) mm	<i>U</i> =0.021mm		
				(0～1000)mm	<i>U</i> =(8～14) μ m		
				(0.001～100)mm/s	<i>U</i> _{rel} =0.2%		
43	*Tension , Compression and Universal Testing Machine	force	Verification Regulation of Tension , Compression and Universal Testing Machines JJG139	9.8 μ N～500kN	<i>U</i> _{rel} =0.18%		
				500kN～2000kN	<i>U</i> _{rel} =0.50%		
				2000kN～10000kN	<i>U</i> _{rel} =0.50%		
		Displacement		0.1mm～0.3mm	<i>U</i> =0.72 μ m		
				0.3mm～800mm	<i>U</i> _{rel} =0.2%		
				Concentricity	(0.1～40) mm		
44	*Material Testing Machine	force	Standard Practices for Force Calibration and Verification of Testing Machines ASTM E4	0.1N～2000kN	<i>U</i> _{rel} =0.4%		
45	*Material Testing Machine	displacement	Standard Practices for Verification of Displacement Measuring Systems and Devices Used in Material Testing Machines ASTM E2309/E2309M	(0.001～750)mm	<i>U</i> =0.06mm		
				(750～1000)mm	<i>U</i> =0.78mm		
				(1000～2000)mm	<i>U</i> =0.80mm		



No. CNAS L0128

第 26 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
46	*Testing Machine(Alignment)	Alignment	Standard Practice for Verification of Testing Frame and Specimen Alignment Under Tensile and Compressive Axial Force Application ASTM E1012	(0.001~2450)μm/m	<i>U</i> _{rel} =1.4%		
47	*material testing machine	speed	Standard Practices for Verification of Speed for Material Testing Machines ASTM E2658	(0.001~100)mm/s	<i>U</i> _{rel} =0.2%		
48	*Electric-hydraulic Servo Universal Test Machine	force	Verification regulation of electro hydraulic servo universal testing machine JJG 1063	9.8 μ N~500kN	<i>U</i> _{rel} =0.18%		
				500kN~2000kN	<i>U</i> _{rel} =0.50%		
				2000kN~3000kN	<i>U</i> _{rel} =0.50%		
		concentricity		(0.1~40) mm	<i>U</i> =0.021mm		
		Displacement		0.1mm~0.3mm	<i>U</i> =0.72 μ m		
				0.3mm~800mm	<i>U</i> _{rel} =0.2%		
49	*Static uniaxial testing machine	force	Metallic materials-Verification of static uniaxial testing machines-Part1 Tensioncompression testing machines-Verification and calibration of the force-measuring system ISO 7500-1	1.96mN~2000kN	<i>U</i> _{rel} =0.4%		
		concentricity		(0.1~40) mm	<i>U</i> =0.021mm		
50	*high temperature creep and endurance strength	force	Verification regulation of high temperature creep and endurance strength testing	0.1N~2000kN	<i>U</i> _{rel} =0.4%		
		Temperature		(-80~200)℃	<i>U</i> =0.3℃		



No. CNAS L0128

第 27 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
	testing machine		machine JJG 276	(200~300)°C	$U=0.6^{\circ}\text{C}$		
				(300~500)°C	$U=1.2^{\circ}\text{C}$		
				(500~1100)°C	$U=2.1^{\circ}\text{C}$		
		concentricity		(0.1~40) mm	$U=0.021\text{mm}$		
		Extensometer		(0.01~50) mm	$U_{\text{rel}}=0.2\%$		
		Time		(1~3600) s	$U=0.13\text{s}$		
51	*Fatigue Testing Machines	force	Axial Force Fatigue Testing Machines JJG 556, Standard Practice for Verification of Constant Amplitude Dynamic Forces in an Axial Fatigue Testing System ASTM E467	(0.5N~100kN)	$U_{\text{rel}}=0.4\%$		
		concentricity		(0.1~40) mm	$U=0.021\text{mm}$		
52	Torque Wrenches	torque	Verification Regulation of Torque Wrenches JJG707	0.06Nm~6Nm	$U_{\text{rel}}=2\%$		
				2Nm~3000Nm	$U_{\text{rel}}=0.9\%$		
53	*Metal Brielle Hardness Tester	Hardness	Verification Regulation of Metal Brielle Hardness Tester JJG150	(50~125)HBW	$U_{\text{rel}}=1.9\%$		
				(125~225)HBW	$U_{\text{rel}}=1.4\%$		
				(225~650)HBW	$U_{\text{rel}}=1.2\%$		
54	*Metal Rockwell Hardness Tests	Hardness	Verification Regulation of Metal Rockwell Hardness Tests(A、B、C、D、E、F、G、H、R、N、T Scale) JJG112	(20~95)HRA, HRB, HRC	$U=(0.6\sim0.8)\text{HR}$		
				(32~91) HRN (34~93) HRT	$U=(0.7\sim0.9)\text{HRN}$ $U=(1.2\sim1.7)\text{HRTw}$		
		Force		(9.8~1571)N	$U_{\text{rel}}=0.3\%$		
		Length		(0~4) μm	$U=0.6\mu\text{m}$		



No. CNAS L0128

第 28 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Angle		(0~150) °	$U=0.03^{\circ}$		
55	Leeb Hardness Tester	Hardness	Verification Regulation of Leeb Hardness Tester JJG747	(330~890) HLD	$U=8\text{HLD}$		
56	*Metal Vickers Hardness Tester	Hardness	Verification Regulation of Metal Vickers Hardness Tester JJG151	HV5~HV100, (50~225)HV	$U_{\text{rel}}=2.7\%$		
				HV5~HV100, (225~1000)HV	$U_{\text{rel}}=2.0\%$		
				HV0.2~HV5, (50~225)HV	$U_{\text{rel}}=5.3\%$		
				HV0.2~HV5, (225~1000)HV	$U_{\text{rel}}=4.0\%$		
				HV0.05~HV0.2, (50~225)HV	$U_{\text{rel}}=5.1\%$		
				HV0.05~HV0.2, (225~1000)HV	$U_{\text{rel}}=6.0\%$		
57	Shore A Durometers	Hardness	Verification Regulation of Shore A Durometers JJG304	(0~100) HA	$U=0.2\text{HA}$		
		Force		(0.1~50) N	$U_{\text{rel}}=0.2\%$		
		Length		(0~10)mm	$U=5\text{ }\mu\text{m}$		
		Angle		(10~40)°	$U=0.04^{\circ}$		
58	*Hydraulic Vibration Testing System	Frequency	Verification Regulation of Hydraulic Vibration Testing System JJG 638	(0.1~2000) Hz	$U_{\text{rel}}=0.012\%$		
		Displacement		(0.1~50) mm (0.1~10k) Hz	$U_{\text{rel}}=3\%$		
		acceleration		(0.05~1000) m/s ² (0.1~10k) Hz	$U_{\text{rel}}=3\%$		
		Velocity		(0.01~20) m/s (0.1~10k) Hz	$U_{\text{rel}}=1.1\%$		



No. CNAS L0128

第 29 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
59	*Metallic Webster Hardness Testing Machines	Hardness	Verification Regulation for Metallic Webster Hardness Testing Machines JJG 944	(5~18) HW	$U=0.3HW$		
		Length		(0.02~1) mm	$U=4\text{ }\mu\text{m}$		
		Angle		(0~180) °	$U=0.03^\circ$		
		Size		(0.02~50) mm	$U=6\text{ }\mu\text{m}$		
60	*Mechanical Vibration Generator for Testing	Frequency	Verification Regulation of Mechanical Vibration Generator for Testing JJG189	(10~200)Hz	$U=0.6\text{Hz}$		
		Displacement		(1~50)mm (0.1~200)Hz	$U_{\text{rel}}=3\%$		
		acceleration		(0.5~1000) m/s ² (0.1~200)Hz	$U_{\text{rel}}=3\%$		
61	*Microhardness Testing Machine in International Rubber Hardness Degree	Length	Verification Regulation for Microhardness Testing Machine in International Rubber Hardness Degree JJG 898	(0~10) mm	$U=6\text{ }\mu\text{m}$		
		Hardness		(30~85)IRHD	$U=1.3\text{IRHD}$		
		Force		0.2mN~10N	$U_{\text{rel}}=0.2\%$		
62	*Plastic Ball Indentation Hardness Testing Machine	Force	Verification Regulation for Plastic Ball Indentation Hardness Testing Machine JJG 369	(4.9~961) N	$U_{\text{rel}}=0.3\%$		
		Size		(0~15) mm	$U=6\text{ }\mu\text{m}$		
		Depth		(0~1) mm	$U=0.6\text{ }\mu\text{m}$		
		Time		(1~120) s	$U_{\text{rel}}=1.0\%$		
63	Piezoelectric Accelerometer	acceleration	Verification Regulation of Piezoelectric Accelerometer JJG233	(0.02~100)m/s ² (0.2~1k) Hz	$U_{\text{rel}}=1.1\%$		
				(2~100)m/s ² (1k~2k) Hz	$U_{\text{rel}}=2.0\%$		
				(100~20000)m/s ² (2k~10k) Hz	$U_{\text{rel}}=3.0\%$		



No. CNAS L0128

第 30 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				(20000~100000)m/s ² (5k~10k) Hz	$U_{rel}=5.0\%$		
64	*Plastic Rock Well Hardness Testing Machine	Hardness	Verification Regulation for Plastic Rock Well Hardness Testing Machine JJG 884	(70~94)HRE	$U=1.1HRE$		
				(100~120)HRL	$U=0.6HRL$		
				(85~110)HRM	$U=0.8HRM$		
				(114~125)HRR	$U=0.7HRR$		
		Force		(50~2000) N	$U_{rel}=0.3\%$		
		Depth		(0~1) mm	$U=0.6 \mu m$		
		Size		(0~800) mm	$U=6 \mu m$		
		Time		(1~3600) s	$U_{rel}=1.0\%$		
65	*Verification Regulation of Vibration Meters	acceleration	Verification Regulation of Measuring Vibration Instruments JJG676	(2~100000)m/s ² (0.2~ 10k) Hz	$U_{rel}=3.0\% \sim 5.0\%$		
		velocity		(0~20)m/s (0.2~ 10k) Hz	$U_{rel}=1.1\% \sim 3.0\%$		
		displacement		(0.002~1000)mm (0.2~10k) Hz	$U_{rel}=1.4\% \sim 3.0\%$		
66	*Pencil Hardness Testers	Mass	Calibration Specification for Pencil Hardness Testers JJF(石化) 007	(0.1~1000) g	$U=0.02g$		
				(1000~2000) g	$U=0.13g$		
		Angle		(0~90) °	$U=0.02^\circ$		
67	*Electrodynamic vibration testing systems	Frequency	Verification Regulation of Electrodynamic Vibration Testing System JJG 948	(0.1~2000) Hz	$U_{rel}=0.012\%$		
		Displacement		(1~50) mm (0.1~ 10k) Hz	$U_{rel}=3\%$		



No. CNAS L0128

第 31 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		acceleration		(0.05~1000) m/s ² (0.1~10k) Hz	$U_{rel}=3\%$		
68	Metallic Rockwell Hardness Reference Blocks	Hardness	Verification Regulation of Metallic Rockwell Hardness Reference Blocks (Scales A,B,C,D,E,F,G,H,K,N,T) JJG 113	(20~88) HRA	$U=0.56$ HRA		
				(85~100) HRB	$U=0.56$ HRB		
				(20~30) HRC	$U=0.56$ HRC		
				(35~55) HRC	$U=0.56$ HRC		
				(60~70) HRC	$U=0.56$ HRC		
				(70~91) HR15N	$U=0.84$ HR15N		
				(42~80) HR30N	$U=0.84$ HR30N		
				(73~93) HR15T	$U=0.84$ HR15T		
				(82~82) HR30T	$U=0.84$ HR30T		
				(55~72) HR45T	$U=0.84$ HR45T		
				(32~61) HR45N	$U=0.84$ HR45N		
69	*Bump Testing Machines	acceleration	Verification Regulation of Shock and Bump Testing Machines JJG 1174	(100~5000) m/s ²	$U_{rel}=4\%$		
		Pulse width		(0.05~240) ms	$U_{rel}=0.8\%$		
70	Aneroid Barometer	Pressure	Verification Regulation of Aneroid Barometer and Aneroid Barograph JJG 272	(500~1060) hPa	$U=0.4$ hPa		
71	Digital Barometers	Pressure	Verification Regulation of Digital Barometers JJG1084	(100~1200) hPa	$U=0.12$ hPa		
72	Multi- Component	Force	Calibration Specification for Multi-component Force Transducer JJF1560	0.1N~1MN	$U_{rel}=0.05\%$		



No. CNAS L0128

第 32 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
73	*Laser Vibrometers	Frequency	Calibration Specification for Laser Vibrometers JJF 1219	(0.1~200k)Hz	$U_{\text{rel}}=3.4 \times 10^{-5}$		
		Voltage		10mV~10V (0.1~200k)Hz	$U_{\text{rel}}=0.4\%$		
		acceleration		(0.1~100000) m/s ² (0.1~200k)Hz	$U_{\text{rel}}=0.5\%$		
74	*Temperature/Humidity/Vibration Combined Environmental Testing System	acceleration	Calibration Specification for Temperature/Humidity/Vibration Combined Environmental Testing System JJF1270	(0.5~1000) m/s ² , (20Hz~2000Hz)	$U_{\text{rel}}=3\%$		
		Temperature		(-75~200) °C	$U=0.3^{\circ}\text{C}$		
		Humidity		(10~90) %RH	$U=1.2\%\text{RH}$		
		Heating rate		(0.5~60) °C/min	$U=(0.12\sim0.38)^{\circ}\text{C/min}$		
		The wind speed		(0.5~20) m/s	$U=(0.08\sim0.60) \text{ m/s}$		
		Dew point temperature		(-20~20)°CDP	$U=0.2^{\circ}\text{CDP}$		
75	*Impact Testing Machine	Energy	Verification Regulation of Pendulum Impact Testing Machines JJG145	(0.25~750)J	$U_{\text{rel}}=0.8\%$		
		Distance		(0~0.5)mm	$U=6 \mu\text{m}$		
		Center Distance		(50~1000)mm	$U_{\text{rel}}=0.15\%$		
		Force		(1~10000) N	$U_{\text{rel}}=0.4\%$		
		Size		(0~150) mm	$U=0.04\text{mm}$		
		Angle		(0.1~180) °	$U_{\text{rel}}=0.1\%$		
76	*Electric and Pneumatic Torque Wrenches	Torque	Calibration Specification for Electric and Pneumatic Torque Wrenches JJF1610	(0.1~3000) Nm	$U_{\text{rel}}=1.5\%$		



No. CNAS L0128

第 33 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
77	Dynamic Force Sensors	Static force value	Verification Regulation of Dynamic Force Sensors JJG632	10N~1000kN	<i>U</i> _{rel} =0.4%		
		Dynamic force value		10N~1000kN（0.1Hz~ 200kHz）	<i>U</i> _{rel} =0.6%~2.0%		
		Frequency		0.1Hz~200kHz	<i>U</i> _{rel} =0.01%		
78	Elevator Overspeed Governor Tester	speed	Calibration Specification for Elevator Overspeed Governor Testers JJF 1374	(0.1000~20.000)m/s	<i>U</i> _{rel} =0.03%~0.32%		
79	Vickers Hardness Reference Blocks	Hardness	Verification Regulation for Vickers Hardness Reference Blocks JJG148	（700~800）HV1	<i>U</i> _{rel} =1.4%		
				(700~800)HV0.5	<i>U</i> _{rel} =1.6%		
				(700~800)HV0.2	<i>U</i> _{rel} =2.3%~2.5%		
				(400~600)HV0.1	<i>U</i> _{rel} =2.5%		
				(175~225)HV0.05	<i>U</i> _{rel} =3.2%		
80	*standard vibrators	acceleration	standard vibrators JJG298	（0.1~1000）m/s ² (0.1~10k)Hz	<i>U</i> _{rel} =3%		
		Frequency		(0.2~10k)Hz	<i>U</i> =2×10 ⁻⁴ Hz		
				(0.2~10k)Hz	<i>U</i> _{rel} =0.01 %		
81	*Buoy Type Oxygen Inhalers	Pressure	Verification Regulation of Buoy Type Oxygen Inhalers JJG 913	(0~25)MPa	<i>U</i> =0.17MPa		
		Flow		(1~10)L/min	<i>U</i> _{rel} =1.2%		
82	Engine Speed Meter	Rotational Speed	Calibration Specification for Motor Vehicle Engine Speed Measuring Instruments JJF1375	500r/min	<i>U</i> _{rel} =0.3%		
				(500~4500)r/min	<i>U</i> _{rel} =0.2%		
				(4500~6000)r/min	<i>U</i> _{rel} =0.1%		



No. CNAS L0128

第 34 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
83	*Syringe Pumps and Infusion Pumps	Flow	Calibration Specification for Syringe Pumps and Infusion Pumps JJF 1259	(5~20)mL/h	$U_{\text{rel}}=2.5\%$		
				(20~200)mL/h	$U_{\text{rel}}=1.2\%$		
				(200~1000)mL/h	$U_{\text{rel}}=2.5\%$		
		Pressure		(0~200) kPa	$U=2.4 \text{ kPa}$		
84	Positive Control of Drug Container - Closure Integrity Test (CCIT)	Leak Rate	Calibration Specification for Positive Control of Drug Container - Closure Integrity Test (CCIT) SQI/JL-JF-64	$(1.4 \times 10^{-5} \sim 1.4 \times 10^{-3}) \text{ Pa} \cdot \text{m}^3/\text{s}$	$U_{\text{rel}}=6.2\%$		
85	*Medical Suction Equipment	pressure	Calibration Specification of Medical Suction Equipment JJF 1810	(-0.095~0) MPa	$U=0.001 \text{ MPa}$		
86	Vacuum Helium Leaks	leak rate	Calibration Specification for Vacuum Helium Leaks JJF 1833	$(1.0 \times 10^{-10} \sim 1.0 \times 10^{-4}) \text{ Pa} \cdot \text{m}^3/\text{s}$	$U_{\text{rel}}=12\%$		
87	Gasoline Vapor Recovery Detectors	Flow-rate	Calibration Specification for Gasoline Vapor Recovery Detectors JJF 1948	(3.3~50) L/min	$U_{\text{rel}}=0.8\%$		
		Time		(10~1800) s	$U=1 \text{ s}$		
		Pressure		(-5~5)kPa	$U=3.4 \text{ Pa}$		
88	*Electromagnetic Horizontal Vibration Generator for Testing	Frequency	Verification Regulation for Electromagnetic Horizontal Vibration Generator for Testing JJG 1000	(0.1~2000) Hz	$U_{\text{rel}}=0.012\%$		
		Acceleration		(0.05~1000) m/s^2	$U_{\text{rel}}=3\%$		
		Displacement		(0.1~50) mm	$U_{\text{rel}}=3\%$		
89	*Measuring instrumentations for strain gauge transducer	voltage ratio	Calibration specification for measuring instrumentations for strain gauge transducer JJF 1469	(-100~-0.1) mV/V	$U_{\text{rel}}=0.0056\%$		
				(0.1~100) mV/V	$U_{\text{rel}}=0.0056\%$		



No. CNAS L0128

第 35 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
90	*Medical Centrifuges	Speed	Calibration Specification for Medical Centrifuges JJF 2004	(20～30000) r/min	<i>U</i> _{rel} =0.1%		
		Time		(180～3600) s	<i>U</i> _{rel} =0.3%		
		Temperature		(-10～100) °C	<i>U</i> =0.3 °C		
		Noise		(40～100) dB	<i>U</i> =3dB		
91	*Interfacial tensiometer	tension	Calibration Specification for Interface Tensiometer JJF 1464	(10～1000) mN/m	<i>U</i> _{rel} =0.2%		
		Length		(0.09～1) mm	<i>U</i> _{rel} =0.01%		
				(1～100) mm	<i>U</i> _{rel} =0.2%		
92	*Spring hammers	Energy	Calibration specification for spring hammers JJF 1475	(0.2～0.5) J	<i>U</i> =0.004J		
				(0.5～1.5) J	<i>U</i> =0.01J		
				(1.5～2.0) J	<i>U</i> =0.02J		
三、Acoustics							
1	*Ultrasonic Flaw Detectors	Attenuation	V.R. for Ultrasonic Flaw Detectors JJG 746	0.5dB～20dB, (2.5MHz～10MHz)	<i>U</i> =0.2dB		
				20dB～50dB, (2.5MHz～10MHz)	<i>U</i> =0.3dB		
				50dB～81dB, (2.5MHz～10MHz)	<i>U</i> =0.5dB		
		Linearity of Horizon Display		2%～100% ,(2.5MHz～10MHz)	<i>U</i> =1.5%		
		Linearity of Vertical Display		5%～100%, (2.5MHz～10MHz)	<i>U</i> =1.6%		



No. CNAS L0128

第 36 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date	
2	Sound Level Meters	Sound Pressure Level	V.R.of Sound Level Meters JJG 188	Acoustic Signal: (30dB~114dB),10Hz~200Hz	<i>U</i> =0.5dB			
				Acoustic Signal: (30dB~124dB), 250Hz~1.25kHz	<i>U</i> =0.4dB			
				Acoustic Signal: (30dB~114dB),1.6kHz~10kHz	<i>U</i> =0.6dB			
				Acoustic Signal: (30dB~114dB),12.5kHz~20kHz	<i>U</i> =1.0dB			
				Steady State Electrical Signal:(1dB~140dB), 31.5Hz~8kHz	<i>U</i> =0.4dB			
				Burst Signal:(10~140)dB,(0.25ms~200ms)	<i>U</i> =0.2dB			
		Time-aver Sound Level		Steady State Electrical signal :(10~140)dB, 4kHz	<i>U</i> =0.2dB			
				Burst Signal:(10~140)dB,4kHz,(0.25ms~200ms)	<i>U</i> =0.2dB			
				Attenuation Rate	F:(30~50)dB/s,4kHz			<i>U</i> =0.4dB/s
		S:(3~6)dB/s,4kHz			<i>U</i> =0.2dB/s			
3	*Anechoic Rooms and Hemi-Anechoic Rooms	Sound Pressure Level	Calibration Specification for Acoustic Performance of Anechoic Rooms and Hemi-anechoic Rooms JJF 1147	Anechoic Rooms: (10~120)dB,20Hz~10kHz	<i>U</i> =0.8dB			
				Anechoic Rooms: (10~120)dB,10kHz~20kHz	<i>U</i> =1.2dB			



No. CNAS L0128

在线扫码获取验证

第 37 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
4	Sound Calibrator	Sound Pressure Level	V.R. of Sound Calibrator JJG 176	Hemi-anechoic Rooms: (10~120)dB, 20Hz~10kHz	$U=1.0\text{dB}$		
		Frequency		Hemi-anechoic Rooms: (10~120)dB, 10kHz~20kHz	$U=1.5\text{dB}$		
		Total distortion + noise		(94~124)dB, 31.5Hz~16kHz	$U=0.15\text{dB}$		
5	*Microphone Preamplifiers	Sound Pressure Level	Calibration specification for Microphone Preamplifiers JJF 1137	Frequency Response: 120dB, 10Hz~50kHz	$U=0.3\text{dB}$		
				Transmission Loss: 120dB, 10Hz~50kHz	$U=0.1\text{dB}$		
6	Noise Level Statistical Analyzers	Sound Pressure Level	Verification Regulation of Noise Level Statistical Analyzers JJG 778	Acoustic Signal: (30dB~114dB), 10Hz~200Hz	$U=0.5\text{dB}$		
				Acoustic Signal: (30dB~124dB), 250Hz~1.25kHz	$U=0.4\text{dB}$		
				Acoustic Signal: (30dB~114dB), 1.6kHz~10kHz	$U=0.6\text{dB}$		
				Acoustic Signal: (30dB~114dB), 12.5kHz~20kHz	$U=1.0\text{dB}$		



No. CNAS L0128

第 38 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
				Steady State Electrical Signal:(1dB~140dB), 31.5Hz~8kHz	<i>U</i> =0.2dB		
				Burst Signal:(10~ 140)dB,4kHz,(0.25ms~ 200ms)	<i>U</i> =0.2dB		
				F:(30~50)dB/s,4kHz	<i>U</i> =0.4dB/s		
				S:(3~6)dB/s,4kHz	<i>U</i> =0.2dB/s		
		Attenuation Rate		Repetitive Burst Sound Signal: 10~140)dB,(0.25 ~200)ms	<i>U</i> =0.2dB		
				Steady State Electrical signal:(10~140)dB, 4kHz	<i>U</i> =0.2dB		
				Percentile Level	(5~140)dB,4kHz		<i>U</i> =0.3dB
7	Working Standard Microphone	Sound Pressure Level	Verification regulation of Working Standard Microphones(Electrostatic Actuator Method) JJG 175	124dB, 250Hz	<i>U</i> =0.17dB		
				(80~130)dB,20Hz~ 4kHz	<i>U</i> =0.20dB		
				(80~130)dB,5kHz~ 8kHz	<i>U</i> =0.30dB		
				(80~130)dB,10kHz~ 20kHz	<i>U</i> =0.50dB		
				124dB,25kHz~40kHz	<i>U</i> =0.70dB		
8	*Acoustic Performance of Reverberation Rooms	Reverberation Time	Calibration Specification for Acoustic Performance of Reverberation Rooms JJF 1143	1s~20s	<i>U</i> _{rel} =9%		



No. CNAS L0128

第 39 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		Sound pressure Level		(10~130)dB,100Hz~8kHz	<i>U</i> =0.60dB		
9	*Ultrasonic Source for Medical Ultrasonic Diagnostic Equipment	Ultrasonic Power	Verification regulation ofUltrasonic Source for Medical Ultrasonic Diagnostic Equipment JJG 639	(10~100)mW	<i>U</i> _{rel} =11%		
10	*Audio - frequency Signal Generator	Voltage	V.R.of Audio-frequency Signal Generator JJG 607	1mV~10mV（10Hz~20kHz）	<i>U</i> _{rel} =0.3%		
				10mV~100V（10Hz~20kHz）	<i>U</i> _{rel} =1.3%		
				Frequency	10Hz~20kHz		<i>U</i> _{rel} =0.2%
		Resistance		0.1 Ω ~10k Ω	<i>U</i> _{rel} =0.3%		
		THD		0.01%~0.05%,(10Hz~20kHz）	<i>U</i> =0.006%		
				0.05%~0.1%,（10Hz~20kHz）	<i>U</i> =0.012%		
				0.1%~0.2%,（10Hz~20kHz）	<i>U</i> =0.024%		
		Attenuation		1dB~10dB，1kHz	<i>U</i> =0.03dB		
				10dB~80dB，1kHz	<i>U</i> =0.06dB		
11	*Audio-frequency power amplifier	Sound pressure level	Calibration specification for audio-frequency power amplifiers JJF 1200	Frequency response: (60~120)dB, (20Hz~20kHz)	<i>U</i> =0.3dB		
				Gain control: (1~100)dB, (20Hz~20kHz)	<i>U</i> =0.1dB		



No. CNAS L0128

第 40 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
				Maximum gain: (1~100)dB, (20Hz~20kHz)	<i>U</i> =0.1dB		
				Gain difference: (1~100)dB, (20Hz~20kHz)	<i>U</i> =0.1dB		
12	*Transducers of Ultrasonic Flaw Detector	Pulse duration	Calibration Specification for Transducers of Ultrasonic Flaw Detector JJF 1294	(25~6500) ns	<i>U</i> _{rel} =4.5%		
		Frequency		(0.5~15) MHz	<i>U</i> _{rel} =4%		
		Sound Pressure Level		(1~130) dB	<i>U</i> =0.4dB		
		Index point		(0.5~80) mm	<i>U</i> _{rel} =1.5%		
		Beam angle		35° ~85°	<i>U</i> _{rel} =3%		
13	*Multi-channels sound analyzer	Frequency weighting and frequency response	calibration specification for multi-channels sound analyzers JJF 1288	Frequency weighting and frequency response:(1~140) dB, (10Hz~20kHz)	<i>U</i> =0.2dB		
				Level linearity:(1~140)dB, (10Hz~20kHz)	<i>U</i> =0.3dB		
				Burst response:(1~140) dB, (10~500)ms	<i>U</i> =0.2dB		
		Time-average Sound Level		Steady State Electrical signal and Repetitive Burst Sound Signal: (10~140) dB, 4kHz, (10~500)ms	<i>U</i> =0.2dB		
		Time-weighting F and S		F: (20~50) dB/s,4kHz	<i>U</i> =0.1dB/s		
				S:(3~6)dB/s, 4kHz	<i>U</i> =0.1dB/s		



No. CNAS L0128

第 41 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Reverberation time		(0.75~22.5) s	$U_{rel}=1.2\%$		
		The frequency of the output signal		(10Hz~20kHz)	$U_{rel}=0.2\%$		
		Amplitude of the output signal		10mV~10V, (10Hz~20kHz)	$U_{rel}=0.3\%$		
		White noise and pink noise output signal frequency band sound level		(20~140)dB, (20Hz~20kHz)	$U=1.3\text{dB}$		
四、Electromagnetism							
1	*Digital AC Electrical Parameters Meter	AC Voltage	C.S. for Digital AC Electrical Parameters Meter JJF 1491	(1~1000)V (45~65)Hz	$U_{rel}=0.05\%$		
		AC Current		(0.2~40)A (45~65)Hz	$U_{rel}=0.05\%$		
		AC Power		(0.2~40000)W (45~65)Hz	$U_{rel}=0.05\%$		
		Frequency		(45~65)Hz	$U=0.003\text{Hz}$		
		Phase (power factor)		0~360° (PF:0~1)	$U=0.0010$		
2	*Online Testers of Winding Temperature Rise	Resistance	Calibration Specification for Online Testers of Winding Temperature Rise JJF 1540	0.1 Ω ~ 10k Ω	$U_{rel}=0.02\%$		
3	*DC Digital Voltmeter	DC Voltage	C.S. of Digital multimeter JJF 1587	20mV~200mV	$U_{rel}=0.0012\%$	6.5 bits and above do not carry out	
				0.2V~2V	$U_{rel}=0.0006\%$		



No. CNAS L0128

第 42 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				2V~20V	$U_{rel}=0.0004\%$	field calibration	
				20V~200V	$U_{rel}=0.0006\%$		
				200V~1000V	$U_{rel}=0.0007\%$		
4	*Impulse Voltage Testers for Winding Interturn Insulation	Peak Voltage	C.S. for Impulse Voltage Testers for Winding Interturn Insulation JJF 1691	(100~15000)V	$U_{rel}=3\%$		
		Wave front time		10ns~100ms	$U_{rel}=3\%$		
5	*DC Digital Amperemeter	DC Current	C.S. of Digital multimeter JJF 1587	20 μ A~200 μ A	$U_{rel}=0.012\%$	6.5 bits and above do not carry out field calibration	
				0.2mA~2mA	$U_{rel}=0.006\%$		
				2mA~20mA	$U_{rel}=0.004\%$		
				20mA~200mA	$U_{rel}=0.006\%$		
				0.2A~2A	$U_{rel}=0.01\%$		
				2A~10A	$U_{rel}=0.05\%$		
				10A~20A	$U_{rel}=0.12\%$		
6	*Grow-wire Apparatus	Temperature	C.S. for Grow-wire Apparatus JJF (ZE) 1050	(0~1000) °C	$U=1.0^{\circ}\text{C}$		
		Force value		(0.1~5)N	$U=0.01\text{N}$		
		Time		(5~3600)s	$U=0.26\text{s}$		
		length		(0.1~150)mm	$U=(0.006\sim0.03)\text{mm}$		
7	*DC Digital Ohmmeter	Resistance	C.S. of Digital multimeter JJF 1587	1 Ω ~ 10 Ω	$U_{rel}=0.0024\%$		
				10 Ω ~ 100 Ω	$U_{rel}=0.0012\%$		
				0.1k Ω ~ 1k Ω	$U_{rel}=0.001\%$		



No. CNAS L0128

第 43 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				1k Ω ~ 10k Ω	$U_{rel}=0.001\%$		
				10k Ω ~ 100k Ω	$U_{rel}=0.0012\%$		
				0.1M Ω ~ 1 M Ω	$U_{rel}=0.0022\%$		
				1M Ω ~ 10M Ω	$U_{rel}=0.0042\%$		
				10M Ω ~ 100M Ω	$U_{rel}=0.03\%$		
8	*Reference Ballasts	Impedance	Calibration Specification for Reference Ballasts JJF 1502	1 Ω ~ 10k Ω	$U_{rel}=0.1\%$		
		pF		(0~1)	$U=0.0012$		
9	*AC Digital Voltmeter	AC Voltage	C.S. of Digital multimeter JJF 1587	20mV~200mV (40Hz~100kHz)	$U_{rel}=0.02\%$	6.5 bits and above do not carry out field calibration	
				0.2V~2V(40Hz~100kHz)	$U_{rel}=0.006\%$		
				2V~20V(40Hz~100kHz)	$U_{rel}=0.006\%$		
				20V~200V(40Hz~100kHz)	$U_{rel}=0.006\%$		
				200V~1000V(40Hz~30kHz)	$U_{rel}=0.01\%$		
10	*Electrical Fast Transient/Burst Simulators	Voltage	Calibration Specification for Electrical Fast Transient/Burst Simulators JJF 1672	(0.1~4)kV	$U_{rel}=3.4\%$		
		Rise Time		(0.5~20)ns	$U_{rel}=4.4\%$		
		Pulse Duration		(20~1000)ns	$U_{rel}=2.6\%$		
		Repetition Frequency		1kHz~2MHz	$U_{rel}=2.6\%$		
		Pulse Group Duration		0.1 μ s~300ms	$U_{rel}=2.6\%$		



No. CNAS L0128

第 44 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Period		(240~360)ms	$U_{rel}=2.6\%$		
11	*AC Digital Amperemeter	AC Current	C.S. of Digital multimeter JJF 1587	10mA~200mA(40Hz~10kHz)	$U_{rel}=0.02\%$	6.5 bits and above do not carry out field calibration	
				0.2A~2A(40Hz~10kHz)	$U_{rel}=0.03\%$		
				2A~10A(40Hz~10kHz)	$U_{rel}=0.05\%$		
				10A~20A(45Hz~5kHz)	$U_{rel}=0.2\%$		
12	Tesla meter	Magnetic field intensity	Calibration Specification of (1mT~2.5T) Magnetometers JJF 1832	(1~2000) mT	$U_{rel}=0.08\%$		
13	*DC low resistance meters	Resistance	Verification Regulation of DC Low Resistance Meters JJG 837	10 $\mu\Omega$ ~ 1m Ω	$U_{rel}=0.3\%$		
				1m Ω ~ 10m Ω	$U_{rel}=0.06\%$		
				10m Ω ~ 1M Ω	$U_{rel}=0.03\%$		
				1M Ω ~ 10M Ω	$U_{rel}=0.06\%$		
14	*Proof Tracking Index Tester	AC voltage	Calibration Specification for Proof Tracking Index Tester JJF(ZE) 1087	(10~600)V,(45~65)Hz	$U_{rel}=0.28\%$		
		AC current		(0.1~10)A,(45~65)Hz	$U_{rel}=0.3\%$		
		Force value		(0.1~5)N	$U=0.01N$		
		Time		(5~3600)s	$U=0.26s$		
		length		(0.1~150)mm	$U=0.0052mm$		
		angle		(1~320) $^{\circ}$	$U=0.2^{\circ}$		
15	*Multifunction Standard Sources	DC Voltage	Calibration Specification for Multifunction Standard Sources JJF 1638	10mV~200mV	$U_{rel}=0.0015\%$		



No. CNAS L0128

第 45 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date	
		DC Current	ilac-MEA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	0.2V~10V	<i>U</i> _{rel} =0.0002%			
				10V~1000V	<i>U</i> _{rel} =0.0005%			
				100 μ A~20 A	<i>U</i> _{rel} =0.003%			
		DC Resistance		1 Ω ~2 Ω	<i>U</i> _{rel} =0.0020%			
				2 Ω ~20 Ω	<i>U</i> _{rel} =0.0012%			
				20 Ω ~200 Ω	<i>U</i> _{rel} =0.0010%			
				0.2k Ω ~2M Ω	<i>U</i> _{rel} =0.0012%			
				2M Ω ~20M Ω	<i>U</i> _{rel} =0.0030%			
				20M Ω ~100M Ω	<i>U</i> _{rel} =0.008%			
				AC Votage	10mV~20mV(50Hz~50kHz)		<i>U</i> _{rel} =0.035%	
					10mV~20mV(50kHz~100kHz)		<i>U</i> _{rel} =0.098%	
		20mV~0.5V(50Hz~20kHz)			<i>U</i> _{rel} =0.0060%			
		20mV~0.5V(20kHz~100kHz)			<i>U</i> _{rel} =0.020%			
		0.5V~60V(50Hz~1kHz)			<i>U</i> _{rel} =0.0028%			
		0.5V~60V(1kHz~20kHz)			<i>U</i> _{rel} =0.0025%			
		0.5V~60V(20kHz~100kHz)		<i>U</i> _{rel} =0.0090%				
		60V~1000V(50Hz~20kHz)		<i>U</i> _{rel} =0.0035%				



No. CNAS L0128

第 46 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
				60V~1000V(20kHz~100kHz)	<i>U</i> _{rel} =0.0080%		
		AC Current		100 μ A ~1mA(50Hz~5kHz)	<i>U</i> _{rel} =0.06%		
				1mA~20A(50Hz~10kHz)	<i>U</i> _{rel} =0.0075%		
16	*AC Voltage Source	Frequency	Calibration Specification for Stable Character of AC Voltage Stable Source JJF(military) 85	40Hz~1kHz	<i>U</i> _{rel} =0.05%		
		Distortion		0.01%~30%	<i>U</i> =0.1%		
		Voltage		15V~1000V	<i>U</i> _{rel} =0.1%		
		Current		10mA~30A	<i>U</i> _{rel} =0.1%		
				30A~400A	<i>U</i> _{rel} =0.8%		
		Power		20W~3000W	<i>U</i> _{rel} =0.2%		
				3kW~240kW	<i>U</i> _{rel} =1.5%		
		Efficiency		(0.01%~120%),(20W~3000W)	<i>U</i> _{rel} =0.2%		
				(0.01%~120%),(3kW~240kW)	<i>U</i> _{rel} =1.5%		
		Input Power adjustment rate		(0.01%~100%),(15V~1000V)	<i>U</i> _{rel} =0.1%		
Load adjustment rate	(0.01%~100%),(15V~1000V)	<i>U</i> _{rel} =0.1%					
Output Voltage Stability	(0.01%~100%),(15V~1000V)	<i>U</i> _{rel} =0.1%					



No. CNAS L0128

第 47 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
17	*DC Electronic Loads	DC Voltage	Calibration Specification for DC Electronic Loads JJF 1462	0.1V~1000V	<i>U</i> _{rel} =0.01%		
		DC Current		1mA~20A	<i>U</i> _{rel} =0.02%		
				20A~500A	<i>U</i> _{rel} =0.12%		
		DC Resistance		0.1 Ω ~100k Ω	<i>U</i> _{rel} =0.03%		
		DC Power		1mW~30kW	<i>U</i> _{rel} =0.2%		
18	*Spark tester	Voltage	Verification procedure for test equipment of rubber plastic wire and cable Part10:Spark tester JB/T 4278.10	(1~20)kV(50Hz)	<i>U</i> _{rel} =1.2%		
19	*Wrist Strap/ Footwear Tester	Resistance	V.R. for High Insulation Resistance Meter JJG 690	0.1M Ω ~10M Ω	<i>U</i> _{rel} =0.24%		
		Resistance		10M Ω ~100M Ω	<i>U</i> _{rel} =1.2%		
20	*High Insulation Resistance Meter	Voltage	V.R. for High Insulation Resistance Meter JJG 690	(10~100)V	<i>U</i> _{rel} =0.04%		
				(100~1000)V	<i>U</i> _{rel} =0.60%		
		Resistance		1M Ω ~10M Ω	<i>U</i> _{rel} =0.26%		
				10M Ω ~100M Ω	<i>U</i> _{rel} =0.60%		
				100M Ω ~1G Ω	<i>U</i> _{rel} =1.2%		
				1G Ω ~10G Ω	<i>U</i> _{rel} =2.4%		
				10G Ω ~100G Ω	<i>U</i> _{rel} =5.8%		
				100G Ω ~1T Ω	<i>U</i> _{rel} =5.9%		
	*DC Standard		Calibration Specification for	10mV~200.0000mV	<i>U</i> =0.0008mV	6.5 bits and above do not	

No. CNAS L0128

第 48 页 共 143 页



The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
	(Generator)		Sources JJF 1638	200.0000mV~329.9999mV	$U=0.0017\text{mV}$	carry out field calibration	
				0.329999V~1.000000V	$U=0.000004\text{V}$		
				1.000000V~3.299999V	$U=0.000015\text{V}$		
				3.299999V~20.00000V	$U=0.00007\text{V}$		
				20.00000V~32.99999V	$U=0.00020\text{V}$		
				32.99999V~200.0000V	$U=0.0010\text{V}$		
				200.0000V~329.9999V	$U=0.0030\text{V}$		
				329.9999V~500.000V	$U=0.004\text{V}$		
				500.000V~1000.000V	$U=0.006\text{V}$		
22	*AC Standard Voltage Source (Generator)	AC Voltage	Calibration Specification for Multifunction Standard Sources JJF 1638	20mV~200mV (10Hz~40Hz)	$U=0.015\%U_x+0.004\text{mV}$	6.5 bits and above do not carry out field calibration	
				0.2V~2V (10Hz~40Hz)	$U=0.013\%U_x+0.00002\text{V}$		
				2V~20V (10Hz~40Hz)	$U=0.013\%U_x+0.0002\text{V}$		
				20V~200V (10Hz~40Hz)	$U=0.013\%U_x+0.002\text{V}$		
				200V~1000V (10Hz~40Hz)	$U=0.013\%U_x+0.02\text{V}$		
				20mV~200mV (40Hz~100Hz)	$U=0.012\%U_x+0.004\text{mV}$		
				0.2V~2V (40Hz~100Hz)	$U=0.01\%U_x+0.00002\text{V}$		
				2V~20V (40Hz~100Hz)	$U=0.01\%U_x+0.0002\text{V}$		



No. CNAS L0128

第 49 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				20V~200V (40Hz~100Hz)	$U=0.01\%U_x+0.002V$		
				200V~1000V (40Hz~100Hz)	$U=0.01\%U_x+0.02V$		
				20mV~200mV (100Hz~2kHz)	$U=0.012\%U_x+0.002mV$		
				0.2V~2V (100Hz~2kHz)	$U=0.01\%U_x+0.00002V$		
				2V~20V (100Hz~2kHz)	$U=0.01\%U_x+0.0002V$		
				20V~200V (100Hz~2kHz)	$U=0.01\%U_x+0.002V$		
				200V~1000V (100Hz~2kHz)	$U=0.01\%U_x+0.02V$		
				20mV~200mV (2kHz~10kHz)	$U=0.014\%U_x+0.004mV$		
				0.2V~2V (2kHz~10kHz)	$U=0.01\%U_x+0.00002V$		
				2V~20V (2kHz~10kHz)	$U=0.01\%U_x+0.0002V$		
				20V~200V (2kHz~10kHz)	$U=0.01\%U_x+0.002V$		
				200V~1000V (2kHz~10kHz)	$U=0.01\%U_x+0.02V$		
				20mV~200mV (10kHz~30kHz)	$U=0.014\%U_x+0.004mV$		
				0.2V~2V (10kHz~30kHz)	$U=0.01\%U_x+0.00002V$		



No. CNAS L0128

第 50 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				2V~20V (10kHz~30kHz)	$U=0.01\%U_x+0.0002V$		
				20V~200V (10kHz~30kHz)	$U=0.01\%U_x+0.002V$		
				200V~1000V (10kHz~30kHz)	$U=0.01\%U_x+0.02V$		
				20mV~200mV (30kHz~100kHz)	$U=0.075\%U_x+0.02mV$		
				0.2V~2V (30kHz~100kHz)	$U=0.055\%U_x+0.0002V$		
				2V~20V (30kHz~100kHz)	$U=0.055\%U_x+0.002V$		
				20V~200V (30kHz~100kHz)	$U=0.055\%U_x+0.02V$		
				200V~1000V (30kHz~100kHz)	$U=0.055\%U_x+0.2V$		
				0.2V~200V (100kHz~300kHz)	$U=0.5\%U_x$		
				0.2V~200V (300kHz~1MHz)	$U=2\%U_x$		
23	*DC Standard Current Source (Generator)	DC Current	Calibration Specification for Multifunction Standard Sources JJF 1638	20 μ A~2.00000mA	$U=0.00003mA$	6.5 bits and above do not carry out field calibration	
				2.00000mA~3.29999mA	$U=0.00008mA$		
				3.29999mA~20.0000mA	$U=0.0003mA$		
				20.0000mA~32.9999mA	$U=0.0022mA$		
				32.9999mA~200.000mA	$U=0.009mA$		
				200.000mA~329.999mA	$U=0.084mA$		



No. CNAS L0128

第 51 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
24	*AC Standard Current Source (Generator)	AC Current	Calibration Specification for Multifunction Standard Sources JJF 1638	0.329999A~0.99999A	$U=0.00009A$		
				0.99999A~1.49999A	$U=0.00022A$		
				1.49999A~2.00000A	$U=0.00043A$		
				2.00000A~2.19999A	$U=0.00143A$		
				2.19999A~7.0000A	$U=0.0036A$		
				7.0000A~10.9999A	$U=0.0053A$		
				10.9999A~20.0000A	$U=0.0092A$		
				100 μA ~200 μA (10Hz~2kHz)	$U=0.026\%Ix+0.02\mu A$	6.5 bits and above do not carry out field calibration	
				0.2mA~2mA (10Hz~ 2kHz)	$U=0.026\%Ix+0.2\mu A$		
				2mA~20mA (10Hz~ 2kHz)	$U=0.026\%Ix+0.002mA$		
				20mA~200mA (10Hz~ 2kHz)	$U=0.026\%Ix+0.02mA$		
				0.2A~2A (10Hz~2kHz)	$U=0.062\%Ix+0.0002A$		
				2A~20A (10Hz~2kHz)	$U=0.082\%Ix+0.002A$		
				100 μA ~200 μA (2kHz~10kHz)	$U=0.026\%Ix+0.02\mu A$		
				0.2mA~2mA (2kHz~ 10kHz)	$U=0.026\%Ix+0.2\mu A$		
				2mA~20mA (2kHz~ 10kHz)	$U=0.026\%Ix+0.002mA$		



No. CNAS L0128

第 52 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				20mA~200mA (2kHz~10kHz)	$U=0.026\%Ix+0.02\text{mA}$		
				0.2A~2A (2kHz~10kHz)	$U=0.075\%Ix+0.0002\text{A}$		
				2A~20A (2kHz~10kHz)	$U=0.3\%Ix+0.002\text{A}$		
				100 μA ~200 μA (10kHz~30kHz)	$U=0.062\%Ix+0.01\mu\text{A}$		
				0.2mA~2mA (10kHz~30kHz)	$U=0.062\%Ix+0.01\mu\text{A}$		
				2mA~20mA (10kHz~30kHz)	$U=0.062\%Ix+0.01\mu\text{A}$		
				20mA~200mA (10kHz~30kHz)	$U=0.062\%Ix+0.01\mu\text{A}$		
25	Electrical Meters for Measuring Alternating-current Electrical Energy	Alternating-current Electrical Energy	Verification Regulation of Electrical Meters for Measuring Alternating-current Electrical Energy JJG 596	$3\times(57.7\sim380)\text{V}$ $3\times(0.1\sim100)\text{A}$ (50Hz、60Hz)	$U_{\text{rel}}=0.14\%$		
26	Electromechanical Meters for Measuring Alternating-current Electrical Energy	Alternating-current Electrical Energy	Verification Regulation of Electromechanical Meters for Measuring Alternating-current Electrical Energy JJG 307	$3\times(57.7\sim380)\text{V}$ $3\times(0.1\sim100)\text{A}$ (50Hz、60Hz)	$U_{\text{rel}}=0.2\%$		
27	*Earth Resistance Meters	Resistance	Earth Resistance Meters JJG 366	$0.01\Omega\sim11111.110\Omega$	$U_{\text{rel}}=0.4\%$		
28	*Electrical Meters for Measuring Alternating-current Electrical Energy	Alternating-current Electrical Energy	Verification Regulation of Direct Meters at Place Of Installation JJG (HU)49	$3\times(57.7\sim380)\text{V}$ $3\times(0.1\sim100)\text{A}$ (50Hz)	$U_{\text{rel}}=0.14\%$		

No. CNAS L0128

第 53 页 共 143 页



The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
29	*Electromechanical Meters for Measuring Alternating-current Electrical Energy	Alternating-current Electrical Energy	Verification Regulation of Direct Meters at Place Of Installation JJG (HU)49	3×(57.7~380)V 3×(0.1~100)A (50Hz)	<i>U</i> _{rel} =0.21%		
30	*Electronic Insulating Resistance Meters	Voltage	Verification Regulation of Electronic Insulating Resistance Meters JJG 1005	50V~10kV	<i>U</i> _{rel} =0.6%		
		Resistance		100 Ω ~10M Ω	<i>U</i> _{rel} =0.3%		
				10M Ω ~100M Ω	<i>U</i> _{rel} =0.6%		
				100M Ω ~1G Ω	<i>U</i> _{rel} =1.2%		
				1000M Ω ~10G Ω	<i>U</i> _{rel} =2.4%		
				10G Ω ~100G Ω	<i>U</i> _{rel} =6%		
31	*Clamp Ammeters	AC Current	Calibration Specification for Clamp Ammeters JJF 1075	0.1A~100A(45Hz~400Hz)	<i>U</i> _{rel} =0.05%		
				100A~1000A(45Hz~400Hz)	<i>U</i> _{rel} =0.6%		
				1000A~2000A(45Hz~400Hz)	<i>U</i> _{rel} =0.8%		
		DC Current		0.1A~100A	<i>U</i> _{rel} =0.05%		
				100A~1000A	<i>U</i> _{rel} =0.6%		
				1000A~2000A	<i>U</i> _{rel} =0.8%		
32	*Voltage meter	DC Voltage	V.R.of Ammeter, Voltmeter, Power meter and Ohmmeter JJG 124	10mV~1000V	<i>U</i> _{rel} =0.042%		
		AC Voltage		10mV~1000V (45Hz~65Hz)	<i>U</i> _{rel} =0.042%		
33	*Current meter	DC current	V.R.of Ammeter, Voltmeter, Power meter and Ohmmeter	(0.001~30) A	<i>U</i> _{rel} =0.044%		



No. CNAS L0128

第 54 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date		
		AC current	JJG 124	(0.001~30) A (45Hz~65Hz)	<i>U</i> _{rel} =0.044%				
34	*Power meter	DC power	V.R.of Ammeter, Voltmeter, Power meter and Ohmmeter	(0.3~15000)W	<i>U</i> _{rel} =0.065%				
		AC power	JJG 124	(0.3~15000)W (45Hz~65Hz)	<i>U</i> _{rel} =0.065%				
35	*Leakage current tester	DC Voltage	V.R. of Leakage Current Meter JJG 843	(10~300)V	<i>U</i> _{rel} =0.11%				
		AC Voltage		(10~300)V(45Hz~400Hz)	<i>U</i> _{rel} =0.11%				
		DC Leakage current		10 μ A~200mA	<i>U</i> _{rel} =0.1%				
		AC Leakage current		10 μ A~200mA(20Hz~45Hz)	<i>U</i> _{rel} =1.5%				
				10 μ A~200mA(45Hz~400Hz)	<i>U</i> _{rel} =0.1%				
				20 μ A~200mA(400Hz~100kHz)	<i>U</i> _{rel} =1.5%				
				20 μ A~20mA, (100kHz~1MHz)	<i>U</i> _{rel} =1.5%				
		Dc resistance		(100~3000) Ω	<i>U</i> _{rel} =0.1%				
		Ac impedance		100 Ω ~20k Ω (20Hz~10kHz)	<i>U</i> _{rel} =0.2%				
				100 Ω ~20k Ω (10kHz~1MHz)	<i>U</i> _{rel} =0.4%				
36	*EARTH CONTINUITY TESTER	Current of Earth Continuity	V.R. of Earth Continuity Meter JJG 984	(1~60)A	<i>U</i> _{rel} =0.12%				



No. CNAS L0128

第 55 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		Earth Continuity Resistance		(10~1000)mΩ	<i>U</i> _{rel} =0.24%		
		Time		10s~60min	<i>U</i> _{rel} =1.2%		
37	*Insulation resistance meter	resistance	V.R. of Megohm -meter JJG 622	100Ω~1MΩ	<i>U</i> _{rel} =2.0%		
				(1~100) MΩ	<i>U</i> _{rel} =2.4%		
				(100~1000) MΩ	<i>U</i> _{rel} =3%		
				(1~10) GΩ	<i>U</i> _{rel} =4%		
		DC Voltage		(100~2500) V	<i>U</i> _{rel} =1.2%		
38	*Withstanding Voltage Tester	DC Voltage	V.R.of Withstanding Voltage Testers JJG 795	VDC (0.5~15) kV	<i>U</i> _{rel} =0.34%		
		AC Voltage		VAC (0.5~15) kV (50Hz)	<i>U</i> _{rel} =0.58%		
		DC Current		IDC (0.5~199) mA	<i>U</i> _{rel} =0.58%		
		AC Current		IAC (0.5~199) mA (50Hz)	<i>U</i> _{rel} =1.2%		
		Time		1s~60min	<i>U</i> _{rel} =1.2%		
39	Current Transformer	The Ratio difference of current	V.R.of Current Transformer for measurement JJG 313	2000A/5A, Rated current 5%	<i>U</i> _{rel} =0.052%		
				2000A/5A, Rated current 20%	<i>U</i> _{rel} =0.026%		
				2000A/5A, Rated current 100%,120%	<i>U</i> _{rel} =0.017%		
		the phase difference of current		0~999.9' , Rated current 5%	<i>U</i> =1.9'		
				0~999.9' , Rated current 20%	<i>U</i> =1.0'		



No. CNAS L0128

第 56 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date	
				0~999.9' , Rated current 100%,120%	<i>U</i> =0.7'			
40	Voltage Transformer	the Ratio difference of voltage	V.R.of Voltage Transformer for measurement JJG 314	10kV/100V, Rated voltage 20%	<i>U</i> _{rel} =0.033%			
				10kV/100V, Rated voltage 50%	<i>U</i> _{rel} =0.025%			
				10kV/100V, Rated voltage 80%,100%,120%	<i>U</i> _{rel} =0.017%			
		the phase difference of voltage		0~999.9' , Rated voltage 20%	<i>U</i> =1.4'			
				0~999.9' , Rated voltage 50%	<i>U</i> =1.0'			
				0~999.9' , Rated voltage 80%,100%,120%	<i>U</i> =0.7'			
41	*Surface Resistance Tester	Resistance	Calibration specification for surface resistance tester JJF 1285	(0.001~10)M Ω	<i>U</i> _{rel} =0.46%			
				(10~100)M Ω	<i>U</i> _{rel} =0.7%			
				(100~1×10 ³)M Ω	<i>U</i> _{rel} =1.2%			
				(1×10 ³ ~1×10 ⁴)M Ω	<i>U</i> _{rel} =2.3%			
				(1×10 ⁴ ~1×10 ⁵)M Ω	<i>U</i> _{rel} =5.8%			
				1×10 ⁶ M Ω	<i>U</i> _{rel} =2.3%			
		DC Voltage		(10~100)V	<i>U</i> _{rel} =0.1%			
				250V	<i>U</i> _{rel} =0.12%			
42	D.C. Resistance Box/D.C. Standard Resistors	Resistance	Verification Regulation of D.C.Resistance Box JJG 982,Verification Regulation	0.001 Ω ~2 Ω				
				2 Ω ~20 Ω	<i>U</i> =0.0008% <i>R</i> _x +0.000014 Ω			



No. CNAS L0128

第 57 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
			of D.C.Standard Resistors JJG 166	20 Ω ~ 200 Ω	$U=0.0008\%Rx+0.00005 \Omega$		
				0.2k Ω ~ 2k Ω	$U=0.0008\%Rx+0.0000005k \Omega$		
				2k Ω ~ 20k Ω	$U=0.0008\%Rx+0.000005k \Omega$		
				20k Ω ~ 200k Ω	$U=0.0008\%Rx+0.00005k \Omega$		
				0.2M Ω ~ 2M Ω	$U=0.0008\%Rx+0.000001M \Omega$		
				2M Ω ~ 20M Ω	$U=0.0011\%Rx+0.0001M \Omega$		
				20M Ω ~ 200M Ω	$U=0.004\%Rx+0.01M \Omega$		
				0.2G Ω ~ 1G Ω	$U=0.06\%Rx+0.001G \Omega$		
				1G Ω ~ 2G Ω	$U=1.0\%Rx+0.01G \Omega$		
				2G Ω ~ 20G Ω	$U=3\%Rx+0.1G \Omega$		
				20G Ω ~ 100G Ω	$U=6\%Rx+0.1G \Omega$		
43	D.C.Potentiometer s	DC voltage	Verification Regulation of D.C.Potentiometers JJG 123	10mV~13.11110V	$U_{rel}=0.007\%$		
44	D.C.Bridges	resistance	Verification Regulation of D.C.Bridges JJG 125	1m Ω ~ 100m Ω	$U_{rel}=0.59\%$		
				100m Ω ~ 1 Ω	$U_{rel}=0.15\%$		
				1 Ω ~ 10 Ω	$U_{rel}=0.059\%$		
				10 Ω ~ 100k Ω	$U_{rel}=0.0016\%$		
45	*Magnetic Particle Flaw Detectors	Magnetization Current	C.S. for Magnetic Particle Flaw Detectors JJF 1273	(500~6000) A	$U_{rel}=2.8\%$		



No. CNAS L0128

第 58 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
46	*AC Digital Power Meter	Power	V.R. of AC Digital Power Meter JJG 780	70V~300V, 0.1A~2A, 0.75~1.00, (45Hz~65Hz)	<i>U</i> _{rel} =0.025%	Level 0.1 and above do not carry out field calibration.	
				1V~1000V, 0.001A~20A, 0.50~1.00, (45Hz~65Hz)	<i>U</i> _{rel} =0.05%		
				1V~1000V, 0.001A~20A, 0.25~0.50, (45Hz~65Hz)	<i>U</i> _{rel} =0.07%		
47	High Voltage Electrostratic voltmete	Voltage	V.R.of High Voltage Electrostratic voltmeter JJG 494	600V~30kV, (DC、45Hz~400Hz)	<i>U</i> _{rel} =0.3%		
48	*Harmonic Analysis System	Voltage Measure	Calibration Specification for Harmonious and Flicker Analysis System JJF 1205	1V~1000V(Fundamental Frequency:45Hz~65Hz)	<i>U</i> _{rel} =0.03%		
				1V~1000V(Fundamental Frequency:45Hz~65Hz)	<i>U</i> _{rel} =0.07%		
		Current Measure		0.001A~20A(Fundamental Frequency:45Hz~65Hz)	<i>U</i> _{rel} =0.05%		
				0.001A~20A(Fundamental Frequency:45Hz~65Hz)	<i>U</i> _{rel} =0.09%		
		Voltage		(0.1~1000) V	<i>U</i> _{rel} =0.1%		
		THD		(0.1~50) % (45Hz~800Hz)	<i>U</i> _{rel} =0.1%		
		peak factor		1.30~1.50	<i>U</i> _{rel} =3.3%		
		peak phase		(80~100)°	<i>U</i> _{rel} =3.0%		



No. CNAS L0128

第 59 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		resistance		(0.1~1) Ω	<i>U</i> _{rel} =0.15%		
		inductance		(0.05~0.5) Ω	<i>U</i> _{rel} =0.2%		
49	*Lightning Surge(Oscilla- tory waves)Test Signal Generator	Voltage	Electromagnetic compatibility -Testing and measurement techniques - Surge immunity test (6.2.3 calibration of generator, 6.4 calibration of CDN, A.2.3 calibration of generator, A.4 calibration of CDN) GB/T 17626.5,Electromagnetic compatibility-Testing and measurement techniques- Part 12:Ring wave immunity test GB/T 17626.12	1V~20kV	<i>U</i> _{rel} =3.3%		
		Current		1A~20kA	<i>U</i> _{rel} =3.3%		
		Time		10ns~10ms	<i>U</i> _{rel} =3%		
50	*Human body impedance simulation network	Input impedance	Methods of measurement of touch current and protective conductor current(Appendix K Network Performance and Calibration) GB/T12113	100 Ω ~20k Ω （10Hz~10kHz）	<i>U</i> _{rel} =0.2%		
				100 Ω ~20k Ω （10kHz~1MHz）	<i>U</i> _{rel} =0.4%		
		Input-output voltage		1mV~10V （10Hz~100kHz）	<i>U</i> _{rel} =0.2%		
				1mV~10V （100kHz~1MHz）	<i>U</i> _{rel} =1.1%		
51	*Electrical fast transient/burst	Voltage	Electromagnetic compatibility-Testing and measurement techniques-	1V~10kV	<i>U</i> _{rel} =4.4%		
		Rise Time		3ns~20ns	<i>U</i> _{rel} =6.2%		



No. CNAS L0128

第 60 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Time	Electrical fast transient/burst immunity test (6.2.3) Calibration of Electrical fast transient/burst generator, 6.3.2 Calibration of CDN, 6.4.2 Calibration of capacitive coupling clamp) GB/T 17626.4	20ns~1s	U _{rel} =1.4%		
52	Electrostatic discharge simulators	Current	C.S. for Electrostatic Discharge Simulators JJF 1397	1A~60A	U _{rel} =6%		
		Rise Time		0.5ns~1ms	U _{rel} =13%		
		Voltage		1V~30kV	U _{rel} =3.3%		
53	*Standard Capacitor	Capacitance	Verification Regulation of Standard Capacitors JJG 183	10pF~1 μ F(1kHz)	U _{rel} =0.05%		
54	*Alternating Current Bridge (LCR Tester)	Resistance	V.R. of Alternating Current Bridge JJG 441	1m Ω (1kHz、100kHz)	U _{rel} =12%		
				10m Ω (1kHz、100kHz)	U _{rel} =3.0%		
				100m Ω (1kHz、100kHz)	U _{rel} =0.6%		
				1 Ω、10 Ω、100 Ω、1k Ω、10k Ω、100k Ω (1kHz、100kHz)	U _{rel} =0.12%		
		Capacitance		1pF、10pF、100pF(1kHz)	U _{rel} =0.12%		
		Capacitance		0.001 μ F、0.01 μ F、0.1 μ F、1 μ F(1kHz)	U _{rel} =0.06%		
		Capacitance		0.1 μ F~1 μ F(100Hz)	U _{rel} =0.6%		
				1 μ F~1mF(100Hz)	U _{rel} =2.3%		



No. CNAS L0128

第 61 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Inductance		0.0001H 、 1H(1kHz)	$U_{rel}=0.23\%$		
		Inductance		0.001H、 0.01H、 0.1H(1kHz)	$U_{rel}=0.12\%$		
55	*Reference ballast	Impedance	Ballasts for tubular fluorescent lamps-Performance requirements Appendix C Reference ballast calibration GB/T 14044	$1\Omega \sim 10k\Omega$	$U_{rel}=0.1\%$		
		PF		0~1	$U=0.0012$		
56	*Power Frequency magnetic field Generator	Magnetic Field	Electromagnetic compatibility Testing and measurement techniques- Power Frequency magnetic field immunity test(part 6.) GB/T 17626.8	$1\mu T \sim 0.1mT$	$U_{rel}=4\%$		
		Current		$(1 \sim 1000)A$ (16.7Hz~60Hz)	$U_{rel}=1.5\%$		
		Coil factor		$(0.4 \sim 3)/m$	$U_{rel}=4\%$		
57	*Pulse magnetic field Generator (include Damped oscillatory magnetic field)	Magnetic Field	Electromagnetic compatibility Testing and measurement techniques- Pulse magnetic field immunity test(part 6.) GB/T 17626.9,Electromagnetic compatibility Testing and measurement techniques- Damped oscillatory magnetic field immunity test(part 6.) GB/T 17626.10	$1\mu T \sim 0.1mT$	$U_{rel}=3.0\%$		
		Pulse Current		1A~20kA	$U_{rel}=3.3\%$		
		Time		1ns~1s	$U_{rel}=3\%$		
		Current		$(1 \sim 1000)A$	$U_{rel}=1.5\%$		
		Coil factor		$(0.4 \sim 3)/m$	$U_{rel}=4\%$		
58	*Resistance Strain Gauge Indicators	Resistance Strain	Verification Regulation of Resistance Strain Gauge Indicators JJG 623	$(1 \sim 500)\mu\epsilon$	$U=1\mu\epsilon$		
				$(500 \sim 1000)\mu\epsilon$	$U=2\mu\epsilon$		



No. CNAS L0128

第 62 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				(1000~2000) $\mu\epsilon$	$U=3 \mu\epsilon$		
				(2000~100000) $\mu\epsilon$	$U_{rel}=0.15\%$		
59	*DC Stabilized Power Supplies	DC Voltage	Calibration Specification DC Stabilized Power Supplies JJF 1597	(0.01~1000) V	$U_{rel}=0.03\%$		
		DC Current		(0.01~20) A	$U_{rel}=0.08\%$		
				(20~500) A	$U_{rel}=0.12\%$		
		load effect for DC Voltage power supplies		10mV~10V	$U_{rel}=0.01\%$		
		load effect for DC Current power supplies		10 μ A~10A	$U_{rel}=0.08\%$		
		source voltage effect for DC Voltage power supplies		10mV~10V	$U_{rel}=0.01\%$		
		source voltage effect for DC Current power supplies		10 μ A~10A	$U_{rel}=0.08\%$		
		PARD for DC Voltage power supplies		10mV~50V (20Hz~20MHz)	$U_{rel}=0.03\%$		
		PARD for DC Current power supplies		200 μ A~10A (20Hz~20MHz)	$U_{rel}=0.08\%$		



No. CNAS L0128

第 63 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
				(10~50) A (20Hz~20MHz)	<i>U</i> _{rel} =0.12%		
60	*Clamp Earth Resistance Meter	Resistance	V.R.of Clamp Earth Resistance Meters JJG 1054	0.001 Ω ~0.01 Ω	<i>U</i> _{rel} =2%		
				0.01 Ω ~10k Ω	<i>U</i> _{rel} =0.3%		
61	*Process Calibrator	DC Votage(measure)	Calibration Specification for Process Calibrators JJF 1472	10mV~300V	<i>U</i> _{rel} =0.002%		
		DC Votage(measure)		100 μ A~100mA	<i>U</i> _{rel} =0.02%		
		AC Votage(measure)		10mV~300V (10Hz~10kHz)	<i>U</i> _{rel} =0.03%		
		AC Votage(measure)		(0.1~200)mA (10Hz~10kHz)	<i>U</i> _{rel} =0.08%		
		Resistance(measure)		10 Ω ~100 k Ω	<i>U</i> _{rel} =0.02%		
		Frequency(measure)		1Hz~500kHz	<i>U</i> _{rel} =0.008%		
		Temperature(thermocouple measure)		(-250~-100)°C	<i>U</i> =0.5°C		
				(-100~120)°C	<i>U</i> =0.2°C		
				(120~1000)°C	<i>U</i> =0.3°C		
				(1000~1700)°C	<i>U</i> =0.6°C		
		Temperature(thermal resistance measure)		(-200~400)°C	<i>U</i> =0.2°C		
				(400~850)°C	<i>U</i> =0.3°C		



No. CNAS L0128

第 64 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		DC Voltage(output)		0.1mV~100V	$U_{rel}=0.002\%$		
		DC Current(output)		100 μ A~100mA	$U_{rel}=0.008\%$		
		Resistance(output)		1 Ω ~100 k Ω	$U_{rel}=0.008\%$		
		Frequency(output)		1Hz~50kHz	$U_{rel}=0.008\%$		
		Temperature(thermocouple output)		(-250~-100) $^{\circ}$ C	$U=0.5^{\circ}$ C		
				(-100~120) $^{\circ}$ C	$U=0.2^{\circ}$ C		
				(120~1000) $^{\circ}$ C	$U=0.3^{\circ}$ C		
				(1000~1800) $^{\circ}$ C	$U=0.6^{\circ}$ C		
		Temperature(thermal resistance output)		(-200~400) $^{\circ}$ C	$U=0.1^{\circ}$ C		
				(400~800) $^{\circ}$ C	$U=0.2^{\circ}$ C		
62	*Standard Inductor	Inductor	Verification Regulation of Standard Inductors JJG 726	100 μ H、1mH、10mH、100mH、1H (1kHz)	$U_{rel}=0.05\%$		
63	*Dips Generator	Voltage	Calibration Specification for Dips Generator JJF (浙)1061	0.1V~1000V (DC、45Hz~400Hz)	$U_{rel}=0.4\%$		
		Time		0.4ns~200s	$U_{rel}=3\%$		
64	*Power quality test analyzer	Power	Verification code for power quality analyzer DL/T 1028	0.1mW~50kW (DC、40Hz~1kHz)	$U_{rel}=0.05\%$		



No. CNAS L0128

第 65 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
				50kW~25MW (DC、40Hz~1kHz)	<i>U</i> _{rel} =0.7%		
		Voltage		0.1V~1000V (DC、40Hz~1kHz)	<i>U</i> _{rel} =0.03%		
		Current		0.1mA~50A (DC、40Hz~1kHz)	<i>U</i> _{rel} =0.03%		
				50A~2500A (DC、40Hz~1kHz)	<i>U</i> _{rel} =0.6%		
		Frequency		40Hz~1kHz	<i>U</i> _{rel} =0.008%		
		Harmonious voltage content		(2~60 times): (0.01~20)% (DC、45Hz~65Hz)	<i>U</i> =0.03%		
		Harmonious current content		(2~60 times): (0.01~20)% (DC、45Hz~65Hz)	<i>U</i> =0.04%		
		Flicker		0.10~5.00 (DC、45Hz~65Hz)	<i>U</i> _{rel} =2%		
65	*Safety performance comprehensive tester	DC Voltage(Withstanding)	Electronic Insulating Resistance Meters JJG 1005,V.R.of Withstanding Voltage Testers JJG 795,V.R. of Leakage Current Meter JJG 843,V.R. of Earth Continuity Meter JJG 984	(0.5~20) kV	<i>U</i> _{rel} =0.4%		
		AC Voltage(Withstanding)		(0.5~20) kV (50Hz)	<i>U</i> _{rel} =0.6%		
		DC Current (Withstanding)		(0.1~199) mA	<i>U</i> _{rel} =0.6%		
		AC Current(Withstanding)		(0.1~199) mA (50Hz)	<i>U</i> _{rel} =1.2%		
		resistance(Insulation)		100 Ω ~10M Ω	<i>U</i> _{rel} =0.3%		

No. CNAS L0128

第 66 页 共 143 页



The scope of the accreditation in Chinese remains the definitive version.



№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
			ilac-MRA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	(10~100) M Ω	$U_{rel}=0.6\%$		
				(100~1000) M Ω	$U_{rel}=1.2\%$		
				(1~10) G Ω	$U_{rel}=2.4\%$		
				(10~100) G Ω	$U_{rel}=6\%$		
		Voltage (IR)		50V~20kV	$U_{rel}=0.6\%$		
		DC Voltage (Leakage)		(10~300) V	$U_{rel}=0.6\%$		
		AC Voltage (Leakage)		(10~300) V (50Hz~400Hz)	$U_{rel}=0.6\%$		
		DC current(Leakage)		(0.1~199) mA	$U_{rel}=0.5\%$		
		AC current(Leakage)		(0.1~199) mA(50Hz~400Hz)	$U_{rel}=0.6\%$		
		Current (Earth Continu ity)		(1~60) A	$U_{rel}=0.12\%$		
66	*Digital Multimeter	DC Voltage	C.S. of Digital multimeter JJF 1587	(10~1000) m Ω	$U_{rel}=0.24\%$		
				10s~60min	$U_{rel}=1.2\%$		
				20 μ V~20mV	$U_{rel}=0.03\%$		
				20mV~200mV	$U_{rel}=0.003\%$		
				0.2V~2V	$U_{rel}=0.0006\%$		
				2V~20V	$U_{rel}=0.0004\%$		



No. CNAS L0128

第 67 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		DC Current		20V~200V	<i>U</i> _{rel} =0.0006%		
				200V~1000V	<i>U</i> _{rel} =0.0007%		
				10 μ A~200 μ A	<i>U</i> _{rel} =0.012%		
				0.2mA~2mA	<i>U</i> _{rel} =0.006%		
				2mA~20mA	<i>U</i> _{rel} =0.004%		
				20mA~200mA	<i>U</i> _{rel} =0.006%		
				0.2A~2A	<i>U</i> _{rel} =0.01%		
				2A~10A	<i>U</i> _{rel} =0.05%		
				>10A~100A	<i>U</i> _{rel} =0.08%		
		DC Resistance		1 Ω ~10 Ω	<i>U</i> _{rel} =0.0024%		
				10 Ω ~100 Ω	<i>U</i> _{rel} =0.0012%		
				0.1k Ω ~1k Ω	<i>U</i> _{rel} =0.001%		
				1k Ω ~10k Ω	<i>U</i> _{rel} =0.001%		
				10k Ω ~100k Ω	<i>U</i> _{rel} =0.0012%		
				0.1M Ω ~1 M Ω	<i>U</i> _{rel} =0.0022%		
				1M Ω ~10M Ω	<i>U</i> _{rel} =0.005%		
		AC Votage		10M Ω ~100 M Ω	<i>U</i> _{rel} =0.03%		
				20mV~200mV (40Hz~20kHz)	<i>U</i> _{rel} =0.02%		



No. CNAS L0128

第 68 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				20mV~200mV (20kHz~50kHz)	$U_{rel}=0.03\%$		
				20mV~200mV (50kHz~100kHz)	$U_{rel}=0.07\%$		
				0.2V~200V(40Hz~ 20kHz)	$U_{rel}=0.006\%$		
				0.2V~200V(20kHz~ 50kHz)	$U_{rel}=0.01\%$		
				0.2V~200V(50kHz~ 100kHz)	$U_{rel}=0.02\%$		
				200V~1000V(40Hz~ 1kHz)	$U_{rel}=0.01\%$		
				200V~1000V(1kHz~ 20kHz)	$U_{rel}=0.02\%$		
				200V~1000V(20kHz~ 30kHz)	$U_{rel}=0.1\%$		
		AC Current		40 μ A~200mA(40Hz~ 1kHz)	$U_{rel}=0.02\%$		
				40 μ A~200mA(1kHz~ 5kHz)	$U_{rel}=0.05\%$		
				40 μ A~200mA(5kHz~ 10kHz)	$U_{rel}=0.2\%$		
				0.2A~2A(40Hz~1kHz)	$U_{rel}=0.03\%$		
				0.2A~2A(1kHz~5kHz)	$U_{rel}=0.06\%$		
				0.2A~2A(5kHz~ 10kHz)	$U_{rel}=0.8\%$		
				2A~10A(40Hz~1kHz)	$U_{rel}=0.05\%$		



No. CNAS L0128

第 69 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
				2A~10A(1kHz~5kHz)	<i>U</i> _{rel} =0.1%		
				2A~10A(5kHz~10kHz)	<i>U</i> _{rel} =0.4%		
				10A~100A(40Hz~1kHz)	<i>U</i> _{rel} =0.1%		
				10A~100A(1kHz~10kHz)	<i>U</i> _{rel} =0.6%		
67	*Industry Frequency Single-Phase Meter	Power Factor	V.R. for Industry Frequency Single-Phase Meter JJG 440	0~1	<i>U</i> =0.002		
		Phase		0~360°	<i>U</i> = 0.12°		
68	A.C. Resistance Boxes	resistance	Calibration Specification for A.C. Resistance Boxes JJF 1636	0.01 Ω ~0.1 Ω	<i>U</i> _{rel} =3.5%		
				0.1 Ω ~1 Ω	<i>U</i> _{rel} =0.38%		
				1 Ω ~100 Ω	<i>U</i> _{rel} =0.06%		
				100 Ω ~100000 Ω	<i>U</i> _{rel} =0.028%		
				100000 Ω ~10000000 Ω	<i>U</i> _{rel} =0.036%		
		time constant		1ns~100 μ s	<i>U</i> _{rel} =0.8%		
		residual resistance		1m Ω ~1 Ω	<i>U</i> _{rel} =0.01%		
		residual inductance		10 μ H~100 μ H	<i>U</i> _{rel} =0.2%		
69	*Durability test equipment for household appliances and lamps	DC Voltage	C.S for Durability test equipment for household appliances and lamps SQI/JL-JF-60	0.1V~1000V	<i>U</i> _{rel} =0.01%		
		AC Voltage		0.1V~1000V(40Hz~1kHz)	<i>U</i> _{rel} =0.02%		
		DC Current		10mA~30A	<i>U</i> _{rel} =0.06%		
				30A~2000A	<i>U</i> _{rel} =2%		



No. CNAS L0128

第 70 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		AC Current	ilac-M	10mA~30A(40Hz~1kHz)	$U_{rel}=0.06\%$		
				30A~2000A(40Hz~1kHz)	$U_{rel}=2\%$		
		impulse voltage		0.1kV~20kV	$U_{rel}=3\%$		
		impulse current		10A~2000A	$U_{rel}=2\%$		
		Power		1W~600kW	$U_{rel}=0.1\%$		
		PF		0~1	$U=0.002$		
		rotate speed		(10~10000)r/min	$U_{rel}=0.12\%$		
		time		1ms~24h	$U=0.006s$		
		temperature		(0~1000)°C	$U=0.3^{\circ}C$		
		force		(0.1~500)N	$U_{rel}=0.2\%$		
		size		(0.1~10000)mm	$U=0.03mm$		
70	*Energy consumption test equipment for refrigeration and washing appliances	voltage	Calibration Specification for Energy consumption test equipment for refrigeration and washing appliances SQI/JL-JF-61	0.01V~1000V	$U_{rel}=0.05\%$		
		current		0.001A~80A	$U_{rel}=0.05\%$		
		power		0.01W~80kW	$U_{rel}=0.05\%$		
		frequency		40Hz~30kHz	$U_{rel}=0.01\%$		
		PF		0~1	$U=0.0006$		
		THD		(0.1~20)% (50Hz~800Hz)	$U=0.5\%$		



No. CNAS L0128

第 71 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		electric energy		0.01V~1000V; 0.001A~80A	<i>U</i> _{rel} =0.05%		
		PT100		(-189.3442~660.323)℃	<i>U</i> =0.2℃		
		thermocouple		(-189.3442~660.323)℃	<i>U</i> =0.4℃		
		pressure		(-0.1~60)MPa	<i>U</i> =0.2%FS		
		flow		0.001m³/h~40m³/h	<i>U</i> _{rel} =0.2%		
		temperature		(-25~100)℃	<i>U</i> =0.2℃		
		humidity		20%RH~100%RH	<i>U</i> =1%RH		
		wind speed		5m/s~25m/s	<i>U</i> _{rel} =1%		
		time		0.01s~24h	<i>U</i> _{rel} =0.1%		
		metering		1time~10000time	<i>U</i> =1 次		
		rotate speed		(20~33000) r/min	<i>U</i> _{rel} =0.02%		
		71		*High-voltage withstand voltage tester	DC voltage		High-voltage withstand voltage tester JJG(军工) 18
AC voltage	(0.5~100)kV(45~400)Hz		<i>U</i> _{rel} =1.2%				
DC current	(0.01~0.02)mA		<i>U</i> _{rel} =1.2%				
	(0.02~1000)mA		<i>U</i> _{rel} =0.6%				
AC current	(0.01~0.02)mA(45~400)Hz		<i>U</i> _{rel} =1.2%				
	(0.02~1000)mA(45~400)Hz		<i>U</i> _{rel} =0.6%				
time	(0.1~3600)s		<i>U</i> _{rel} =1.2%				



No. CNAS L0128

第 72 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
72	*Power Frequency A.C. Electrical Quantities Measuring Transducers	AC voltage	V.R.of Power Frequency A.C. Electrical Quantities Measuring Transducers JJG 126	0.1V~1000V(45Hz~ 65Hz)	<i>U</i> _{rel} =0.03%		
		AC current		0.2A~50A(45Hz~ 65Hz)	<i>U</i> _{rel} =0.03%		
				50A~2500A(45Hz~ 65Hz)	<i>U</i> _{rel} =0.6%		
		AC power		0.1W~50kW(0.1V~ 1000V, 0.2A~50A, 45Hz~65Hz)	<i>U</i> _{rel} =0.06%		
				5W~2.5MW(0.1V~ 1000V, 50A~2500A, 45Hz~65Hz)	<i>U</i> _{rel} =0.7%		
		Power Factor		0~1	<i>U</i> =0.001		
73	*Charge & Discharge of Battery Tester	DC Voltage	Calibration Specification for Charge&Discharge of Battery Tester JJF(军工) 108	10mV~1000V	<i>U</i> _{rel} =0.01%		
		DC Current		10 μ A~20A	<i>U</i> _{rel} =0.02%		
				20A~500A	<i>U</i> _{rel} =0.15%		
				500A~2000A	<i>U</i> _{rel} =0.5%		
				Resistance	0.1 Ω ~100 Ω		<i>U</i> _{rel} =0.4%
		100 Ω ~1k Ω			<i>U</i> _{rel} =0.04%		
		Power		10mW~200W(1V~ 10V, 10mA~20A)	<i>U</i> _{rel} =0.04%		
				200W~10kW(1V~ 1000V, 20A~2000A)	<i>U</i> _{rel} =0.5%		
		Capability		0.1Ah~10Ah	<i>U</i> _{rel} =0.06%		



No. CNAS L0128

第 73 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				10Ah~1kAh	$U_{rel}=0.6\%$		
		Rise time		$1\mu s\sim 10ms$	$U_{rel}=4\%$		
		temperature		$-20^{\circ}C\sim 100^{\circ}C$	$U=0.2^{\circ}C$		
74	Coil Number Testing Instructing	Coil Number	Calibration Specification for Coil Number Testing Instrument JJF(浙) 1065	(10~500)Coil Number	$U=1\text{Coil Number}$		
				(1000~2000)Coil Number	$U=3\text{Coil Number}$		
				(3000~5000)Coil Number	$U=4\text{Coil Number}$		
				(6000~7000)Coil Number	$U=5\text{Coil Number}$		
				(7000~8000)Coil Number	$U=6\text{Coil Number}$		
				10000 Coil Number	$U=7\text{Coil Number}$		
75	*AC Charging Spot for Electric Vehicles	Electric energy	V.R.of AC Charging Spot for Electric Vehicles(for Trial Implementation) JJG 1148	$3\times 220V$, $3\times (0.1A\sim 60)A$, $0.5C\sim 1.0\sim 0.5L$	$U_{rel}=0.2\%$		
76	*Off-board Charger for Electric Vehicles	Electric energy	V.R.of Off-board Charger for Electric Vehicles(for Trial Implementation) JJG 1149	$150V\sim 1000V$, $2.5A\sim 250A$	$U_{rel}=0.2\%$		
77	DC shunt	resistance	Verification Regulation of DC shunts JJG 1069	$(0.0001\sim 10)\Omega$ (Testing Current $\leq 5000A$)	$U_{rel}=0.06\%$	DC standard source method	
				$(0.00001\sim 0.0001)\Omega$ (Testing Current $\leq 5000A$)	$U_{rel}=0.1\%$		
				$(0.000001\sim 0.00001)\Omega$ (Testing Current $\leq 5000A$)	$U_{rel}=0.12\%$		



No. CNAS L0128

第 74 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
78	*Damped oscillatory wave generator	short circuit current	Electromagnetic compatibility—Testing and measurement techniques—Damped oscillatory wave immunity test GB/T 17626.18	(1~80)A	U _{rel} =3.5%		
		open circuit voltage		(0.2~4)kV	U _{rel} =3.7%		
		Time		10ns~10s	U _{rel} =3.1%		
		Frequency		1Hz~50MHz	U _{rel} =3.1%		
79	*Loop Resistance Tester and DC Resistance Meters	Resistance	Verification Regulation of Loop Resistance Tester and DC Resistance Meters JJG 1052	10 μ Ω ~ 1m Ω	U _{rel} =0.12%		
				1m Ω ~ 600m Ω	U _{rel} =0.06%		
		Current		1A~600A	U _{rel} =0.13%		
80	*Battery Internal Resistance Testers	Resistance	Calibration Specification for Battery Internal Resistance Testers JJF 1620	1m Ω ~ 10m Ω	U _{rel} =0.2%		
				10m Ω ~ 3k Ω	U _{rel} =0.06%		
		Voltage		±(0.1~800)V	U _{rel} =0.003%		
81	*Moment Disconnection Tester	Resistance	Calibration Specification for Moment Disconnection Tester JJF (军工) 198	1 Ω ~ 5k Ω	U _{rel} =1.3%		
		DC Voltage		3V~30V	U _{rel} =0.7%		
		DC Electric current		10mA~1A	U _{rel} =0.7%		
		Time		0.01 μ s~ 100 μ s	U _{rel} =0.6%		
82	*Roadvehicles-Electrical Transient Conduction Disturbance Simulators	DC Voltage	Calibration Specification of Road vehicles-Electrical Transient Conduction Disturbance Simulators JJF(电子) 00019	(10~60)V	U _{rel} =0.6%		
		Pulse voltage		P1,P2,P4,P5:(0.1~700)V	U _{rel} =3.2%		
				P3:(0.1~250)V	U _{rel} =4.6%		
		Time		(0.5~20)ns	U _{rel} =4.4%		



No. CNAS L0128

第 75 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
				20ns~2s	<i>U</i> _{rel} =3.2%		
83	*Standard AC Power Source	AC Power	Verification Regulation for Standard AC Power Source JJG (军工)6	5W~30kW, (10Hz~10kHz)	<i>U</i> _{rel} =0.06%		
		AC Voltage		10mV~200V, (10Hz~40Hz)	<i>U</i> _{rel} =0.02%		
				10mV~200V, (40Hz~10kHz)	<i>U</i> _{rel} =0.013%		
				200V~1000V, (10Hz~10kHz)	<i>U</i> _{rel} =0.015%		
		AC current		1mA~100A, (10Hz~40Hz)	<i>U</i> _{rel} =0.02%		
				1mA~100A, (40Hz~10kHz)	<i>U</i> _{rel} =0.015%		
		Phase		(-120~180)°	<i>U</i> =0.02°		
		Frequency		10Hz~10kHz	<i>U</i> _{rel} =0.06%		
84	*Test Devices for Action Characteristics of Low Voltage Circuit Breakers	current	Calibration Specification for Test Devices for Action Characteristics of Low Voltage Circuit Breakers JJF 1799	1A~2kA	<i>U</i> _{rel} =0.14%		
				2kA~10kA	<i>U</i> _{rel} =0.4%		
				10kA~40kA	<i>U</i> _{rel} =3.2%		
				40kA~90kA	<i>U</i> _{rel} =6.2%		
		time		1ms~1s	<i>U</i> _{rel} =2%		
				1s~1h	<i>U</i> _{rel} =0.3%		
85	*Impulse Current Tester	current	Calibration Specification for Impulse Current Tester JJF (浙)1110	1kA~40kA	<i>U</i> _{rel} =3.2%		
				40kA~90kA	<i>U</i> _{rel} =6.2%		



No. CNAS L0128

第 76 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		time		10ns~100ms	$U_{\text{rel}}=3.8\%$		
86	*Cable Testers	AC Voltage	Calibration Specification for Cable Testers JJF 1457	10V~1kV (50Hz、 60Hz)	$U_{\text{rel}}=0.3\%$		
				1kV~1.5kV(50Hz、 60Hz)	$U_{\text{rel}}=0.7\%$		
		Distortion		0.1%~20%	$U=0.3\%$		
		AC Current		(0.1~2)mA	$U_{\text{rel}}=0.4\%$		
		Voltage(IR)		10V~1kV	$U_{\text{rel}}=0.2\%$		
				1kV~1.5kV	$U_{\text{rel}}=0.6\%$		
		resistance(Ins ulation)		10k Ω ~ 10M Ω	$U_{\text{rel}}=0.5\%$		
				10M Ω ~ 100M Ω	$U_{\text{rel}}=0.7\%$		
				100M Ω ~ 200M Ω	$U_{\text{rel}}=1.3\%$		
		Resistance		0.1 Ω ~ 1M Ω	$U_{\text{rel}}=0.4\%$		
		Capacitance		1pF~1 μ F	$U_{\text{rel}}=0.3\%$		
87	*Surge Simulators	Voltage	Calibration Specification for Surge Simulators JJF 1741	(0.5~4)kV	$U_{\text{rel}}=3.5\%$		
		Current		(0.25~2)kA	$U_{\text{rel}}=4.4\%$		
		Time		50ns~100s	$U_{\text{rel}}=3.2\%$		
88	*Power Frequency Magnetic Field Simulators	Current	Calibration Specification for Power Frequency Magnetic Field Simulators JJF 1737	10mA~1600A(16Hz~ 60Hz)	$U_{\text{rel}}=2\%$		
		Magnetic Field Intensity		(1~1000)A/m (16Hz~ 50Hz)	$U_{\text{rel}}=12\%$		



No. CNAS L0128

第 77 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				(1~1000)A/m (50Hz~60Hz)	$U_{rel}=6\%$		
89	*Testers for Relaying Protection	DC Voltage	Verification Regulation of Testers for Relaying Protection JJG 1112	10mV~1000V	$U_{rel}=0.004\%$		
		AC Voltage		10mV~1000V(50Hz~20kHz)	$U_{rel}=0.010\%$		
		DC Current		10 μ A~100A	$U_{rel}=0.008\%$		
		AC Current		100 μ A~100A(50Hz~5kHz)	$U_{rel}=0.016\%$		
		Harmonic Distortion		(0.1~20)%	$U=0.1\%$		
		Phase		0° ~360°	$U=0.02^\circ$		
		Time		1ms~9999.9s	$U_{rel}=0.02\%$		
		Rate		10Hz~1000Hz	$U_{rel}=0.0003\%$		
90	*AC Electronic Load	Voltage	Calibration Specification for AC Electronic Load JJF(电子)0002	0.1V~750V(40Hz~1000Hz)	$U_{rel}=0.03\%$		
		Current		0.1A~90A(40Hz~1000Hz)	$U_{rel}=0.08\%$		
		Frequency		40Hz~1000Hz	$U_{rel}=0.06\%$		
		Resistance		2.5 Ω ~1000 Ω	$U_{rel}=0.2\%$		
		Power		1W~13kW(40Hz~1000Hz)	$U_{rel}=0.2\%$		
		Power Factor		0.01~1.00	$U=0.02$		
91	*Magnetic Yoke Detector	Magnetizing Current	Calibration Specification for Magnetic Yoke Detectors JJF1458	(0.1~20)A	$U_{rel}=5\%$		
		Lifting Force		10N~300N	$U_{rel}=4\%$		



No. CNAS L0128

第 78 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
92	*High Current Generator	AC Current	Calibration Specification of High Current Generator JJF(Mechanic) 1037	1A~2kA	$U_{rel}=0.15\%$		
93	Contactless Electrostatic Voltage Measuring Instruments	Electrostatic Voltage	Calibration Specification for Contactless Electrostatic Voltage Measuring Instruments JJF 1517	100V~20kV	$U_{rel}=0.6\%$		
94	*Medical Leakage Current Testers	Leakage Current	Verification Regulation of Medical Leakage Current Testers JJG 1188	DC: (0.01~20) mA	$U_{rel}=0.2\%$		
		DC resistance		AC: (0.01~20) mA (50Hz/60Hz)	$U_{rel}=0.5\%$		
		Input impedance		(0.5~2) k Ω	$U_{rel}=0.4\%$		
		Transmission Impedance		(0.5~2) k Ω	$U_{rel}=0.5\%$		
		Frequency Response		(-70~0) dB (10Hz~1MHz)	$U=0.2\text{dB}$		
		Voltage		AC: (0.01~600) V (50Hz/60Hz)	$U_{rel}=0.5\%$		
95	*Eddy Current Conductivity Meters	Conductivity	Calibration Specification for Eddy Current Conductivity Meters JJF 1692	0.58 MS/m~58MS/m	$U=(0.09\sim0.32)\text{ MS/m}$		
五、Radio							
1	*EMI TEST RECEIVER	reference frequency	Calibration Specification for EMI Testing Receivers JJF 1144	10MHz	$U_{rel}=6\times 10^{-10}$		
		IF output frequency		0.1MHz~1GHz	$U_{rel}=6\times 10^{-10}$		



No. CNAS L0128

第 79 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		1st IF image frequency interference		(0~130)dB(0.1MHz~1GHz)	<i>U</i> =5dB		
		2nd IF image frequency interference		(0~130)dB(0.1MHz~1GHz)	<i>U</i> =5dB		
		IF rejection		(0~130)dB(0.1MHz~1GHz)	<i>U</i> =5dB		
		bandwidth of 6dB filter		200Hz,9kHz,120kHz,1MHz	<i>U</i> _{rel} =1.5%		
		noise indication		(-40~30)dB μV(0.1MHz~500MHz)	<i>U</i> =3dB		
		attenuator		(0~110)dB(0.1MHz~1GHz)	<i>U</i> =0.5dB		
		level measurement		(30~120)dB μV(0.1MHz~1GHz)	<i>U</i> =0.30dB		
		pulse characteristic		9kHz~30MHz(-10~40)dB	<i>U</i> =0.8dB		
				30MHz~1GHz(-10~40)dB	<i>U</i> =1.0dB		
	input VSWR		1~10(0.1MHz~1GHz)	<i>U</i> _{rel} =7%			
2	*Low-frequency volt-meters	Voltage	C.S. for Low-frequency volt-meters JJF 1925	1mV ~10mV(10Hz~1MHz)	<i>U</i> _{rel} =1.4%		
				10mV ~100V(10Hz~1MHz)	<i>U</i> _{rel} =0.4%		
3	*Oscilloscope voltage probe	Voltage decay ratio	C.S. for Oscilloscope Voltage Probes JJF 1437	1:1~1:10000 (1V~30000V)	<i>U</i> _{rel} =0.7%		
		Frequency response		0.1V~50V (50kHz~100MHz)	<i>U</i> =1.0dB		



No. CNAS L0128

第 80 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		Resistance		50 Ω ~100M Ω	<i>U</i> _{rel} =0.6%		
		Rise time		1ns~10ms	<i>U</i> _{rel} =10%		
4	*TDMA-GSM Radio Communication Testers	SG frequency	Calibration Specification for TDMA-GSM Radio Communication Testers JJF 1131	10MHz ~2.7GHz	<i>U</i> _{rel} =1.2×10 ⁻⁸		
		SG Level		(-97~13)dBm(10MHz~ 2.7GHz)	<i>U</i> =0.26dB		
				(-107~- 97)dBm(10MHz~ 2.7GHz)	<i>U</i> =0.38dB		
				(-117~- 107)dBm(10MHz~ 2.7GHz)	<i>U</i> =0.51dB		
				(-120~- 117)dBm(10MHz~ 2.7GHz)	<i>U</i> =0.65dB		
		SG phase (peak)		0.3° ~60°	<i>U</i> =1.7°		
		SG phase (rms)		0.3° ~60°	<i>U</i> =0.5°		
		SG modulate frequency		1Hz~250kHz	<i>U</i> =5Hz		
		SG modulate level		0dB~10dB(1Hz~ 250kHz)	<i>U</i> =0.2dB		
		Analyzer frequency		10MHz ~2.7GHz	<i>U</i> _{rel} =1.6×10 ⁻⁸		
Analyzer level	(-20~10)dBm (10MHz~ 2.7GHz)	<i>U</i> =0.28dB					
	(-70~-20)dBm (10MHz~2.7GHz)	<i>U</i> =0.30dB					



No. CNAS L0128

第 81 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date	
		Analyzer phase (peak)		0.3° ~60°	<i>U</i> =1.7°			
		Analyzer phase (rms		0.3° ~60°	<i>U</i> =0.5°			
		Analyzer modulate frequency		1Hz~250kHz	<i>U</i> =5Hz			
5	*CDMA Digital Radio Communication Testers	SG frequency	Calibration Specification of CDMA Digital Radio Communication Testers JJF 1177	10MHz ~2.7GHz	<i>U</i> _{rel} =1.2×10 ⁻⁸			
		SG Level		(-97~13)dBm(10MHz~ 2.7GHz)	<i>U</i> =0.26dB			
				(-107~- 97)dBm(10MHz~ 2.7GHz)	<i>U</i> =0.38dB			
				(-117~- 107)dBm(10MHz~ 2.7GHz)	<i>U</i> =0.51dB			
				(-120~- 117)dBm(10MHz~ 2.7GHz)	<i>U</i> =0.65dB			
		SG wave quality		0.9~1.0	<i>U</i> =0.001			
		SG EVM		1.5%~50%	<i>U</i> =2.0%			
		SG modulate frequency		1Hz~250kHz	<i>U</i> =5Hz			
		SG phase (rms)		0.3° ~60°	<i>U</i> =0.5°			
Analyzer frequency	10MHz ~2.7GHz	<i>U</i> _{rel} =1.6×10 ⁻⁸						



No. CNAS L0128

第 82 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
6	*Vector Signal Generator	Analyzer level	Calibration Specification for Vector Signal Generator JJF 1174	(-20~10)dBm(10MHz~2.7GHz)	$U=0.28\text{dB}$		
		Analyzer wave quality		(-70~-20)dBm(10MHz~2.7GHz)	$U=0.30\text{dB}$		
		Analyzer EVM		0.9~1.0	$U=0.001$		
		Analyzer phase (rms)		1.5%~50%	$U=2.0\%$		
		Analyzer modulate frequency		0.3° ~60°	$U=0.5^\circ$		
				1Hz~250kHz	$U=5\text{Hz}$		
		Frequency		250kHz~20GHz	$U_{\text{rel}}=1\times 10^{-8}$		
		Output High Level		(-20~30)dBm (10MHz~6GHz)	$U=0.4\text{dB}$		
		Output Low Level		(0~110)dB (2GHz~6GHz)	$U=0.3\text{dB}$		
		GSM/CDMA Modulated Frequency Error		1Hz~250kHz	$U=5\text{Hz}$		
		CDMA EVM		1.5%~50%	$U=2.0\%$		
		CDMA Wave Quality Rho		0.9~1.0	$U=0.001$		
		GSM/CDMA Phase Error(RMS)		0.3° ~60°	$U=0.5^\circ$		



No. CNAS L0128

第 83 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
7	*WLAN Test Set	Frequency	C.S. for WLAN Test Set JJF 1277	10MHz~8GHz	$U_{rel}=1.2\times 10^{-8}$		
		Level		(-100~20)dBm (10MHz~3.6GHz)	$U=0.28\text{dB}$		
				(-100~20)dBm (3.6GHz~8GHz)	$U=0.31\text{dB}$		
				(-90~0)dBc (10MHz~8GHz)	$U=1.2\text{dB}$		
		Spectral Duration		(-90~0)dBc (frequency deviation: 10Hz~2MHz)	$U=3\text{dB}$		
		Single-side Phase Noise					
		EVM		0~8%(BW:5MHz~80MHz)	$U=0.4\%$		
				0~8%(BW:80MHz~160MHz)	$U=1.2\%$		
		Frequency Error		-100kHz~100kHz	$U=3.3\text{Hz}$		
		Level Measure		(-90~10)dBm (200kHz~6GHz)	$U=0.57\text{dB}$		
EVM Measure	0~8%(BW:5MHz~160MHz)	$U=0.7\%$					
Frequency Error Measure	-100kHz~100kHz	$U=9.2\text{Hz}$					
8	*Bluetooth Test Set	Frequency	C.S. for Bluetooth Test Set JJF 1278	10MHz~3GHz	$U_{rel}=1.2\times 10^{-8}$		
		Level		(-100~20)dBm (10MHz~3GHz)	$U=0.28\text{dB}$		
		Spectral Purity		(-90~0)dBc (10MHz~3GHz)	$U=1.2\text{dB}$		
		Single-side Phase Noise		(-90~0)dBc (frequency deviation: 10Hz~2MHz)	$U=3\text{dB}$		



No. CNAS L0128

第 84 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		GFSK Frequency Error		-250kHz~250kHz (10MHz~3GHz)	<i>U</i> _{rel} =1.2%		
		EVM		0.3%~12%	<i>U</i> =1.0%		
		Frequency Error		-100kHz ~ 100kHz	<i>U</i> =12Hz		
		Level Measure		(-90~10)dBm (200kHz~3GHz)	<i>U</i> =0.56dB		
		GFSK Frequency Error Measure		-250kHz~250kHz (10MHz~3GHz)	<i>U</i> _{rel} =1.2%		
		EVM Measure		0.3%~12% (10MHz~ 3GHz)	<i>U</i> =1.2%		
		Frequency Error Measure		-100kHz ~ 100kHz	<i>U</i> =12Hz		
		9		*LTE Radio Communication Testers	Frequency		C.S. for LTE Radio Communication Testers JJF 1443
Level	(-100~20)dBm (10MHz~3.6GHz)		<i>U</i> =0.28dB				
	(-100~20)dBm (3.6GHz~3.8GHz)		<i>U</i> =0.31dB				
Spectral Duration	(-90~0)dBc (10MHz~ 3.8GHz)		<i>U</i> =1.2dB				
Single-side Phase Noise	(-90~0)dBc (frequency deviation:10Hz~2MHz)		<i>U</i> =3dB				
EVM(RMS)	0.3%~17.5% (600MHz~3.8GHz)		<i>U</i> =0.80%				
IQ Offset	(-5~-80) dB (600MHz~3.8GHz)		<i>U</i> =1.5dB				



No. CNAS L0128

第 85 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Frequency Error		(-80~80)kHz (600MHz~3.8GHz)	$U=8.0\text{Hz}$		
		Level Measure		(-90~10)dBm(600MHz~3.8GHz)	$U=0.57\text{dB}$		
		EVM Measure		0.3%~17.5% (600MHz~3.8GHz)	$U=0.80\%$		
		Frequency Error Measure		(-80~80)kHz (600MHz~3.8GHz)	$U=8.0\text{Hz}$		
10	*Coaxial resistive attenuator	Attenuation	V.R. of coaxial resistive attenuator JJG 387	(0~90)dB (9kHz~100kHz)	$U=0.04\text{dB}$		
				(0~10)dB (100kHz~20GHz)	$U=0.02\text{dB}$		
				(10~40)dB (100kHz~20GHz)	$U=0.04\text{dB}$		
				(40~60)dB (100kHz~20GHz)	$U=0.05\text{dB}$		
				(60~70)dB (100kHz~20GHz)	$U=0.18\text{dB}$		
				(70~80)dB (100kHz~20GHz)	$U=0.20\text{dB}$		
				(80~90)dB (100kHz~20GHz)	$U=0.28\text{dB}$		
				(90~100)dB (100kHz~20GHz)	$U=0.30\text{dB}$		
		Standing-wave ratio		(1~2) (9kHz~20GHz)	$U_{\text{rel}}=1\%$		
11	*Spectrum analyzer	Frequency	Calibration specification of spectrum analyzer JJF 1396	100kHz~10MHz	$U_{\text{rel}}=1 \times 10^{-6}$		



No. CNAS L0128

第 86 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date	
		Power level	ilac-MRA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	10MHz~20GHz	<i>U</i> _{rel} =2×10 ⁻⁷			
				(0~ -20)dBm(10MHz ~6GHz)	<i>U</i> =0.18dB			
				(0~ -20)dBm(6GHz ~12GHz)	<i>U</i> =0.35dB			
				(0~ -20)dBm(12GHz ~20GHz)	<i>U</i> =0.5dB			
		Level		(0~60)dB (250kHz~1GHz)	<i>U</i> =0.25dB			
				(60~100)dB(250kHz~1GHz)	<i>U</i> =0.45dB			
		Frequency Span		10Hz~20GHz	<i>U</i> _{rel} =1.5%			
		Resolution Bandwidth		1Hz~10MHz	<i>U</i> _{rel} =1.5%			
		Noise sideband		100Hz~1MHz	<i>U</i> =3dB			
		Residual FM		0.2Hz~2Hz	<i>U</i> =0.2Hz			
				2Hz~50Hz	<i>U</i> =2Hz			
		Display average noise level		-160dBm~-50dBm (9kHz~20GHz)	<i>U</i> =3dB			
		Harmonic Distortion		-100dBc~0dBc(10MHz~20GHz)	<i>U</i> =3dB			
12	*Network analyzer	Frequency	Calibration specification of vector network analyzer JJF 1495	300 kHz~20GHz	<i>U</i> _{rel} =1×10 ⁻⁶			
		Output Amplitude of signal source		(20 ~ -70) dBm(300 kHz~20GHz)	<i>U</i> =0.18dB			



No. CNAS L0128

第 87 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		Transmission and Reflection		Attenuation:(20 ~ 50) dB(50MHz~18GHz)	<i>U</i> =0.5dB		
				Reflection: (0~1) (50MHz~18GHz)	<i>U</i> =0.06		
				Transmission: (-180° ~180°) (50MHz~18GHz)	<i>U</i> =3°		
13	*Signal Generator	Frequency	Calibration specification for signal generators JJF 1931	0.1MHz~26.5GHz	<i>U</i> _{rel} =6×10 ⁻⁹		
		Power		(0~+20)dBm (0.1MHz~2GHz)	<i>U</i> =0.14dB		
				(0~+20)dBm (2GHz~4.2GHz)	<i>U</i> =0.28dB		
				(0~+20)dBm (4.2GHz~18GHz)	<i>U</i> =0.30dB		
				(0~+20)dBm (18GHz~26.5GHz)	<i>U</i> =0.40dB		
				(-120~0)dBm (0.1MHz~2GHz)	<i>U</i> =0.20dB		
				(-120~0)dBm (2GHz~18GHz)	<i>U</i> =0.34dB		
				(-120~0)dBm (18GHz~26.5GHz)	<i>U</i> =0.43dB		
				AM depth	(5~99)% carrier frequency: 0.1MHz~10MHz modulation frequency: 400Hz、1kHz		<i>U</i> _{rel} =0.86%



No. CNAS L0128

第 88 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				(20~99)% carrier frequency: 10MHz~ 3GHz modulation frequency: 400Hz、 1kHz	$U_{rel}=0.58\%$		
				(5~20)% carrier frequency: 10MHz~ 3GHz modulation frequency: 400Hz、 1kHz	$U_{rel}=2.9\%$		
				(20~99)% carrier frequency: 3GHz~ 26.5GHz modulation frequency: 400Hz、 1kHz	$U_{rel}=1.7\%$		
				(5~20)% carrier frequency: 3GHz~ 26.5GHz modulation frequency: 400Hz、 1kHz	$U_{rel}=5.2\%$		
		FM deviation		(0.1~400)kHz carrier frequency: 0.1MHz~ 26.5GHz modulation frequency: 400Hz、 1kHz	$U_{rel}=1.2\%$		



No. CNAS L0128

第 89 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		Harmonic、 Non- Harmonic、 Inter- Harmonic Distortion		(-90~-20)dBc(0.1MHz~ 13GHz)	<i>U</i> =1.6dB		
14	*Low Frequency Signal Generator	Frequency	V.R. of Low Frequency Signal Generator JJG 602	1Hz~1MHz	<i>U</i> _{rel} =1×10 ⁻⁵		
		Distortion		(0.001 ~30)% (5Hz ~ 100kHz)	<i>U</i> =1dB		
		Voltage		1mV ~10mV (10Hz~ 100kHz)	<i>U</i> _{rel} =2%		
				10mV~300V (10Hz~ 100kHz)	<i>U</i> _{rel} =0.16%		
				1mV ~30V (100kHz~ 1MHz)	<i>U</i> _{rel} =0.9%		
				30V~300V (100kHz~ 1MHz)	<i>U</i> _{rel} =3%		
		Attenuation		0~100dB (10Hz~1MHz)	<i>U</i> =0.2dB		
15	*Television Video Generator	Luminance Amplitude	C.S. for Television Video Generator JJF 1235	60mV~1V	<i>U</i> _{rel} =0.6%		
		Chrominance Amplitude		60mV~1V	<i>U</i> _{rel} =1.2%		
		Chrominance Phase		(0~360)°	<i>U</i> =0.6°		
		Pulse Width		1ns~30 μ s	<i>U</i> _{rel} =0.4%		
16	*Function generator	Frequency	V.R. of Function Generators JJG840	0.1Hz~240MHz	<i>U</i> _{rel} =1.0×10 ⁻⁸		



No. CNAS L0128

第 90 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date	
		Voltage (amplitude of output waveform)	ilac-M	Peak-Peak value: 3mV ~30V、 1kHz (50 Ω Load measurement)	<i>U</i> _{rel} =0.24%			
				Peak-Peak Value: 3mV ~30V、 1kHz (Open circuit measurement)	<i>U</i> _{rel} =0.084%			
				DC voltage (1mV~5V)	<i>U</i> _{rel} =0.1%			
		Voltage (sine wave amplitude flatness)		1V (10Hz~100kHz)	<i>U</i> =0.02 dB			
				1V (100kHz~200kHz)	<i>U</i> =0.05dB			
				1V (200kHz~240MHz)	<i>U</i> =0.18dB			
		Distortion		(0.01~30)% (5Hz~ 20kHz)	<i>U</i> _{rel} =13%			
				(0.01~30)% (20kHz~ 50kHz)	<i>U</i> _{rel} =26%			
				(0.01~30)% (50kHz~ 150kHz)	<i>U</i> _{rel} =41%			
		Harmonic distortion		(-90~-20)dBc (150kHz~240MHz)	<i>U</i> =2.1dB			
		Rise time		135ps~100 μ s	<i>U</i> =0.2ns			
		square wave pulse duty cycle		(0.1~99.9)%	<i>U</i> _{rel} =0.7%			
17	*Low Power Mount	VSWR	Verification regulation for lower power mount GJB/J 3598	(1~10) (50MHz~2GHz)	<i>U</i> _{rel} =3.5%			
				(1~10) (2GHz~8GHz)	<i>U</i> _{rel} =5.9%			
				(1~10) (8GHz~18GHz)	<i>U</i> _{rel} =7.2%			



No. CNAS L0128

第 91 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		CAL Factor		(10%~200%) (10MHz~2GHz)	$U=2.1\%$		
				(10%~200%) (2GHz~6GHz)	$U=2.4\%$		
				(10%~200%) (6GHz~12.4GHz)	$U=3.1\%$		
				(10%~200%) (12.4GHz~15GHz)	$U=3.4\%$		
				(10%~200%) (15GHz~18GHz)	$U=4.1\%$		
18	*FM Stereo Signal Generator	Frequency	XT-24 V.R. of Stereo Signal Generator JJG(DZ)12011	10Hz~100kHz	$U_{rel}=1 \times 10^{-8}$		
		Stereo Separation		(30~60) dB (50Hz~15kHz)	$U_{rel}=0.4\%$		
		Distortion		(-10~-100)dB(5Hz ~150kHz)	$U=1.2\text{dB}$		
19	*Audio Analyzer	Output Frequency	C.S. for Audio Analyzer JJF 1395	5Hz~150kHz	$U_{rel}=1 \times 10^{-7}$		
		Output Voltage		10mV~10V (1kHz)	$U_{rel}=1.3\%$		
		Source Distortion		(-10~-90)dB (10Hz~100kHz)	$U=0.6\text{dB}$		
		Voltage Measurement		10mV~100V (1kHz)	$U_{rel}=0.66\%$		
		Distortion Measurement		(10~100)% (10Hz,100Hz,1kHz,10kHz)	$U_{rel}=1.3\%$		
				(3~10)%(10Hz,100Hz,1kHz,10kHz)	$U_{rel}=0.88\%$		



No. CNAS L0128

第 92 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				(1~3)%, (10Hz, 100Hz, 1kHz, 10kHz)	$U_{rel}=1.3\%$		
				(0.3~1)%, (10Hz, 100Hz, 1kHz, 10kHz)	$U_{rel}=0.88\%$		
				(0.10~0.30)%, (10Hz, 100Hz, 1kHz, 10kHz)	$U_{rel}=1.4\%$		
				(0.03~0.10)%, (10Hz, 100Hz, 1kHz, 10kHz)	$U_{rel}=1.0\%$		
				(0.01~0.03)%, (10Hz, 100Hz, 1kHz, 10kHz)	$U_{rel}=1.6\%$		
				(0.003~0.01)%(10Hz,100Hz,1kHz,10kHz)	$U_{rel}=1.3\%$		
				(10~100)%, (20kHz, 100kHz)	$U_{rel}=1.6\%$		
				(3~10)%, (20kHz, 100kHz)	$U_{rel}=1.3\%$		
				(1~3)%, (20kHz, 100kHz)	$U_{rel}=1.6\%$		
				(0.30~1)%, (20kHz, 100kHz)	$U_{rel}=1.3\%$		
				(0.10~0.30)%, (20kHz, 100kHz)	$U_{rel}=1.6\%$		
				(0.03~0.10)%, (20kHz, 100kHz)	$U_{rel}=2.4\%$		



No. CNAS L0128

第 93 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
				(0.01~0.03)%, (20kHz, 100kHz)	<i>U</i> _{rel} =3.6%		
				(0.10~100)%, (150kHz, 200kHz)	<i>U</i> _{rel} =3.6%		
				(0.03~0.10)%, (150kHz, 200kHz)	<i>U</i> _{rel} =5.8%		
				(0.01~0.03)%, (150kHz, 200kHz)	<i>U</i> _{rel} =9.3%		
20	*Video Signal measure meter	Luminance Level	C.S.for Television Video Signal Analyzer JJF 1455	140mV ~2V	<i>U</i> _{rel} =0.6%		
		Chrominance Level		(1~714) mV	<i>U</i> _{rel} =0.6%		
		Phase		Vector phase: (0~360) °	<i>U</i> =0.6°		
		Time		1 μ s~20 μ s	<i>U</i> =10ns		
		Distortion		Difference Gain Distortion: (1~30)%	<i>U</i> =0.4%		
				Difference Phase Distortion: (0~30)°	<i>U</i> =0.4°		
21	*Analogue Oscilloscope	Time	V.R. of Analogue Oscilloscope JJG 262	(500ps~55s)/div	<i>U</i> _{rel} =0.3%		
		Rise Time		150ps~50ns	<i>U</i> =25ps		
		Voltage		(1mV ~5V)/div (1kHz)	<i>U</i> _{rel} =0.4%		
		Bandwidth		50kHz~6.0GHz	<i>U</i> _{rel} =5%		
22	Oscilloscope Calibrator	voltage	V.R.of Oscilloscope Calibrators JJG 278	10mV~200V	<i>U</i> _{rel} =0.1%		
		Sine-wave flatness		10mV~1V (50kHz~ 100kHz)	<i>U</i> _{rel} =1.0%		



No. CNAS L0128

第 94 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				(0~+20)dBm (100kHz~3.2GHz)	$U=0.30\text{dB}$		
		Rise time		300ps~1ns	$U_{\text{rel}}=4.0\%$		
		Time		2ns~0.5s	$U_{\text{rel}}=0.0012\%$		
23	*Distortion Meter	Distortion	Calibration Specification for Distortion Meters JJF 1852	(10~100)%, (10Hz, 100Hz, 1kHz, 10kHz)	$U_{\text{rel}}=1.3\%$		
				(3~10)%, (10Hz, 100Hz, 1kHz, 10kHz)	$U_{\text{rel}}=0.88\%$		
				(1~3)%, (10Hz, 100Hz, 1kHz, 10kHz)	$U_{\text{rel}}=1.3\%$		
				(0.3~1)%, (10Hz, 100Hz, 1kHz, 10kHz)	$U_{\text{rel}}=0.88\%$		
				(0.10~0.30)%, (10Hz, 100Hz, 1kHz, 10kHz)	$U_{\text{rel}}=1.4\%$		
				(0.03~0.10)%, (10Hz, 100Hz, 1kHz, 10kHz)	$U_{\text{rel}}=1.0\%$		
				(0.01~0.03)%, (10Hz, 100Hz, 1kHz, 10kHz)	$U_{\text{rel}}=1.6\%$		
				(0.003~0.01)%, (10Hz, 100Hz, 1kHz, 10kHz)	$U_{\text{rel}}=1.3\%$		
				(10~100)%, (20kHz, 100kHz)	$U_{\text{rel}}=1.6\%$		



No. CNAS L0128

第 95 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				(3~10)%, (20kHz, 100kHz)	$U_{rel}=1.3\%$		
				(1~3)%, (20kHz, 100kHz)	$U_{rel}=1.6\%$		
				(0.30~1)%, (20kHz, 100kHz)	$U_{rel}=1.3\%$		
				(0.10~0.30)%, (20kHz, 100kHz)	$U_{rel}=1.6\%$		
				(0.03~0.10)%, (20kHz, 100kHz)	$U_{rel}=2.4\%$		
				(0.01~0.03)%, (20kHz, 100kHz)	$U_{rel}=3.6\%$		
				(10~100)%, (150kHz, 200kHz)	$U_{rel}=3.6\%$		
				(3~10)%, (150kHz, 200kHz)	$U_{rel}=3.5\%$		
				(1~3)%, (150kHz, 200kHz)	$U_{rel}=3.6\%$		
				(0.30~1)%, (150kHz, 200kHz)	$U_{rel}=3.5\%$		
				(0.10~0.30)%, (150kHz, 200kHz)	$U_{rel}=3.6\%$		
				(0.03~0.10)%, (150kHz, 200kHz)	$U_{rel}=5.8\%$		
				(0.01~0.03)%, (150kHz, 200kHz)	$U_{rel}=9.3\%$		
		Voltage		1mV~300V, (1kHz)	$U_{rel}=0.66\%$		



No. CNAS L0128

第 96 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				1V, (10Hz~20kHz)	$U_{rel}=0.66\%$		
				1V, (20kHz~200kHz)	$U_{rel}=1.4\%$		
24	Radio-frequency coupling decoupling device	Common Mode Impedance	Electromagnetic compatibility Testing and measurement techniques-Immunity to conducted disturbances, induced by radio-frequency fields (part 6.) GB/T 17626.6	25 Ω ~ 300 Ω (150kHz~230MHz)	$U_{rel}=4\%$		
		Attenuation(Coupling)		0dB~60dB (150kHz~230MHz)	$U=0.5\text{dB}$		
25	Curve Tracer	Voltage(X-axis)	Calibration Specification for Semiconductor Device Curve Tracers JJF 1236	(0.1~1000)V	$U_{rel}=0.8\%$		
		Current(Y-axis)		5 μA ~ 20A	$U_{rel}=0.8\%$		
		Voltage (output)		(0.1~20) V	$U_{rel}=0.2\%$		
		resistance		100 Ω ~ 10M Ω	$U_{rel}=0.2\%$		
26	*Network Cable Analyzers	DC Loop Resistance	Calibration Specification for Network Cable Analyzers JJF 1494	(0~50) Ω	$U=0.08\Omega$		
		Delay Time		(50~500)ns	$U=3.0\text{ns}$		
		Insertion Loss		(0~40)dB	$U=0.16\text{dB}$		
		Near-end Crosstalk		(20~70)dB	$U=0.39\text{dB}$		
		Equivalent Far-end Crosstalk		(10~60)dB	$U=0.46\text{dB}$		
		Return Loss		(8~26)dB	$U=0.06\text{dB}$		



No. CNAS L0128

第 97 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
27	*Digital Oscilloscope	Voltage	Digital Oscilloscope JJF 1057,Verification regulation for digital oscilloscope GJB 7691	(1mV~10V)/div (1kHz)	<i>U</i> _{rel} =0.4%		
		time		(500ps~55s)/div	<i>U</i> _{rel} =0.3%		
		Time base		(500ps~55s)/div	<i>U</i> _{rel} =2.6×10 ⁻⁶		
		Bandwidth		50kHz ~6.0GHz	<i>U</i> _{rel} =6.0%		
		Rise Time		65ps~0.1ms	<i>U</i> _{rel} =4.0%		
28	Microwave Radiation and Leakage Energy Measuring Instruments	Power density	Microwave Radiation and Leakage Energy Measuring Instruments JJG 776	0.1mW/cm ² ~3mW/cm ² (2.45GHz)	<i>U</i> =1.3dB		
29	*Artificial Mains Network	Modulus of Impedance	Calibration Specification for Artificial Mains Networks JJF 1705	(0.1~100)Ω , (9kHz~108MHz)	<i>U</i> _{rel} =6%		
		Phase Angle of Impedance		(0.1~360)° , (9kHz~108MHz)	<i>U</i> _{rel} =6%		
		Voltage division factor		(0.01~20)dB , (9kHz~108MHz)	<i>U</i> =0.3dB		
30	*Microwave Power Amplifier	Rated output power	V.R. of RF and Microwave Power Amplifiers JJF 1678	P:(1W~1000W) (9kHz~3GHzHz)	<i>U</i> =0.32dB		
				P:(1W~500W) (3GHz~10GHz)	<i>U</i> =0.41dB		
				P:(1W~150W) (10GHz~18GHz)	<i>U</i> =0.41dB		
		Gain		(0~80)dB (9kHz~240MHz)	<i>U</i> =0.32dB		
				(0~80)dB (240MHz~18GHz)	<i>U</i> =0.41dB		



No. CNAS L0128

第 98 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		Gain flatness	ilac-M	±(0.5～12)dB (240MHz～18GHz)	<i>U</i> =0.32dB		
				±(0.5～12)dB (240MHz～18GHz)	<i>U</i> =0.41dB		
		1dB compression point output power		P:(1W～1000W) (9kHz～3GHz)	<i>U</i> =0.32dB		
				P:(1W～500W) (3GHz～10GHz)	<i>U</i> =0.41dB		
				P:(1W～150W) (10GHz～18GHz)	<i>U</i> =0.41dB		
		Maximum output power		P:(1W～1000W) (9kHz～3GHz)	<i>U</i> =0.32dB		
				P:(1W～500W) (3GHz～10GHz)	<i>U</i> =0.41dB		
				P:(1W～150W) (10GHz～18GHz)	<i>U</i> =0.41dB		
		Harmonic distortion and clutter suppression		(10～78)dBc (9kHz～26.5GHz)	<i>U</i> =1.6dB		
		VSWR of Input voltage		VSWR:(1.05～3) (30kHz～2GHz)	<i>U</i> _{rel} =3.5%		
				VSWR:(1.05～3) (2GHz～9GHz)	<i>U</i> _{rel} =4.7%		
				VSWR:(1.05～3) (9GHz～18GHz)	<i>U</i> _{rel} =5.9%		
31	*Directional Coupler	Port return loss	Calibration specification for directional coupler and standing wave ratio bridge	(15～60)dB (300kHz～18GHz)	<i>U</i> =0.2dB		



No. CNAS L0128

第 99 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Insertion Loss	JJF 1680	(0.01~3)dB (300kHz~18GHz)	$U=0.12\text{dB}$		
		Coupling Coefficient		(10~40)dB (300kHz~18GHz)	$U=0.2\text{dB}$		
		Directional		(20~50)dB (300kHz~18GHz)	$U=0.3\text{dB}$		
32	RF Communication Test Set	reference frequency	Calibration Specification for RF Communication Test Set JJF 1065	10MHz	$U_{\text{rel}}=1 \times 10^{-9}$		
		input frequency		10kHz~10MHz	$U_{\text{rel}}=2 \times 10^{-6}$		
				10MHz~1000MHz	$U_{\text{rel}}=2 \times 10^{-9}$		
		span		10kHz~1000MHz	$U_{\text{rel}}=1.5\%$		
		RBW		100Hz~1MHz	$U_{\text{rel}}=1.5\%$		
		reference level		(-70~+20)dBm	$U=0.3\text{dB}$		
		input level		(0~100)dB	$U=0.3\text{dB}$		
		input frequency response		(-70~20)dBm (100kHz~1000MHz)	$U=0.2\text{dB}$		
		input attenuation		(0~70)dB	$U=0.2\text{dB}$		
		DANL		(-160~-114)dBm	$U=3\text{dB}$		
		residuals		(-160~-90)dBm	$U=3\text{dB}$		
		OBW		(5~200)kHz	$U_{\text{rel}}=1.5\%$		
		ACP		(-70~0)dBc	$U=3\text{dB}$		



No. CNAS L0128

第 100 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		output frequency		250kHz~1000MHz	$U_{rel}=6 \times 10^{-9}$		
		output level		(-137~+19)dBm	$U=0.6\text{dB}$		
		harmonic distortion		(-80~-30)dBc	$U=1.6\text{dB}$		
		non-harmonic distortion		(-90~-60)dBc	$U=1.6\text{dB}$		
		SSB noise		(-140~-70)dBc/Hz(45~900)MHz	$U=3\text{dB}$		
		AM depth		(0.1~99)%(CF: 450MHz;fm: 1kHz)	$U_{rel}=3\%$		
		FM deviation		1Hz~400kHz(f: 45MHz, 450MHz, 900MHz)	$U_{rel}=1.2\%$		
		RF power		1μW~60W(10~900)MHz	$U_{rel}=4\%$		
		AM Measure		(0.1~95)%(carrier frequency: 5MHz~1000MHz, modulation frequency: 1kHz)	$U_{rel}=3\%$		
		FM Measure		20Hz~75kHz(carrier frequency: 5MHz~1000MHz, modulation frequency: 1kHz)	$U_{rel}=1.2\%$		
		output voltage		0.1mV~4V(DC~25kHz)	$U_{rel}=0.1\%$		
		Output frequency response		20Hz~25kHz	$U=0.01\text{dB}$		



No. CNAS L0128

第 101 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		output distortion	ilac-MEA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	0.0001%~0.13%	$U=1.5\text{dB}$		
		AF input frequency		20Hz~400kHz	$U_{\text{rel}}=1 \times 10^{-6}$		
		input voltage		20mV~30V(20Hz~25kHz)	$U_{\text{rel}}=0.05\%$		
		frequency response		20Hz~25kHz	$U=0.02\text{dB}$		
		bandwidth		20Hz~50kHz	$U_{\text{rel}}=6\%$		
		voltage		1mV/div~10V/div	$U_{\text{rel}}=0.4\%$		
		time		10 $\mu\text{s/div}$ ~100ms/div	$U_{\text{rel}}=0.3\%$		
33	*Charge Amplifiers	Charge	Verification Regulation of Charge Amplifiers JJG 338	($10 \sim 10^5$)pC	$U_{\text{rel}}=0.05\%$		
		Voltage		10mV~10V, (10Hz~100kHz)	$U_{\text{rel}}=0.08\%$		
34	*Dynamical Signal Analyzer	Frequency	Verification Regulation of Dynamical Signal Analyzer JJG 834	0.1Hz~200kHz	$U_{\text{rel}}=3.4 \times 10^{-5}$		
		Voltage		1mV~10V, (10Hz~200kHz)	$U_{\text{rel}}=0.4\%$		
35	Absorbing Clamp	Absorption attenuation	Calibration Specification for Absorbing Clamp in the Range of 30MHz to 1.0GHz JJF 1155	(14~22)dB, (30MHz~1GHz)	$U=1.8\text{dB}$	合格评定国家认可委员会 认可证书专用章	
36	*Antenna feeder tester	Frequency	Calibration specification for antenna feeder tester JJF 1740	2MHz~18GHz	$U_{\text{rel}}=6 \times 10^{-9}$		
		Power		-30dBm~20dBm, (2MHz~18GHz)	$U=0.30\text{dB}$		
		VSWR		1~10, (2MHz~18GHz)	$U_{\text{rel}}=5.1\%$		



No. CNAS L0128

第 102 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		Attenuation		0dB~30dB, (2MHz~18GHz)	<i>U</i> =0.20dB		
		Fault location distance		15m, (2MHz~18GHz)	<i>U</i> _{rel} =1.1%		
		RF power		-50dBm~10dBm, (2MHz~18GHz)	<i>U</i> =0.18dB		
37	*Waveform Recorder	DC voltage	Calibration Specification for Waveform Recorders JJF1876	±(10mV~300mV)	<i>U</i> _{rel} =0.4%		
				±(0.3V~3V)	<i>U</i> _{rel} =0.05%		
				±(3V~30V)	<i>U</i> _{rel} =0.01%		
				±(30V~1000V)	<i>U</i> _{rel} =0.06%		
		AC voltage		10mV~33mV, (10Hz~10kHz)	<i>U</i> _{rel} =4%		
				33mV~330mV, (10Hz~10kHz)	<i>U</i> _{rel} =0.4%		
				0.33V~3.3V, (10Hz~10kHz)	<i>U</i> _{rel} =0.05%		
				3.3V~33V, (10Hz~10kHz)	<i>U</i> _{rel} =0.02%		
				33V~330V, (10Hz~10kHz)	<i>U</i> _{rel} =0.03%		
				330V~1000V, (10Hz~10kHz)	<i>U</i> _{rel} =0.04%		
		Time base		1 μ s/div~10h/div	<i>U</i> _{rel} =3×10 ⁻⁷		
		Frequency Bandwidth		10Hz~2MHz	<i>U</i> _{rel} =1%		
		Resistance		50 Ω ~1M Ω	<i>U</i> _{rel} =0.01%		



No. CNAS L0128

第 103 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date		
		DC gain		0.5～1.5	<i>U</i> _{rel} =0.4%				
		Record Time		(1～1000)s	<i>U</i> _{rel} =3×10 ⁻⁷				
38	*High Voltage Dielectric Loss Tester	capacitance	V.R.of High Voltage Dielectric Loss Tester JJG 1126	100pF、1000pF、10nF	<i>U</i> _{rel} =0.14%				
				100nF	<i>U</i> _{rel} =0.16%				
				500nF	<i>U</i> _{rel} =0.11%				
		loss factor		0%～0.2%(100pF、1000pF)	<i>U</i> =0.003%				
				0.5%(100pF、1000pF)	<i>U</i> =0.004%				
				1%(100pF、1000pF)	<i>U</i> =0.005%				
				2%(100pF、1000pF)	<i>U</i> =0.007%				
				5%(100pF、1000pF)	<i>U</i> =0.013%				
				10%(100pF、1000pF)	<i>U</i> =0.023%				
				0.1%～0.2%(10nF、100nF、500nF)	<i>U</i> =0.004%				
				0.5%(10nF、100nF、500nF)	<i>U</i> =0.006%				
				1%(10nF、100nF、500nF)	<i>U</i> =0.008%				
				2%(10nF、100nF、500nF)	<i>U</i> =0.013%				
				5%(10nF、100nF、500nF)	<i>U</i> =0.028%				
				10%(10nF、100nF、500nF)	<i>U</i> =0.053%				



No. CNAS L0128

第 104 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
39	*RF and Microwave Power Meters	Output power of reference source	Calibration Specification for RF and Microwave Power Meters JJF 1885	1mW（50MHz）	<i>U</i> _{rel} =1.0%	5.4.1 Alternating comparison method is NOT suitable for this item	
		Calibration factor		0.5~2（Frequency range: 0.1MHz~2GHz）	<i>U</i> _{rel} =1.4%		
				0.5~2（Frequency range: 2GHz~6GHz）	<i>U</i> _{rel} =1.8%		
				0.5~2（Frequency range: 6GHz~12.4GHz）	<i>U</i> _{rel} =2.0%		
				0.5~2（Frequency range: 12.4GHz~15GHz）	<i>U</i> _{rel} =2.4%		
				0.5~2（Frequency range: 15GHz~18GHz）	<i>U</i> _{rel} =2.5%		
		Power Linearity		Power range: -20dBm~10dBm（50MHz or 1GHz）	<i>U</i> =0.04dB		
		VSWR		1~2（Frequency range: 0.1MHz~2GHz）	<i>U</i> _{rel} =3.5%		
				1~2（Frequency range: 2GHz~8GHz）	<i>U</i> _{rel} =5.9%		
				1~2（Frequency range: 8GHz~18GHz）	<i>U</i> _{rel} =7.2%		



No. CNAS L0128

第 105 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
40	coupling decoupling device	Attenuation(Coupling)	Road vehicles -- Component test methods for electrical disturbances from narrowband radiated electromagnetic energy -- Part 4: Harness excitation methods (part 9.3.1.2) ISO 11452-4	0dB~60dB (9kHz~400MHz)	<i>U</i> =0.5dB		
六、Time and Frequency							
1	Microwave frequency meter	Frequency	Verification regulations for microwave frequency counters JJG 841	100kHz~20GHz	<i>U</i> _{rel} =2×10 ⁻⁶ ~5×10 ⁻¹¹		
2	Electronic stopwatch	Time interval	verification regulation of stopwatches JJG 237	(1~3600)s	<i>U</i> =0.01s		
		Day Error		(-20.00~20.00)s	<i>U</i> =0.06s		
3	Mechanical stopwatch	Time interval	verification regulation of stopwatches JJG 237	(1~1800)s	<i>U</i> =0.1s		
4	*Electronic time relay	Time	Calibration Specification for Electronic time relay JJF 1282	0.001s~24h	<i>U</i> = (3.5×10 ⁻⁵ <i>R</i> +0.003s)		
5	*Pulse Counters	frequency	Calibration Specification for Pulse Counters JJF 1686	10MHz	<i>U</i> _{rel} =6.8×10 ⁻⁹		
		Pulse Count		0~1000000	<i>U</i> =1		
6	Universal counter	frequency	V.R.of Universal Counters JJG 349	0.1Hz~18GHz	<i>U</i> _{rel} =2×10 ⁻⁶ ~5×10 ⁻¹¹		
		cycle		1ns~100s	<i>U</i> _{rel} =2×10 ⁻⁶ ~5×10 ⁻¹¹		
		Time interval		1ms~10s	<i>U</i> =3×10 ⁻⁸ s		
七、Optic							



No. CNAS L0128

第 106 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
1	Illuminance meter	Illuminance	Verification Regulation of Illuminance Meter JJG 245	(4~3000)lx	$U_{rel}=1.6\%$		
		Cosine characteristics (directional response)		(0~10)%	$U=1.0\%$		
2	Ultraviolet Irradiance Meters	Radiosity	Verification Regulation of Ultraviolet Irradiance Meters JJG 879	(55~2000) $\mu W/cm^2$	$U_{rel}=16\%$		
				(2~80)mW/cm ²	$U_{rel}=18\%$		
3	Luminance Meter	Luminance	Verification Regulation of Luminance Meter JJG 211	(0.5~1000)cd/m ²	$U_{rel}=3.0\%$		
		chrominance		x、y: 全色域	$U(x)=0.0092, U(y)=0.0072$		
4	*Standard Light Sources Boxes	Illuminance	Calibration Specification for Standard Light Sources Boxes JJF(FZ) 055	(50~3000)lx	$U=47lx$		
		Correlated Colour Temperature		(2700~7500)K	$U=1.2 \times 10^2 K$		
5	Lamp of Luminous Intensity	Luminance	Verification Regulation of Luminance Meter JJG 246	(5~150)cd	$U_{rel}=3.0\%$		
				(150~750)cd	$U_{rel}=2.6\%$		
				(750~1500)cd	$U_{rel}=2.9\%$		
6	*Spectral radiation system/Spectrometer	Wavelength	Verification Regulation of Analyzer of Spectrum Radiation JJG (Zhe) 85	(200~1500)nm	$U=0.45nm$		
		Color temperature distribution		(500~9000)K	$U=22K$		
		Chromaticity coordinate		Full color gamut	$U=0.0025$		



No. CNAS L0128

第 107 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
7	Lamp for Total Luminous Flux	Total Luminous Flux	Verification Regulation of Standard Incandescent Lamp for Total Luminous Flux JJG 247	10lm~23500lm	$U_{rel}=2.2\%$		
8	*Artificial Accelerated Weathering Apparatus of Xenon Arc Lamp	irradiance	Calibration Specification for Irradiance of Artificial Accelerated Weathering Apparatus of Xenon Arc Lamp JJF 1525	$0.1 \text{ mW} \cdot \text{m}^{-2} \sim 1 \text{ kW} \cdot \text{m}^{-2}$, (300nm~800nm)	$U_{rel}=11\%$	spectroradiometer method	
9	*Abbe Refractometer	index of refraction	Verification Regulation of Abbe Refractometer JJG625	(1.3300~1.8000)nD	$U=9.2 \times 10^{-5} \text{ nD}$		
10	*Colorimeters and Color Difference Meters	Colourimetric	Verification Regulation of Colorimeters and Color Difference Meters JJG595	Y:1.0~100.0			
				x,y:full color gamut	$U=0.0053(0/d); U=0.0070(d/0)$		
11	*Specular Gloss Meters	Gloss	Verification Regulation of Specular Gloss Meters and Gloss Plates JJG696	(0~120.0) gloss unit	$U=1.5 \text{ GU}$		
12	*Hazemeter	Haze	Calibration Specification for Hazemeter JJF1303	0.1~30	$U=0.31$		
		Transmittance		(1~100)%	$U=0.8\%$		
13	*Color Light-box with Standard Light Sources	Luminance	Calibration Specification for Color Light-box with Standard Light Sources JJF (Min) 1062	(50~30000)cd/m ²	$U_{rel}=8\%$		
		Colour coordinates		Full color gamut	$U=0.0035$		
14	*solar simulators	spectral match	Calibration Specification for Solar Simulators JJF 1615	(400~1100)nm	$U_{rel}=8\%$		
		irradiance non-uniformity		0.1%~50%	$U_{rel}=3\%$		
		irradiance instability		0.1%~50%	$U_{rel}=2.1\%$		



No. CNAS L0128

第 108 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
15	*Whiteness Meters	Whiteness	Verification Regulation of Whiteness Meters JJG 512	1.0~100.0	(0/d): $U_{R457}=1.0$, (d/0): $U_{R457}=1.0$, (45/0): $U_{R457}=1.3$		
16	*Lasers for Medicine	laser power	Verification Regulation of Lasers for Medicine JJG 581	1mW~130W	$U_{rel}=6.0\%$		
		Laser energy		20mJ~10J	$U_{rel}=6.0\%$		
17	*Reflection Densitometers	Optical density	Calibration Specification for Reflection Densitometers JJF 1492	0.00~1.00	$U=0.03$		
				1.01~2.00	$U=0.04$		
18	*Calibration Specification for UV	Peak wavelength	Calibration Specification of UV Analyzer JJF 1936	(250~380) nm	$U=0.7nm$	Only test reflection	
		UV irradiance		(40~1000) μ W/cm ²	$U_{rel}=11\%$		
八、Chemistry							
1	*Ultraviolet, Visible, Near-Infrared Spectrophotometers	Wavelength	V.R. of Ultraviolet, Visible, Near-Infrared Spectrophotometers JJG 178	Grating-type instrument Band A: (190~340) nm	$U=0.4nm$		
				Grating-type instrument Band B: (340~900) nm	$U=0.4nm$		
				Prism-type instrument Band B: (340~900) nm	$U=0.8nm$		
				Grating-type instrument Band C: (900~2600) nm	$U=0.6nm$		
		Transmittance		A 段: (0~100) %	$U=0.3\%$		
				B 段: (0~100) %	$U=0.3\%$		
2	*Chlorine Alarm Detectors	Concentration	Calibration Specification for Chlorine Alarm Detectors JJF1433	(0.1~100) μ mol/mol	$U_{rel}=2.4\%$		
		time		(1~3600) s	$U=2s$		



No. CNAS L0128

第 109 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
3	*Laboratory pH Meters	pH	Verification Regulation of Laboratory pH Meters JJG119	Electricity pH: (0~14)	$U=0.001$		
				Instrument pH: (3~10)	$U=0.004$		
		Voltage		(-1900~1900) mV	$U=0.04\%FS$		
4	*Particulate Analyzer	Particle concentration	Calibration Specification for Particulate Analyzer JJF 1290	(10~10000) particles/mL	$U_{rel}=7\%$		
5	*Electrolytic Conductivity Meters	Electrolytic Conductivity	Verification Regulation of Electrolytic Conductivity Meters JJG376	Electricity: (1~2×10 ⁴) μS/cm	$U_{rel}=0.08\%$		
				Electricity: (5×10 ⁻² ~1) μS/cm	$U_{rel}=0.13\%$		
				Instrument: (100~1.3×10 ⁴) μS/cm	$U_{rel}=0.33\%$		
6	*Sulfur Dioxide Gas Detectors	Concentration	Verification Regulation of Sulfur Dioxide Gas Detectors JJG551	(1~300) μmol/mol	$U_{rel}=2.2\%$		
		time		(1~3600) s	$U=2s$		
7	*Gas Chromatograph	Detection limit	V.R.of Gas Chromatograph JJG700	FID: ≤5×10 ⁻¹⁰ g/s	$U_{rel}=4\%$		
				FPD: ≤5×10 ⁻¹⁰ g/s (S), , ≤1×10 ⁻¹⁰ g/s (P)	$U_{rel}=3\%(S), U_{rel}=4\%(P)$		
				NPD: ≤5×10 ⁻¹² g/s(N) ≤1×10 ⁻¹¹ g/s(P)	$U_{rel}=4\%$		
				ECD: ≤5×10 ⁻¹² g/mL	$U_{rel}=4\%$		
		sensitivity		TCD: ≥800mV•mL/mg	$U_{rel}=4\%$		
8	*Dynamic Light Scattering Particle Size Analyzers	Particle size	Verification Regulation of Dynamic Light Scattering Particle Size Analyzers JJG	(0.1~100) nm	$U_{rel}=5\%$		
				(200~300) nm	$U_{rel}=4\%$		



No. CNAS L0128

第 110 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			1104	(400~500) nm	$U_{rel}=4\%$		
				(600~800) nm	$U_{rel}=4\%$		
9	*Liquid Chromatographs	minimum detectable concentration	V.R.of Liquid Chromatographs JJG705	UV-VIS: $\leq 5 \times 10^{-8}$ g/mL	$U_{rel}=6\%$		
				FLD: $\leq 5 \times 10^{-9}$ g/mL	$U_{rel}=6\%$		
				RID: $\leq 5 \times 10^{-6}$ g/mL	$U_{rel}=7\%$		
				ELSD: $\leq 5 \times 10^{-6}$ g/mL	$U_{rel}=7\%$		
		Temperature		(0~100) °C	$U=0.3$ °C		
		Flow rate		(0.1~10) mL/min	$U_{rel}=0.3\%$		
10	Rotational Viscometers	Dynamic Viscosity	Verification Regulation of Rotational Viscometers JJG 1002	(50~400) mPa•s	$U_{rel}=1.2\%$		
				(400~6000) mPa•s	$U_{rel}=1.3\%$		
				(6000~20000) mPa•s	$U_{rel}=1.5\%$		
				(20000~40000) mPa•s	$U_{rel}=1.7\%$		
				(40000~90000) mPa•s	$U_{rel}=1.8\%$		
				(90000~125000) mPa•s	$U_{rel}=1.9\%$		
11	*Mercury Analyzers	Detection limit	Verification Regulation of Mercury Analyzers JJG 548	absorption type: ≤ 1.0 ng	$U=0.2$ ng		
				fluorescent type: ≤ 0.1 ng	$U=0.02$ ng		
12	*Polarimeter	optical activity	Verification Regulation of Polarimeter and Saccharimeter JJG536	-35° ~+35°	$U=0.005$ °		



No. CNAS L0128

第 111 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
13	Wood Moisture Content Measuring Meters	Moisture Content	Verification Regulation of Wood Moisture Content Measuring Meters JJG 986	6%~28%	$U=0.3\%$		
14	*Measurement Instruments	Temperature	V. R. of Melting-point Measurement Instruments JJG701	(50~300) °C (heating rate: 0.2°C/min)	$U=0.16^{\circ}\text{C}$		
				Capillary method: (50~300) °C (heating rate: 1.0°C/min)	$U=0.26^{\circ}\text{C}$		
				Hot table method: (50~300) °C (heating rate: 1.0°C/min)	$U=0.12^{\circ}\text{C}$		
15	*Gas Chromatography-Mass Spectrometers	Signal-to-noise ratio	Calibration Specification for Gas Chromatography-Mass Spectrometers JJF1164	Ion trap, single quadrupole, triple quadrupole: $\geq 10:1$	$U_{\text{rel}}=6\%$		
				Time of flight, electrostatic field of orbital hydrazine: $\geq 50:1$	$U_{\text{rel}}=6\%$		
16	*Atomic Absorption Spectrophotometers	Detection limit	Verification Regulation for Atomic Absorption Spectrophotometers JJG 694	Cu: (0~5) $\mu\text{g/mL}$	$U=0.005 \mu\text{g/mL}$		
				Cd: (0~50) pg	$U=0.2\text{pg}$		
17	*Turbidimeters	turbidity	Verification Regulation of Turbidimeters JJG880	(0.001~400) NTU	$U_{\text{rel}}=4\%$		
18	*Air Samplers	flow	Verification Regulation of Air Sampler JJG956	(50~6000) mL/min	$U_{\text{rel}}=1.2\%$		
		time		(300~3600) s	$U_{\text{rel}}=0.01\%$		



No. CNAS L0128

第 112 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
19	*Alarmer Detectors of Combustible Gas	Gas Concentration	Alarmer Detectors of Combustible Gas JJG693	CH ₄ (0.01~95) %LEL	$U_{rel}=2.7\%$		
				C ₃ H ₈ (0.01~95) %LEL	$U_{rel}=2.4\%$		
				i-C ₄ H ₁₀ (0.01~95) %LEL	$U_{rel}=2.1\%$		
				H ₂ (0.01~95) %LEL	$U_{rel}=2.7\%$		
				C ₂ H ₂ (0.01~95) %LEL	$U_{rel}=2.7\%$		
20	Airborne Particle Counter	Particle Size Particle Concentration	Calibration Specification for Airborne Particle Counter JJF1190	0~100) %	$U=6\%$		
		Concentration		(1000~100000) particle/28.3L	$U_{rel}=14\%$		
21	*Instrument for KF Coulometry Titration	water content	Verification Regulation of Instrument for KF Coulometry Titration JJG1044	(10~5000) μ g	$U_{rel}=3.1\%$		
22	*Inductively coupled plasma- atomic emission spectrometry	Detection limit	Verification Regulation of Emission Spectrometer JJG768	(0~50) mg/L	$U=0.00006\text{mg/L(Ba)},$ $U=0.00026\text{mg/L(Cr)},$ $U=0.00028\text{mg/L(Cu)},$ $U=0.00004\text{mg/L(Mn)},$ $U=0.00015\text{mg/L(Ni)},$ $U=0.00013\text{mg/L(Zn)}$		
23	*Liquid chromatography- mass spectrometers	Signal-to- noise ratio	Calibration Specification for Liquid Chromatography - Mass Spectrometers JJF1317	16/2000 Single four- stage rod, Ion Trap: $\geq$ 10:1	$U_{rel}=7\%$		
				Triple quadrupole(ESI-): $\geq 10:1$	$U_{rel}=7\%$		
				Triple quadrupole (ESI+ 、APCI+): $\geq 30:1$	$U_{rel}=7\%$		



No. CNAS L0128

第 113 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
24	*Ion Chromatograph	Minimum detectable concentration	Verification Regulation of Ion Chromatograph JJG823	Electrical Conductivity Detector: $\leq 0.02 \mu\text{g/mL}$	$U_{\text{rel}}=4\%$		
				UV-VIS Detector: $\leq 0.02 \mu\text{g/mL}$	$U_{\text{rel}}=4\%$		
				Electrochemical Detector: $\leq 0.02 \mu\text{g/mL}$	$U_{\text{rel}}=4\%$		
		Temperature		$(0 \sim 50) ^\circ\text{C}$	$U=0.3 ^\circ\text{C}$		
25	*Hand Saccharinmeter	Flow rate	Verification Regulation of Hand Saccharinmeter (Content-meter) and Hand Refractometer JJG820	$(0.1 \sim 10) \text{ mL/min}$	$U_{\text{rel}}=0.3\%$		
		concentration		$(0 \sim 65) \%$	$U=0.3\%$		
26	*Fourier Transform Infrared Spectrometers	wavenumber	Calibration Specification for Fourier Transform Infrared Spectrometers JJF1319	resolution 0.01: $(4000 \sim 400) \text{ cm}^{-1}$	$U=0.6 \text{ cm}^{-1}$		
				resolution 0.5: $(4000 \sim 400) \text{ cm}^{-1}$	$U=0.6 \text{ cm}^{-1}$		
				resolution 1.0: $(4000 \sim 400) \text{ cm}^{-1}$	$U=0.8 \text{ cm}^{-1}$		
				resolution 2.0: $(4000 \sim 400) \text{ cm}^{-1}$	$U=1.2 \text{ cm}^{-1}$		
27	*Atomic Fluorescence Spectrophotometer	Detection limit	V.R. of Atomic Fluorescence Spectrophotometers JJG 939	As: $\leq 0.4 \text{ ng}$	$U=0.006 \text{ ng}$		
				Sb: $\leq 0.4 \text{ ng}$	$U=0.004 \text{ ng}$		
28	*Automatic Potentiometric Titrators	electric potential	V.R. of Automatic Potentiometric Titrator JJG814	$(-1900 \sim 1900) \text{ mV}$	$U=0.6 \text{ mV}$		
		volume		$(2 \sim 100) \text{ mL}$	$U=0.02 \text{ mL}$		
		molar concentration		0.1 mol/L	$U_{\text{rel}}=1.2\%$		



No. CNAS L0128

第 114 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
29	*Direct Reading Spectrometer	Detection limit	Verification Regulation of Emission Spectrometer JJG768	C: ≤0.02%	U=0.0009%		
				Si: ≤0.02%	U=0.0012%		
				Mn: ≤0.02%	U=0.0009%		
				Cr: ≤0.01%	U=0.0009%		
				Ni: ≤0.02%	U=0.0011%		
				V: ≤0.01%	U=0.0004%		
30	Flow Cups Viscosimeter	Viscosity	Verification Regulation for Flow Cups Viscosimeter JJG 743	(10~600) mm²/s	U _{rel} =1.5%		
31	*ELISA Analytical Instruments	absorbance	Verification Regulation of ELISA Analytical Instruments JJG861	0.2~1.5	U=0.020		
		Wavelength		(400~700)nm	U=0.8nm		
32	*Dust Sampler	flow	Veritification Regulation of Dust Sample JJG520	(0~80) L/min	U=1.3%FS		
		time		(1~1800) s	U=0.1s		
33	*Laboratory Ion Meters	pX	Verification Regulation of Laboratory Ion Meters JJG757	Electricity: 0~14	U=0.005		
		potential		Instrument: 2~4	U=0.02		
				(-1900~1900) mV	U=0.4mV		
34	*Differential Scanning Calorimeters	temperature	Veritification Regulation of Differential Scanning Calorimeters JJG936	(150~450) °C	U=0.29°C(In), U=0.13°C(Sn), U=0.47°C(Pb), U=0.61°C(Zn)		
		melting heat		(20~110) J/g	U _{rel} =1.2%(In), U _{rel} =0.4%(Sn), U _{rel} =1.4%(Pb), U _{rel} =1.3%(Zn)		
35	*Inductively coupled plasma massspectrometry	Detection limit	Veritification Regulation of Inductively coupled plasma massspectrometry JJF1159	Be (0~10) μ g/mL	U=0.48ng/L		
				In (0~10) μ g/mL	U=0.08ng/L		

No. CNAS L0128

第 115 页 共 143 页



The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				Bi: (0~10) μ g/mL	$U=0.06\text{ng/L}$		
36	*Analyser of Oxygen-Nitrogen-Hydrogen	content	Verification Regulation of Elemental Analyser JJF1321	O: (0.0005~0.02) %	$U_{\text{rel}}=7.9\%$		
				N: (0.001~0.01) %	$U_{\text{rel}}=8.3\%$		
				H: (1.0~9.0) $\times 10^{-6}$	$U_{\text{rel}}=12\%$		
37	Dissolved Oxygen Meter	The Concentration of Dissolved Oxygen	Verification Regulation of Dissolved Oxygen Meter JJG291	(5~16) mg/L	$U=0.17\text{mg/L}$		
		temperature		(0~50) $^{\circ}\text{C}$	$U=0.16^{\circ}\text{C}$		
38	*Sulfur Hydrogen Gas Detectors	Concentration	Verification Regulation of Sulfur Hydrogen Gas Detectors JJG695	(0.1~100) μ mol/mol	$U_{\text{rel}}=3.5\%$		
39	Pyknometer	Volume	Paints and varnishes- Determination of density- Pyknometer method (appendix A: calibration for pyknometer GB/T 6750	(0~100) mL	$U=0.12\text{mL}$		
40	*Flue Gas Analyzers	Concentration	Verification Regulation of Flue Gas Analyzers JJG968	SO ₂ : (0.1~300) μ mol/mol	$U_{\text{rel}}=2.1\%$		
		Concentration		NO: (0.1~300) μ mol/mol	$U_{\text{rel}}=1.8\%$		
		Concentration		CO: (0.1~1000) μ mol/mol	$U_{\text{rel}}=1.6\%$		
		Concentration		O ₂ : (0.1~25) %	$U_{\text{rel}}=1.9\%$		
		Concentration		NO ₂ : (0.1~100) μ mol/mol	$U_{\text{rel}}=2\%$		



No. CNAS L0128

第 116 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
41	*Carbon Monoxide Detectors	Gas Concentration	Verification Regulation of Carbon Monoxide Detectors JJG915	(0.1~1000) μmol/mol	U _{rel} =1.9%		
42	*Elemental Analyser	content	Verification Regulation of Elemental Analyser JJF1321	N: (0.1~46.6) %	U _{rel} =2.4%;	N Analyser,CH N Analyser,H Analyser	
		C: (40~85) %		U _{rel} =2.4%			
		N: (0.8~8.1) %		U _{rel} =6.2%			
		H: (2~5) %		U _{rel} =3.1 %			
43	*Ammonia Gas Detectors	Concentration	Verification Regulation of Ammonia Gas Detectors JJG1105	(1~300) μmol/mol	U _{rel} =3.1%		
		Time		(1~3600) s	U=3s		
44	*Electrochemical Oxygen Meter	Concentration	Verification Regulation of Electrochemical Oxygen Meter JJG365	(25~100) %	U _{rel} =3.1%		
		Concentration		(0.1~25) %	U _{rel} =3.3%		
		Time		(1~3600) s	U=2s		
45	*Carbon Monoxide and Carbon Dioxide Infrared Gas Analyzer	Concentration	Verification Regulation of Carbon Monoxide and Carbon Dioxide Infrared Gas Analyzer JJG 635	CO: (1~1000) μmol/mol	U _{rel} =1.7%		
		Concentration		CO ₂ : (0.1~100) %	U _{rel} =2.6%		
		Time		(1~3600) s	U=2s		
46	Samplers for Stack Dust	flow rate	Verification Regulation of Samplers for Stack Dust JJG680	(6~100) L/min	U _{rel} =1.4%		
		temperature		Flowmeter Temperature: (10~35) °C	U=0.2 °C		
		temperature		Flue Gas Temperature: (10~300) °C	U=0.3 °C		



No. CNAS L0128

第 117 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		pressure	Verification Regulation of Total Suspended Particulates JJG943	Dynamic Pressure: (0.1~1) kPa	$U=1.2\text{Pa}$		
				Dynamic Pressure: (1~ 2.5) kPa	$U=12\text{Pa}$		
				Flow meter pressure, static pressure: (-100~ 100) kPa	$U=0.06\text{kPa}$		
		time		(1~3600) s	$U=0.3\text{s}$		
47	*Total Suspended Particulates	Flow Rate	Verification Regulation of Total Suspended Particulates JJG943	(80~150) L/min	$U_{\text{rel}}=1.2\%$		
		Flow Rate		(150~1200) L/min	$U_{\text{rel}}=1.3\%$		
		Pressure		(87~105) kPa	$U=2.5\text{hPa}$		
		Temperature		(0~50) °C	$U=0.2^{\circ}\text{C}$		
		Time		(1~3600) s	$U=0.3\text{s}$		
48	*Laser particle size analyzer	particle size	Calibration Specification for Laser Particle Size JJF1211	(1~5) μm	$U_{\text{rel}}=8\%$		
				(5~20) μm	$U_{\text{rel}}=5\%$		
				(20~120) μm	$U_{\text{rel}}=3\%$		
				50.1 μm (玻璃微珠)	$U_{\text{rel}}=6\%$		
49	*Carbon-Sulfur Analyzers	content	Verification Regulation of Carbon-Sulfur Analyzers JJG 395	C: (0.005~0.01) %	$U_{\text{rel}}=5.3\%$	Infrared carbon-sulfur analyzer, Auto- matic high speed carbon-sulfur	
				C: (0.01~0.1) %	$U_{\text{rel}}=3.6\%$		
				C: (0.1~1) %	$U_{\text{rel}}=1.1\%$		
				C: (1~4) %	$U_{\text{rel}}=0.9\%$		



No. CNAS L0128

第 118 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		content		S: (0.003~0.01) %	$U_{rel}=6.0\%$	analyzer	
				S: (0.01~0.1) %	$U_{rel}=4.7\%$		
				S: (0.1~0.2) %	$U_{rel}=3.7\%$		
				C: (0.030~0.1) %	$U_{rel}=5.0\%$		
				C: (0.1~0.5) %	$U_{rel}=2.7\%$		
				C: (0.5~1) %	$U_{rel}=2.7\%$		
				C: (1~4) %	$U_{rel}=1.6\%$		
				S: (0.003~0.01) %	$U_{rel}=6.7\%$		
				S: (0.01~0.05) %	$U_{rel}=5.5\%$		
				S: (0.05~0.1) %	$U_{rel}=4.2\%$		
				S: (0.1~0.2) %	$U_{rel}=4.0\%$		
50	*Flash Point Tester	Flash point	Calibration Specification for Open/Closed Cup Flash Point Testers JJF 1384	Closed cupflash point: (70~110) °C	$U=5.4^{\circ}\text{C}$		
				Closed cupflash point: (110~170) °C	$U=7.7^{\circ}\text{C}$		
				Open cupflash point: (110~250) °C	$U=8.3^{\circ}\text{C}$		
51	*Osmometers	molar concentration	Verification Regulation of Osmometers JJG1089	(90~400) mOsmol/kg	$U=2.4\text{mOsmol/kg}$		
				(400~710) mOsmol/kg	$U_{rel}=0.7\%$		
52	*Thermogravimetric Analyzers	Curie temperature	Verification Regulation for Thermogravimetric Analyzers JJG 1135	(250~500) °C	$U=1.3^{\circ}\text{C}$		
				(50~250) °C	$U=1.1^{\circ}\text{C}$		



No. CNAS L0128

第 119 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Melting point		(500~800) °C	U=2.0°C		
				(50~200) °C	U=0.6°C		
				(200~300) °C	U=0.5°C		
				(300~400) °C	U=0.7°C		
				(400~500) °C	U=0.8°C		
		mass		(1~50) mg	U=0.020mg		
53	*Analyzers for oil content in water	Concentration	Analyzers for Oil Content in water JJG 950	(0.1~10) mg/L	U=0.17mg/L		
				(10~1000) mg/L	U _{rel} =3.1%		
54	*On-line pH Meters	pH	Calibration Specification for On-line pH Meters JJF1547	Electricity pH:(0~14)	U=0.01		
				Instrument pH:(3~10)	U=0.03		
		Voltage		(-1900~1900) mV	U=0.04%FS		
55	*Karl Fischer Volumetric Titrators for Water Content	Water Content	Karl Fischer Volumetric Titrators for Water Content JJG1154	(0.01~1) %	U _{rel} =7.6%		
56	*Residual Chlorine Meters	Concentration	Calibration Specification for Residual Chlorine Meters JJF 1609	Total residual chlorine (0.1~500) mg/L	U _{rel} =3.2%		
		Concentration		Free residual chlorine: (0.1~50) mg/L	U _{rel} =3.3 %		
57	*Polymerase Chain Reaction Analyzers	Particle concentration	Calibration Specification for Polymerase Chain Reaction Analyzers JJF 1527	(10~100) °C	U=0.2°C		
		Concentration		(0.1~10 ⁸) copies/ μ L	U _{rel} =8.0%		
58	*Fluorescence spectrophotometer	Detection limit	Verification Regulation of Fluorescence	A Type: ≤5×10 ⁻¹⁰ g/mL	U=5.2×10 ⁻¹² g/mL		

No. CNAS L0128

第 120 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.




No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			spectrophotometer JJG 537	BType: $\leq 1 \times 10^{-8} \text{g/mL}$	$U=6.0 \times 10^{-10} \text{g/mL}$		
		Wavelength		(200~700) nm	$U=0.4 \text{nm}$		
59	*Chemical Oxygen Demand (COD) Meters	Temperature		(100~200) °C	$U=0.6^\circ\text{C}$		
		time		(1~3600) s	$U=0.5 \text{s}$		
		concentration (Type A)	Chemical Oxygen Demand (COD) Meters JJG 975	(0.1~50) mg/L	$U_{\text{rel}}=2.6\%$		
				(50~300) mg/L	$U_{\text{rel}}=2.0\%$		
				(300~1000) mg/L	$U_{\text{rel}}=1.8\%$		
		concentration (Type B)		(0.1~100) mg/L	$U=2.6\%$		
60	*Class II Biosafety Cabinets	Wind speed		(0.2~1.5) m/s	$U=0.04 \text{m/s}$		
		Noise		(40~100) dB	$U=3 \text{dB}$		
		illumination	Calibration specification for Class II Biosafety Cabinets JJF 1815	(50~200) lx	$U_{\text{rel}}=14\%$		
				(200~2000) lx	$U_{\text{rel}}=12\%$		
		cleanliness		(0.3~5) μm	$U_{\text{rel}}=29\%$		
		Leakage rate		(0.0001~1) %	$U_{\text{rel}}=30\%$		
61	*Volatile Organic Compounds Photo Ionization Detectors	Concentration	Calibration Specification for Volatile Organic Compounds Photo Ionization Detectors JJF 1172	(0.1~2000) $\mu\text{mol/mol}$	$U_{\text{rel}}=3.1\%$		
62	*Total organic carbon analyzer	Concentration of Inorganic carbon	Verification Regulation of Total Organic Carbon Analyzer JJG 821	(0.01~1000) mg/L	$U_{\text{rel}}=1.6\%$		



No. CNAS L0128

第 121 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Concentration of Organic carbon		(0.01~1000) mg/L	$U_{rel}=U_{rel}=1.8\%$		
63	*Time-of-Flight Mass Spectrometers	mass-to-charge ratio	Calibration Specification for Time-of-Flight Mass Spectrometers JJF 1528	ESI-TOF-MS: (172~4070) u	$U_{rel}=3\times10^{-5}$		
		MALDI-TOF-MS: (1000~67000) u		$U_{rel}=8\times10^{-3}$			
		Signal to noise ratio		ESI-TOF-MS: 1~10 ⁶	$U_{rel}=15\%$		
		MALDI-TOF-MS: 1~10 ⁶		$U_{rel}=18\%$			
64	*On-line Conductivity Meters	Conductivity	Calibration Specification for Online Conductivity Meters JJF(Xin) 19	Electricity: (1~2×10 ⁴) μ S/cm	$U=0.24\%FS$		
		Conductivity		Instrument: (0.9 ~1.3×10 ⁴) μ S/cm	$U=0.8\%FS$		
		Temperature		(0~50) °C	$U=0.3^{\circ}C$		
65	*On-line Automatic Determinators of Chemical Oxygen Demand(COD)	concentration	Verification Regulation of On-line Automatic Determinators of Chemical Oxygen Demand(COD) JJG 1012	(16~1000) mg/L	$U_{rel}=4\%$		
66	*Micro-spectrophotometers	concentration	Calibration Specification for Micro-spectrophotometers JJF 1836	(10~2000) ng/ μ L	$U_{rel}=10\%$		
67	*Alarmer Detector of Sulfur Hexafluoride	concentration	Calibration Specification for the Alarmer Detector of Sulfur Hexafluoride JJF 1263	(0.1~1000) μ mol/mol	$U_{rel}=2.4\%$		
		time		5s~3600s	$U=1s$		



No. CNAS L0128

第 122 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
68	*Bacterial Endotoxin Analyzers	temperature	Calibration Specification for Bacterial Endotoxin Analyzers JJF 1529	(20.0~50.0) °C	$U=0.2^{\circ}\text{C}$		
69	*Colony Counters	Bacterial Colony	Calibration Specification for Colony Counters JJF 1751	(20~300) CFU	$U_{\text{rel}}=3\%$		
70	*Blood Cell Analyzers	concentration	Verification Regulation of Blood Cell Analyzers JJG 714	WBC: (2×10 ⁸ ~2.5×10 ¹⁰) 个/L	$U_{\text{rel}}=4\%$		
				RBC: (1×10 ¹¹ ~6×10 ¹³) 个/L	$U_{\text{rel}}=3\%$		
				HGB: (20~300) g/L	$U_{\text{rel}}=3\%$		
				PLT: (5×10 ⁹ ~5×10 ¹²) 个/L	$U_{\text{rel}}=8\%$		
71	*Hydrogen Chloride Gas Detectors and Alarms	Concentration	Calibration Specification for Hydrogen Chloride Gas Detectors and Alarms JJF 1888	(0.1~100) μmol/mol	$U_{\text{rel}}=3.0\%$		
		Time		(1~3600) s	$U=1\text{s}$		
72	*(Automatic) Nucleic acid extractors	Temperature	Calibration Specification for (Automatic) Nucleic Acid Extractors JJF 1874	(20~100) °C	$U=1.5^{\circ}\text{C}$		
		Vibration frequency		(0.5~100) Hz	$U=0.6\text{Hz}$		
		Volume		(50~200) μL	$U=1.3\text{ μL}$		
		Recovery rate		(10~100) %	$U=2.4\%$		
73	Flue Gas Samplers	flow rate	Verification Regulation of Flue Gas Samplers JJG 1169	(50~2000) mL/min	$U_{\text{rel}}=1.2\%$		
		time		(10~3600) s	$U=0.3\text{ s}$		
		temperature		(0~50) °C	$U=0.2^{\circ}\text{C}$		
		pressure		(-40~40) kPa	$U=0.06\text{ kPa}$		



No. CNAS L0128

第 123 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		atmospheric pressure		(80~106) kPa	U=0.16 kPa		
74	*Plate Electrophoresis Apparatus	Voltage	Calibration Specification for Plate Electrophoresis Apparatus JJF 1654	(0.2~20) V	U=0.3 V		
				(20~1000) V	U=0.4 V		
		Electricity		(2~20) mA	U=0.3 mA		
				20mA~2A	U=0.6 mA		
				(2~10) A	U=0.7mA		
75	*Elemental Analyzers of C、H、N、S	Content	Calibration Specification for Elemental Analyzers JJF 1321	C: (0.1~42) %	U _{rel} =4%		
				H: (0.1~5) %	U _{rel} =1.5%		
				N: (0.1~9) %	U _{rel} =3%		
				S: (0.1~19) %	U _{rel} =4%		
76	*Ultraviolet Fluorescence Sulfur Analyzers	Indication error	Calibration Specification for Ultraviolet Fluorescence Sulfur Analyzers JJF 1685	(0~10.0) mg/L	U=0.2mg/L		
				(10.0~100) mg/L	U _{rel} =4%		
				(100~200) mg/L	U _{rel} =3%		
77	*Flame Photometer	Detection limit	Verification Regulation of Flame Photometer JJG 630	K: ≤0.004mmol/L	U=0.0005mmol/L		
				Na: ≤0.008mmol/L	U=0.0023mmol/L		
78	*Ammonia-Nitrogen Automatic Analyzers	Concentration	Verification Regulation of Ammonia-Nitrogen Automatic Analyzers JJF 631	(0~2) mg/L	U=0.05 mg/L		
				(2~100) mg/L	U _{rel} =2.8 %		
79	*Energy Dispersive X-ray Fluorescence	Concentration	Calibration Specification for Energy Dispersive X-ray Fluorescence Spectrometers	polymer: (5~1200) mg/kg	U _{rel} =6%		



No. CNAS L0128

第 124 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	Spectrometers	Concentration	JJF 2024	metal: (0.001~100) %	U_{rel} =5%		
		Detection Limit		polymer: ≤50mg/kg	U_{rel} =20%		
		Detection Limit		metal: ≤100mg/kg	U_{rel} =25 %		
九、Special test							
1	*Electrosurgical Generator	Output Power	C.S.for Electrosurgical Generator JJF 1217	(1~500)W	U_{rel} =6%		
2	*Friability Surveymeter	rotate speed	C.S.for Disintegration Analyzers JJF(Lu)92	(20~200)r/min	U_{rel} =0.7%		
		Time		(0~300)mm	U =0.2mm		
		Length		1s~10min	U =0.2s		
3	*Cardiac defibrillators	delivered energy	C.S.for Cardiac Defibrillators JJF 1149	(2~40)J	U =2J		
				(40~360)J	U_{rel} =5%		
		pulse frequency		(40~200)min ⁻¹	U_{rel} =1%		
		pulse width		(20~50)ms	U =0.5ms		
		pulse current amplitude		(1~10)mA	U =0.2mA		
				(10~100)mA	U_{rel} =1.3%		
		direct voltage		(0.5~2.0) mV	U_{rel} =2.8%		
		Scanning speed		25mm/s	U_{rel} =0.3%		
	Amplitude-frequency characteristics	1mV (1Hz~25Hz)	U_{rel} =1.4%				



No. CNAS L0128

第 125 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Heart Rate		$(30 \sim 200) \text{min}^{-1}$	$U=1 \text{time/min}$		
4	*Multifunction Patient Monitoring Instruments	Electrocardio voltage	V.R.of Multifunction Patient Monitoring Instruments JJG 1163	$(0.5 \sim 2.0) \text{mV}$	$U_{\text{rel}}=2.8\%$	/	
		Scanning speed		$(25 \sim 50) \text{mm/s}$	$U_{\text{rel}}=0.3\%$		
		Heart Rate		$(30 \sim 200) \text{min}^{-1}$	$U=1 \text{time/min}$		
		Pressure		$(6 \sim 40) \text{kPa}$	$U=0.26 \text{kPa}$		
		Pulse rate		$(30 \sim 200) \text{min}^{-1}$	$U=1 \text{time/min}$		
		Amplitude-frequency characteristics		$1 \text{mV} (1 \text{Hz} \sim 25 \text{Hz})$	$U_{\text{rel}}=1.4\%$		
5	*Ventilators	Tidal Volume	C.S.for Ventilators JJF1234	$(50 \sim 1000) \text{mL}$	$U_{\text{rel}}=5\%$	/	
		Frequency		$(10 \sim 40) \text{min}^{-1}$	$U_{\text{rel}}=4\%$		
		Inspiratory Flow Oxygen Concentration		$21\% \sim 100\%$	$U=4\%$		
		Pressure		$(0.2 \sim 3) \text{kPa}$	$U_{\text{rel}}=6\%$		
6	*Hemodialysis Equipment	Dialysate Conductivity	C.S.for Hemodialysis Equipment JJF 1353	$(12.5 \sim 15.5) \text{mS/cm}$	$U=0.15 \text{mS/cm}$	/	
		Dialysate Temperature		$(25 \sim 40) ^\circ\text{C}$	$U=0.2 ^\circ\text{C}$		
		Dialysate Pressure		$(20 \sim 100) \text{kPa}$	$U=0.4 \text{kPa}$		
		Static (arterial) pulse pressure		$(20 \sim 100) \text{kPa}$	$U=0.4 \text{kPa}$		



No. CNAS L0128

第 126 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Dialysate Flow		(400～800) mL/min	$U_{\text{rel}}=5\%$		
		Anticoagulant pump injection flow		(400～800) mL/min	$U_{\text{rel}}=5\%$		
		Dialysate 1pH Value		pH:(6～10)	$U=0.06\text{pH}$		
		Mass		2kg 、 10kg	$U=1.0\text{g}$		
		Dehydrating capacity		500mL/h、 1000mL/h	$U_{\text{rel}}=2.7\%$		
7	*Tester for Ductility of Bituminous Materials	Temperature	Verification Regulation of Tester for Ductility of Bituminous Materials JJG (JT) 023	(0～50)℃	$U=0.11^{\circ}\text{C}$		
		Speed		(10～50) mm/min	$U=0.02\text{mm/min}$		
		Length		(0～200)mm	$U=0.03\text{mm}$		
				(200～3000)mm	$U=1.1\text{mm}$		
		Roughness		(0.1～10) μ m	$U_{\text{rel}}=6.6\%$		
8	*Electric Compaction Instrument	Weight	Verification Regulation of Compaction Instrument of Soil JJG (JT) 058	(0～5000) g	$U=0.3\text{g}$	/	
		Length		(0～1000) mm	$U=1.0\text{mm}$		
		Length		(0～500) mm	$U=0.03\text{mm}$		
		Length		(0.05～14.80)mm	$U=0.02\text{mm}$		
9	*Apparatus for Softening Point of Bitumen	Weight	Verification Regulation of Apparatus for Softening Point of Bitumen JJG (JT) 057	(0～210) g	$U=0.02\text{g}$	/	
		Length		(0～200) mm	$U=0.03\text{mm}$		
		Temperature		(0～100) ℃	$U=0.1^{\circ}\text{C}$		



No. CNAS L0128

第 127 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Volume		(0~1000) mL	U=2.0mL		
		Heating Rate		(0~10) °C/min	U=0.4°C/min		
10	*Cement Consistency and Setting Time Detector	Weight	Verification Regulation for cement testing apparatus of the onrml consistency and setting time JJG (JC) 105	(0~500) g	U=0.3g		
		Angle		(0~120)°	U=3'		
		Length		(0~100) mm	U=0.04mm		
11	*Jolting Table for Compacting Mortars Specimen	Length	Calibration Specification for Jolting Table for Compacting Mortars Specimen JJF(JC) 124	(0.5~100) mm	U=0.12mm		
		Weight		(0~35000) g	U=0.3g		
		Time		(0~3600) s	U=0.12s		
12	*Biological, Chemical, Sterile Pharmaceutical Production Equipment	Temperature	Calibration Specification for Biological, Chemical, Sterile Pharmaceutical Production Equipment SQI/JL-JF-42	(-100~600)°C	U= (0.03~0.08) °C		
		Pressure		Element pressure gauges; (-0.1~60) MPa	U=0.68%FS		
		Pressure		Digital pressure gauges: (0~60) MPa	U=0.06%FS		
		Revolution Speed		(2.5~30000) r/min	U _{rel} =0.3%		
		Flow		(0.5~38) m³/h	U=0.34%FS		
13	*Inspecting Instrument of Construction Engineering and Building Materials	Length	Calibration Specification for Inspecting Instrument of Construction Engineering and Building Materials SQI/JL-JF-41	Height:(0~1000) mm	U=0.2mm		
		Length		Amplitude:(0~10) mm	U=0.02mm		
		Length		Displacement:(0.5~291.8) mm	U=2.0mm		
		Length		(0~3.5) m	U= 2mm		
		Length		Diameter:(0~500) mm	U=0.04mm		



No. CNAS L0128

第 128 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Length	ilac-MRA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	Gap Value:(0.05~2)mm	$U=0.06\text{mm}$		
		Vertical Degree		(0~200) mm	$U=0.03\text{mm}$		
		Revolution Speed		(2.5~30000) r/min	$U_{\text{rel}}=0.3\%$		
		Angle		(-120~120) °	$U=0.6'$		
		Pressure		Pressure Gauge (-0.1~60) MPa	$U=0.68\%\text{FS}$		
		Pressure		Digital Pressure Gauges (0~60) MPa	$U=0.06\%\text{FS}$		
		Temperature		(-100~600) °C	$U=0.08^{\circ}\text{C}$		
		Time		(0~10) h	$U=0.06\text{s}$		
		Weight		(0~35)kg	$U=1.5\text{ g}$		
14	*Board Compression Strength Tester	force	Verification Regulation of Board Compression Strength Tester JJG(QG)49	(0.5~6000)N	$U_{\text{rel}}=0.4\%$		
15	*Compression Strength Tester for Corrugated Box	force	Verification Regulation of Compression Strength Tester for Corrugated Box JJG(QG)115	(0.0005~50)kN	$U_{\text{rel}}=0.4\%$		
16	*MIT Type Folding Apparatus	force	Verification Regulation of MIT Type Folding Apparatus JJG(QG)59	15N	$U=0.1\text{N}$		
		Rotation Speed		175r/min	$U_{\text{rel}}=0.5\%$		
17	*Paper and Cardboard Bursting Strength Meter	Pressure	Calibration Specification for Bursting Strength Tester for Paper and Board JJF(QG) 116	(0.1~6)MPa	$U=3.9\text{kPa}$		



No. CNAS L0128

第 129 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
18	*Board Puncture Strength Tester	force	Verification Regulation of Board Puncture Strength Tester JJG(QG)56	(0.5~500)N	$U_{rel}=0.4\%$		
19	*Pendulum-type Paper Tensile Breaking Strength Testing Machine	Force	Calibration Specification for Tensile Strength Testing Machines for Paper and Board JJF(QG)115	0.5N~50kN	$U_{rel}=0.4\%$		
20	*Horizontal Tensile Breaking Strength Testing Machine	force	Verification Regulation of Horizontal Tensile Breaking Strength Testing Machine JJG(QG)58.2	(0.5~1000)N	$U_{rel}=0.4\%$		
21	*Paper and Cardboard Thickness Tester	Length	Verification Regulation of Paper and Cardboard Thickness Tester JJG(QG)50.1	(0.5~100)mm	$U=6.2\ \mu\text{m}$		
22	*Corrugated Plate Thickness Tester	Length	Verification Regulation of Corrugated Plate Thickness Tester JJG(QG)50.2	(0.5~100)mm	$U=6.2\ \mu\text{m}$		
23	*Configuration Fatigue Tester for Part of Bicycle	Static force value	Calibration Specification for Configuration Fatigue Tester for Part of Bicycle JJF(QG)106	(0.1~50000) N	$U_{rel}=0.2\%$		
		Dynamic force value		(50~10000) N	$U_{rel}=1.0\%$		
		Frequency		(10~10000) r/min	$U_{rel}=0.2\%$		
		Frequency		(100~10000) 次	$U=1\ \text{次}$		
24	*Static Load Tester for Part of Bicycle	Force value	Calibration Specification for Static Load Tester for Part of Bicycle JJF(QG)121	(10~50000) N	$U_{rel}=0.2\%$		
		Time		(0.01~3600) s	$U=0.13\text{s}$		
25	*Wearing Tester for Axle Bowl, Pedal and Axle of	Force value	Calibration Specification for Wearing Tester for Axle Bowl, Pedal and Axle of	(50~50000) N	$U_{rel}=0.4\%$		



No. CNAS L0128

第 130 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
	Bicycle	Rotation Speed	Bicycle JJF(QG)123	(10~5000) r/min	$U_{rel}=0.6\%$		
26	*Wheel Tires Fatigue Testing Machine	force	Calibration Specification for Wheel Tires Fatigue Tester SQI/JL-JF-17	1N~50kN	$U_{rel}=0.5\%$		
		speed		(0.1~10)m/s	$U_{rel}=3.2\%$		
		Length		(1~300)mm	$U=0.04\text{mm}$		
27	*Brake Performance Tester	force	Calibration Specification for Brake Performance Tester SQI/JL-JF-22	1N~50kN	$U_{rel}=0.69\%$		
		speed		(0.1~10)m/s	$U_{rel}=2.0\%$		
28	*Impact Tester for Part of Bicycle	Length	Calibration Specification for Impact Tester for Part of Bicycle JJF(QG)122	(0~600) mm	$U=0.02\text{mm}$		
				(600~5000) mm	$U=0.7\text{mm}$		
		Quality		10g~1000kg	$U_{rel}=0.2\%$		
		Hardness		(20~70) HRC	$U=0.7\text{HRC}$		
29	*Vibration Tester of Bicycle	Rotation speed	Calibration Specification for Vibration Tester of Bicycle JJF(QG)107	(10~10000) r/min	$U_{rel}=0.2\%$		
		Acceleration		(0.5~10) m/s^2	$U_{rel}=1.0\%$		
		Force value		(50~50000) N	$U_{rel}=0.4\%$		
		Amplitude		(2~20) mm	$U_{rel}=0.5\%$		
30	*Motor Vehicle Testers for Steering Force and Steering Angle	Torque	Calibration Specification for Motor Vehicle Testers for Steering Force and Steering Angle JJF1196	(1~1000) Nm	$U_{rel}=0.72\%$		
		Force		(0.5~1000) N	$U_{rel}=0.60\%$		
		Angle		(1~1000) °	$U=0.1^\circ$		
31	*Roller Type Speedmeter Tester	Speed	Roller Type Speedmeter Tester JJG909	(0.1~200) km/h	$U_{rel}=0.31\%$		



No. CNAS L0128

第 131 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
32	*Measurer for Adhesion Performance of Retroreflective Sheeting	Length	Verification Regulation of Measurer for Adhesion Performance of Retroreflective Sheeting JJG (JT) 083	(0~500)mm	$U=0.07\text{mm}$		
		Weight		(0~850)g	$U=0.2\text{g}$		
33	*Measurer for Impact Resistance of Retroreflective Sheeting	Length	Verification Regulation of Measurer for Impact Resistance of Retroreflective Sheeting JJG (JT) 084	(0~500)mm	$U=0.08\text{mm}$		
		Weight		(0~1000)g	$U=0.2\text{g}$		
34	*Measurer for Resistance to Impact of Raised Pavement Markers	Length	Verification Regulation of Measurer for Resistance to Impact of Raised Pavement Markers JJG (JT) 080	(0~1500)mm	$U=1.0\text{mm}$		
		Weight		(0~2000)g	$U=0.6\text{g}$		
35	*Fibrics Bursting Tester	Pressure	Calibration Specification for Fabric Bursting Tester JJF (FZ) 048	(0~2.5)MPa	$U=0.003\text{MPa}$		
		Time		(0~60)s	$U=0.04\text{s}$		
		Length		Bursting Expansion: (0.5~100)mm	$U=0.02\text{mm}$		
		Length		Aperture: (0~500)mm	$U=0.06\text{mm}$		
36	*Yarn Twist Tester	Force	Calibration Specification for Yarn Twist Tester JJF (FZ) 010	(0.1~20)N	$U_{\text{rel}}=0.2\%$		
		Rotation Rate		(10~5000)r/min	$U_{\text{rel}}=0.3\%$		
		Length		(0~500)mm	$U=0.06\text{mm}$		
37	*Fibre Strength Tester	Length	Calibration Specification for Fibre Strength Tester JJF (FZ) 016	(0~40)mm	$U=0.03\text{mm}$		
		Force		(0.01~200)cN	$U_{\text{rel}}=0.6\%$		
		Mass		(0~100)g	$U=0.020\text{g}$		
		Velocity		(0.05~600)mm/min	$U_{\text{rel}}=0.2\%$		



No. CNAS L0128

第 132 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
38	*Sock Stretch Tester	Length	Calibration Specification for Sock Stretch Tester JJF (FZ) 017	(0~500)mm	<i>U</i> =0.06mm		
		Velocity		(1~200)mm/s	<i>U</i> _{rel} =0.2%		
		Force		(1~100)N	<i>U</i> _{rel} =0.2%		
39	*Wrap Reelers	Force	Calibration Specification for Wrap Reelers JJF (FZ) 019	(5~200)cN	<i>U</i> _{rel} =0.4%		
		Length		(0~1000)mm	<i>U</i> =0.3mm		
		Rotation Rate		(30~250)r/min	<i>U</i> _{rel} =0.5%		
		Time		(0~60)s	<i>U</i> =0.4s		
40	*Textile Yarn Length Tester	Length	Calibration Specification for Textile Yarn Length Tester JJF (FZ) 021	(0~1000)mm	<i>U</i> =0.06mm		
		Force		(0.01~200)cN	<i>U</i> _{rel} =0.6%		
41	*Roller Fiber Length Analyzer	Length	Calibration Specification for Roller Fiber Length Analyzer JJF (FZ) 024	(0~150)mm	<i>U</i> =0.04mm		
42	*Corlour Fastness to Friction Testers	Rotation Rate	Calibration Specification for Corlour Fastness to Friction Testers JJF (FZ) 027	(10~200) r/min	<i>U</i> _{rel} =0.2%		
		Force		(0.5~100)N	<i>U</i> _{rel} =0.4%		
		Length		(0~150)mm	<i>U</i> =0.04mm		
43	*Perspiration Fastness Instruments	Length	Calibration Specification for Perspiration Fastness Instruments JJF (FZ) 028	(0~150)mm	<i>U</i> =0.04mm		
		Force		(0.1~100)N	<i>U</i> _{rel} =1.3%		
44	*Circular Locus Method Fuzzing and Pilling Testers	Speed	Calibration Secification for Circular Locus Method Fuzzing and Pilling Testers JJF (FZ) 031	(0.1~100) r/min	<i>U</i> _{rel} =0.2%		
		Length		(0.1~5) mm	<i>U</i> =0.03mm		
					(5~150) mm		<i>U</i> =0.04mm



No. CNAS L0128

第 133 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		Force		(0.1～1000) cN	<i>U</i> =0.64cN		
45	*Vertical Fabric Crease Recovery Tester	Length	Calibration Specification for Vertical Fabric Crease Recovery Tester JJF (FZ) 032	(1～150)mm	<i>U</i> =0.06mm		
		Mass		1g～2kg	<i>U</i> =0.15g		
		Time		(1～600)s	<i>U</i> =0.13s		
		Angle		(0～200)°	<i>U</i> =0.02°		
46	*Fabric Drapability Tester	Rotation Rate	Calibration Specification for Fabric Drapability Tester JJF (FZ) 033	(10～500)r/min	<i>U</i> _{rel} =1.4%		
		Length		180mm, 120mm	<i>U</i> =0.06mm		
		Drape coefficient		30%～100%	<i>U</i> _{rel} =1.5%		
		Time		(0～300)s	<i>U</i> =0.4s		
47	*Fabric Flat-Rubbing Tester	Rotation Rate	Calibration Specification for Fabric Flat-Rubbing Tester JJF (FZ) 036	(10～500)r/min	<i>U</i> _{rel} =0.3%		
		Length		(0.1～150)mm	<i>U</i> =0.04mm		
		Mass		(0.1～600)g	<i>U</i> =0.020g		
				(600～1000)g	<i>U</i> =0.13g		
				(1000～2000)g	<i>U</i> =0.15g		
				(2000～3000)g	<i>U</i> =0.20g		
48	*Fuzzing and Pilling Tester	Length	Calibration Specification for Fuzzing and Pilling Tester JJF (FZ) 053	(0～300)mm	<i>U</i> =0.05mm		
		Friction Coefficient		0.4～1.0	<i>U</i> _{rel} =0.2%		
		Rotation Rate		(10～100) r/min	<i>U</i> _{rel} =1.4%		



No. CNAS L0128

第 134 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Mass		(0.1~100)g	$U=0.020\text{g}$		
		hardness		(20~80) HA	$U=1.2\text{HA}$		
49	*Tester of Filament Crimp Shrinkage	Force	Calibration Specification for Tester of Filament Crimp Shrinkage JJF (FZ) 057	(0.025~50)N	$U_{\text{rel}}=0.58\%$		
		Length		(0~1000)mm	$U=0.22\text{mm}$		
		Time		0.1s~1h	$U=0.18\text{s}$		
50	*Electronic Fabric Strength Machines	Time	Calibration Specification for Electronic Fabric Strength Machines JJF (FZ) 062	(0~3600)s	$U=0.13\text{s}$		
		Length		(0~500)mm	$U=0.06\text{mm}$		
		Force		(0.1~5000)N	$U_{\text{rel}}=1.3\%$		
		Mass		(0.1~3000)g	$U=0.15\text{g}$		
51	*Down-proof Properties of Fabrics(Rubbing Test)	Length	Calibration Specification for Down-proof Properties of Fabrics(Rubbing Test) JJF (FZ) 064	(0~500)mm	$U=0.06\text{mm}$		
		Rotation Rate		(10~500)r/min	$U_{\text{rel}}=0.3\%$		
52	*Fabric Shedding Testers	Length	Calibration Specification for Fabric Shedding Testers JJF (FZ) 084	(0~500)mm	$U=0.06\text{mm}$		
		Rotation Rate		(10~100)cpm	$U_{\text{rel}}=0.2\%$		
		Mass		(0.5~1200)g	$U=0.15\text{g}$		
53	*Fabric Induction Electrometer Meter	Time	Calibration Specification for Fabric Induction Electrometer Meter JJF (FZ) 070	(0~60)s	$U=0.16\text{s}$		
		Rotate speed		(50~2000) r/min	$U_{\text{rel}}=0.12\%$		
		Length		(0~300)mm	$U=0.03\text{mm}$		
		Voltage		(0.05~12)kV	$U_{\text{rel}}=0.6\%$		
54	*Rotary Friction	Time	Calibration Specification for	(0~600)s	$U=0.16\text{s}$		



No. CNAS L0128

第 135 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
	Electrostatic Tester	Rotate speed	Rotary Friction Electrostatic Tester JJF (FZ) 069	(50~500) r/min	$U_{rel}=0.12\%$		
		Length		(0~300)mm	$U=0.03\text{mm}$		
		Voltage		(0.01~1000)V	$U_{rel}=0.2\%$		
		Mass		(0~1000)g	$U=0.024\text{g}$		
55	*Textiles 45° Combustion Tester	Length	Calibration Specification for Textiles 45° Combustion Tester JJF(FZ)087	(0~150)mm	$U=0.04\text{mm}$		
		Length		(150~1000)mm	$U=0.5\text{mm}$		
		Time		(0~60)min	$U=0.16\text{s}$		
		Weight		(0~500)g	$U=0.018\text{g}$		
		Degree		(0~90)°	$U=0.2^\circ$		
56	*Vertical Combustion Testers	Length	Calibration Specification for Vertical Combustion Testers JJF(FZ)068	(0~150)mm	$U=0.06\text{mm}$		
		Length		(150~900)mm	$U=0.5\text{mm}$		
		Time		(0~10)min	$U=0.17\text{s}$		
		weight		(0~500)g	$U=0.012\text{g}$		
		Degree		(0~90)°	$U=0.2^\circ$		
		Speed		(0~60)mm/s	$U=0.36\text{mm/s}$		
57	*Textile Frictional Static Charges Tester	Static Charges	Calibration Specification for Textile Frictional Static Charges Tester JJF (FZ) 071	(0.05~1.1) μC	$U=0.002 \mu\text{C}$		
		Length		(0~1000)mm	$U=0.5\text{mm}$		
58	*Scorch and Sublimation Tester	temperature	Calibration Specification for Scorch and Sublimation Tester JJF (FZ) 029	deviation: (0~250) °C	$U=1.3^\circ\text{C}$		



No. CNAS L0128

第 136 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				volatility: (0~250) °C	U=0.5°C		
		Time		1ms~30s	U=0.40s		
		gravity value		(0.4~16) N	U=0.4N		
59	*Eight-Basket Oven	Temperature	Calibration Specification of Eight-Basket Oven JJF	(0~300) °C	U=0.3 °C		
		Basket quality	(Textile) 011	(1~500)g	U=2.0mg		
60	*Constant Temperature Humidity Chambers	Temperature	Calibration Specification of Constant Temperature Humidity Chamber JJF(Textile)060	(0~300) °C	U=0.22 °C		
		Humidity		(10~90) %RH	U=1.2%RH		
		wind speed		(0.4~20) m/s	U=0.10m/s		
61	*Light and Weather Fastness Testers	temperature		Test chamber: (0~80) °C	U=0.5 °C		
				Black plate: (20~150) °C	U=1.1 °C		
				Black standard: (20~150) °C	U=0.9 °C		
		humidity	Calibration Specification for Light and Weather Fastness Testers JJF (Textile) 051	(10~90)%RH	U=1.5%RH		
		Speed		(0.1~10) r/min	U _{rel} =0.7%		
		Time		0.1s~1h	U=0.2s		
		Iiradiance		300nm~400nm: (0.1~50.00) W/m ²	U _{rel} =10%		
				420nm: (0.01~2.00) W/m ²	U _{rel} =10%		
	*Textiles Solar	Wavelength	Calibration Specification for Textiles Solar Ultraviolet	(200~500) nm	U=0.8nm		



No. CNAS L0128

第 137 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	Radiation Protective Properties Testers	Transmittance	Radiation Protective Properties Testers JJF (FZ) 081	8%~35%	U _{rel} =1.6%		
				0.8%~1.1%	U _{rel} =14%		
63	*Medicinal Magnetic Resonance Imaging System (MRI)	Magnetic field strength	V. R. for Medicinal Magnetic Resonance Imaging System JJG (HU) 54	10mT~1900mT	U _{rel} =1.5%		
		Length		0.5mm~100mm	U=0.3mm		
64	*Voltage Dips, Short Interruptions and Voltage Variations Test Generators	Voltage	Calibration Specification for Voltage Dips, Short Interruptions and Voltage Variations Test Generators JJF 1673	10mV~360V	U _{rel} =0.2%		
		Rising and falling time		Rise time (1~50) μs, Descent time (1~50) μs	U _{rel} =4.0%		
		Time		1ms~60s	U _{rel} =2.0%		
		Phase angle		(0~360) °	U=2°		
		Peak impulse current		(10~1200) A	U _{rel} =3.2%		
65	*Oscilloscope Current Probes	DC Current	Calibration Specification of Oscilloscope Current Probes JJF (Electronics) 0036	±(10mA~100A)	U _{rel} =0.08%		
				±(100A~750A)	U _{rel} =0.8%		
		AC Current		10mA~100A	U _{rel} =0.08%		
				100 A~500 A	U _{rel} =0.8%		
		DC Current attenuation coefficient		1:1~1000:1 ± (10mA~100A)	U _{rel} =0.08%		
				1:1~1000:1 ± (100A~750A)	U _{rel} =0.8%		
		AC Current attenuation		1:1~1000:1 (10mA~100A)	U _{rel} =0.08%		



No. CNAS L0128

第 138 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		coefficient		1:1~1000:1 (100 A~500 A)	$U_{rel}=0.8\%$		
		Bandwidth		DC~200MHz	$U=0.21\text{dB}$		
		Rise/fall time		(1~50)ns	$U_{rel}=6.3\%$		
66	*Flow analyzers with Spectrophotography	Wavelength	Calibration Specification for Flow analyzers with Spectrophotography JJF 1568	(360~800) nm	$U=1.6\text{nm}$		
		Detectionlimit		Cyanogen: (0~0.1) mg/L	$U=0.0003\text{mg/L}$		
		Detectionlimit		Volatile Phenol: (0~0.1) mg/L	$U=0.0003\text{mg/L}$		
		Detectionlimit		Hexavalent Chromium: (0~1) mg/L	$U=0.001\text{mg/L}$		
		Detectionlimit		Sulfide: (0~1) mg/L	$U=0.002\text{mg/L}$		
		Detectionlimit		Total Phosphorus: (0~5) mg/L	$U=0.002\text{mg/L}$		
		Detectionlimit		Total Nitrogen : (0~5) mg/L	$U=0.02\text{mg/L}$		
		Detectionlimit		Ammonia Nitrogen: (0~5) mg/L	$U=0.02\text{mg/L}$		
		Detectionlimit		Anion Active Derergent: (0~5) mg/L	$U=0.04\text{mg/L}$		
		Detectionlimit					
67	*Calibration Specification of Top Open Washing Shrinkage Testers	water level	Calibration Specification of Top Open Washing Shrinkage Testers JJF(FZ) 092	(30~80)L	$U=0.5\text{L}$		
		Speed		(20~180)次/min	$U=0.5$ 次/min		
		time		(180~3600)s	$U=5.0\text{s}$		
		rotation rate		(20~1000)r/min	$U_{rel}=0.3\%$		



No. CNAS L0128

第 139 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
68	*Down Filling Power Tester	Length	Calibration Secification for Down Filling Power Tester JJF (FZ) 074	Heigth (0~1000) mm	U=0.3mm		
				Inner diameter: (0~300) mm	U=0.10mm		
				Plate diameter: (0~300) mm	U=0.10mm		
		Quality		(1~500) g	U=0.020g		
69	*Crimp Elastic Tester	Fixed length distance	Calibration Specification for Crimp Elastic Tester JJF (FZ) 040	Length: (0~50) mm	U=0.04mm		
		Quality		Descending stroke: (0~30) mm	U=0.02mm		
				(0~1000) mg	U=1.7mg		
				Time	(0~360) s		
		Force		(0.1~100) mN	U=0.01mN (0.2index)		
		70		*Damped Oscillatory Wave Simulators	Voltage		
Current	(1.25~80) A		U _{rel} =3.5%				
Time	10ns~10s		U _{rel} =3.1%				
Frequency	100kHz~30MHz		U _{rel} =3.1%				
71	*Impulse voltage measurement system	Voltage	Calibration Regulation of impulse voltage measurement systems JJF(JX)1029	(1~300) kV	U _{rel} =1.0%		
		Time		100ns~1s	U _{rel} =2.0%		
72	Pitot tubes	Calibration coefficient	Verification Regulation of Pitot Tubes JJG 518	L-type : 0.99~1.01	U _{rel} =2.7%	Except for L Standard Pitot Tube	
				S-type: 0.81~0.86	U _{rel} =2.4%		
73	Thermo-	wind speed	Calibration Specification for Thermo-anemoscopes JJF	(0.2~5) m/s	U=0.10m/s		



No. CNAS L0128

第 140 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			1939	(5~30) m/s	U= (0.12~0.16) m/s		
74	*Urine Analyzers	urine protein concentration	Calibration Specification of Urine Analyzers JJF 1129	(0.1~3.0) g/L	U _{rel} =10%		
		pH		4.5~8.0	U _{rel} =5%		
		urine glucose concentration		(0.1~56) mmol/L	U _{rel} =6%		
		specific gravity		1.000~1.030	U=0.007		
		red blood cell concentration		(10~200) count/μL	U _{rel} =11%		
		white blood cell concentration		(15~300) 个/μL	U _{rel} =12%		
十、Gvometric sense							
1	GNSS receivers	consistency of antenna phase center	Verification Regulation of GNSS Receivers (Geodetic or Navigational Type) JJG1200	Antenna phase center consistency: (0~10) mm	U=1.8mm		
		length		(0~46.5) km	U= (1.5~16) mm		
2	Fineness of Grind Gage	length	Verification Regulation for Fineness of Grind Gage JJG905	(0~150) μ m	U=(0.5~1.1) μ m		
十一、Engineering parameter							
1	*ExtensoMeter	displacement	Verification Regulation of Extensometers JJG762,Standard Practice for	(0.02~0.3) mm	U=0.72um		
				(0.3~50) mm	U _{rel} =0.17%		



No. CNAS L0128

第 141 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
			Verification and Classification of Extensometer Systems ASTM E83	(50~800) mm	$U_{rel}=0.20\%$		
十二、Ionizing radiation							
1	*X-ray Radiation Sources for Medical Computed Radiography System and Digital Radiography System	ESAK	Verification Regulation of X-ray Radiation Sources for Medical Computed Radiography System and Digital Radiography System JJG1078	0.1 μ Gy~999Gy	$U_{rel}=5.5\%$		
		Tube voltage		(50~120)kV	$U_{rel}=2.4\%$		
2	*X-ray radiation source of medical diagnostic digital subtraction angiography (DSA) system	air kerma rate	Verification regulation of X-ray radiation source of medical diagnostic digital subtraction angiography (DSA) system JJG 1067	60 μ Gy/min~1Gy/min	$U_{rel}=5.9\%$		
		tube voltage		(50~120)kV	$U_{rel}=2.4\%$		
3	*Medical Diagnostic X-ray Source for Dental Panorama	air kerma rate	Verification Regulation of Medical Diagnostic X-ray Source for Dental Panorama JJG 1101	60 μ Gy/min~1Gy/min	$U_{rel}=5.9\%$		
		tube voltage		(50~120)kV	$U_{rel}=2.4\%$		
		time		50ms~1s	$U_{rel}=1.2\%$		
4	*Medical X-ray Radiation Sources for Mammographic Equipment	average glandular dose	Verification Regulation of Medical X-ray Radiation Sources for Mammographic Equipment JJG 1145	0.1mGy~0.1Gy	$U_{rel}=9.3\%$		
		tube voltage		(22~40)kV	$U_{rel}=2.5\%$		



No. CNAS L0128

第 142 页 共 143 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
5	*Medical Diagnostic X-ray Radiation Source for Spiral Computed Tomography(CT)	CT Dose Index	Verification Regulation of Medical Diagnostic X-ray Radiation Source for Spiral Computed Tomography(CT) JJG 961	0.01mGy~300mGy	$U_{rel}=4.6\%$		

SCHEDULE OF ACCREDITATION CERTIFICATE



在线扫码获取验证

No. CNAS L0128

The scope of the accreditation in Chinese remains the definitive version.